**Automatic determination of arsenate in drinking water by flow analysis with
dual membrane-based separation**

Ruben Vera^{a,b}, Yanlin Zhang^a, Clàudia Fontàs^b, M. Inês G.S. Almeida^a, Enriqueta Anticó^b,
Robert W. Catrall^a, Spas D. Kolev^{a,*}

^a School of Chemistry, The University of Melbourne, Victoria 3010, Australia

^b Chemistry Department, University of Girona, C/ Maria Aurèlia Capmany, 69, 17003 Girona,
Spain.

Email addresses:

Ruben Vera - ruben.vech@gmail.com

Yanlin Zhang - yanlinz@unimelb.edu.au

Clàudia Fontàs - claudia.fontas@udg.edu

M. Inês G.S. Almeida - maria.gameirodesaalmeida@unimelb.edu.au

Enriqueta Anticó - enriqueta.antico@udg.edu

Robert W. Catrall - r.catrall@unimelb.edu.au

Spas D. Kolev – s.kolev@unimelb.edu.au

* Corresponding authors: S.D. Kolev: Tel. +61 3 8344 7931; E-mail address s.kolev@unimelb.edu.au;

Abstract

The sequential application of a polymer inclusion membrane (PIM), composed of poly(vinylidene fluoride-co-hexafluoropropylene) and the anionic extractant Aliquat 336, and a microporous polytetrafluoroethylene (PTFE) gas-permeable membrane was utilized for the first time to develop a flow analysis (FA) system, for the automatic determination of trace levels of arsenate (As(V)) in drinking water as arsine. The system incorporated a flow-through extraction cell for separation and preconcentration of arsenate and a gas-diffusion cell for the separation of arsine prior to its spectrophotometric determination based on the discoloration of a potassium permanganate solution. Under optimal conditions the FA system is characterized by a limit of detection of $3.0 \mu\text{g L}^{-1}$ As(V) and repeatability of 1.8% ($n=5$, $25 \mu\text{g L}^{-1}$ As(V)) and 2.8% ($n=5$, $50 \mu\text{g L}^{-1}$ As(V)). The newly developed FA method was successfully applied to the determination of arsenate in drinking water samples in the $\mu\text{g L}^{-1}$ concentration range.

Keywords: drinking water; arsenate; flow analysis; polymer inclusion membrane (PIM); gas-diffusion separation; hydride generation.

1. Introduction

Arsenic is a naturally occurring toxic element, which is present in natural waters around the world (Villaescusa & Bollinger, 2008). Inorganic arsenic species, such as arsenate (As(V)) and arsenite (As(III)), are the most common and toxic forms of arsenic found in aquatic systems (Vera, Fontas, & Antico, 2017). Arsenic is considered a leading pollutant since it is often found at elevated levels in natural waters and long-term exposure to its forms have been associated with skin, lung, urinary tract, kidney, and liver cancer (Bissen & Frimmel, 2003). Therefore, the World Health Organization (WHO) has set the guideline concentration for arsenic in drinking water at $10 \mu\text{g L}^{-1}$ (WHO, 2011). It should be pointed out that arsenic in drinking water is present very often almost entirely as arsenate (As(V)) (Döker & Yılmaz, 2018; Komorowicz & Baralkiewicz, 2016). The low regulated level of arsenic and its complex chemistry represent a challenge from an analytical point of view. Hence, a great number of highly sensitive analytical techniques have been developed and employed for the determination of arsenic in environmental samples, namely graphite furnace atomic absorption spectrometry (GFAAS) (Alves, Neri, Borges, Carvalho, & Coelho, 2017), hydride generation atomic absorption spectrometry (HG-AAS) (Susko, Bloom, Neamtiu, Appleton, Surdu, Pop, et al., 2017), hydride generation atomic fluorescence spectrometry (HG-AFS) (Chen, Lai, Mao, Chen, & Chen, 2017), inductively coupled plasma atomic emission spectrometry (ICP-AES) (Güell, Anticó, Kolev, Benavente, Salvadó, & Fontàs, 2011), and inductively coupled plasma mass spectrometry (ICP-MS) (Fontàs, Vera, Batalla, Kolev, & Anticó, 2013; Vera, Fontas, & Antico, 2017). These techniques provide the sensitivity required to directly measure arsenic concentrations in water samples at the $\mu\text{g L}^{-1}$ level. However, the techniques mentioned above require expensive equipment and highly trained laboratory technicians.

Flow injection analysis (FIA) is a technique suitable for performing analysis on-line in an automatic fashion and it is highly efficient in minimizing both reagent and sample consumption

as well as the overall analysis time and associated costs (Cerdeja & Estela, 2008; Valcarlos & Luque de Castro, 1987). Different detection techniques have been successfully applied in FIA systems for the determination and speciation of arsenic (e.g., voltammetry (Fogg & Bsebsu, 1981), amperometry (Farrell, Iles, & Yuan, 1996; Rupasinghe, Cardwell, Catrall, & Kolev, 2009), chemiluminescence (Lomonte, Currell, Morrison, McKelvie, & Kolev, 2007), or spectrophotometry (Boonjob, Miró, & Kolev, 2013; Rupasinghe, Cardwell, Catrall, Luque de Castro, & Kolev, 2001; Rupasinghe, Cardwell, Catrall, Potter, & Kolev, 2004). A great number of spectrophotometric methods for arsenic are based on the method proposed by Johnson and Pilson (Johnson & Pilson, 1972), in which an arsenomolybdenum blue complex is formed. However, this method is affected by severe interferences from silicate or phosphate, often present in arsenic samples, which impose serious limitations on the applicability of this method. To avoid the interference of phosphate and silicate, some authors have used anion-exchange columns to retain the interfering anions (Frenzel, Titzenthaler, & Elbel, 1994; Narusawa, 1988) or optimized the molybdenum blue method to improve its selectivity for arsenate over phosphate, as reported by Dhar et al. (Dhar, Zheng, Rubenstone, & Van Geen, 2004). Rupasinghe et al. (Rupasinghe, Cardwell, Catrall, Potter, & Kolev, 2004) and Toda et al. (Toda & Ohba, 2005) have reported on the development of FIA systems based on hydride generation where arsenic is converted into arsine followed by bleaching an oxidant acceptor solution containing KMnO_4 . The concentration of arsenic in many water samples is at trace level and preconcentration is often required.

Membrane-based extraction procedures involving liquid membranes have emerged as promising alternatives to ion-exchange based separation and preconcentration where retention and stripping of the analyte take place sequentially. In liquid membrane-based separation the extraction and back-extraction of the analyte from a donor aqueous stream into an acceptor aqueous stream occur simultaneously. Supported liquid membranes (SLMs), which are

considered as the most frequently used type of liquid membranes at present, have been used successfully in the determination of arsenate in drinking water (Kamyabi & Aghaei, 2016). However, in this type of membranes the membrane liquid phase, consisting of an extractant and diluent, is retained in the micrometre size pores of a hydrophobic polymeric membrane and this leads to leaching of the membrane liquid phase into the donor and acceptor aqueous phases, thus causing potential deterioration in the performance of the SLM (Almeida, Cattrall, & Kolev, 2017).

Recently, polymer inclusion membranes (PIMs) have been shown to have a better stability than SLMs (Almeida, Cattrall, & Kolev, 2012). PIMs are cast from a solution of a base-polymer, extractant and in some cases plasticizer or modifier in a suitable solvent (Almeida, Cattrall, & Kolev, 2012; Nghiem, Mornane, Potter, Perera, Cattrall, & Kolev, 2006). The reason behind their superior stability compared to SLMs stems from the fact that the membrane liquid phase of PIMs (i.e., extractant and plasticizer/modifier) is retained between the entangled polymer chains of the base-polymer, thus minimizing significantly its leaching to the adjacent aqueous solutions. The base-polymer provides mechanical strength to the PIM, while the extractant (carrier) is responsible for the extraction/transport of the chemical species of interest. The plasticizer or modifier are often added to the PIM composition to provide elasticity or increased solubility of the extracted species in the membrane liquid phase, respectively (Nghiem, Mornane, Potter, Perera, Cattrall, & Kolev, 2006). PIMs have been successfully employed in flow analysis (FA) systems for the on-line separation and preconcentration of Zn(II) (L. L. Zhang, Cattrall, Ashokkumar, & Kolev, 2012; L. L. Zhang, Cattrall, & Kolev, 2011), orthophosphate (Nagul, Fontàs, McKelvie, Cattrall, & Kolev, 2013) and vanadium(V) (Yaftian, Almeida, Cattrall, & Kolev, 2018).

The present paper reports on the development of a spectrophotometric FA system implementing on-line preconcentration of arsenate using a PIM consisting of poly(vinylidene

fluoride-co-hexafluoropropylene) (PVDF-HFP) and Aliquat 336 followed by on-line generation of arsine which diffuses across a gas-permeable membrane into a KMnO_4 solution causing its discoloration. To the best of our knowledge this is the first use of a PIM in an FA system for the determination of arsenate in drinking waters at low $\mu\text{g L}^{-1}$ levels and the first coupling of on-line membrane-based extractive separation with on-line membrane-based gas-diffusion separation.

2. Experimental

2.1. Reagents and solutions

All reagents and solvents used in this study were of analytical reagent grade. The polymers PVDF-HFP (Aldrich, USA) and poly(vinyl chloride) PVC (Fluka, Italy), the extractant Aliquat 336 (Aldrich, USA), and the modifier 1-tetradecanol (Aldrich, USA) were used as constituents of the PIMs studied. Tetrahydrofuran (THF) without a stabilizer, purchased from VWR (Australia), was used as the membrane casting solvent. The acceptor solution used in the PIM-based separation step contained 0.1 mol L^{-1} NaCl (Chem-Supply, Australia) as the stripping reagent for arsenate. The reduction of As(V) to As(III) was conducted using a reductant solution composed of 4 mol L^{-1} HCl (32%, RCI Labscan, Thailand), 1% (w/v) KI (Aldrich, USA), and 0.5% (w/v) ascorbic acid (AA) (Ajax Finechem, Australia). The sodium borohydride reagent stream used for arsine generation contained 0.5% (w/v) NaBH_4 and 0.05 mol L^{-1} NaOH (Chem-Supply, Australia). Arsine was absorbed and oxidized in the gas-diffusion acceptor stream containing 0.2 mmol L^{-1} KMnO_4 (Chem-Supply, Australia) and 0.05 mol L^{-1} NaOH (Chem-Supply, Australia).

The interference studies were performed with working solutions prepared by dilution of stock solutions containing 500 mg L^{-1} H_2PO_4^- , Cl^- , NO_3^- , HCO_3^- , or SO_4^{2-} . These stock solutions were prepared by dissolving Na_2HPO_4 (BDH, Australia), NaCl, NaNO_3 (Ajax,

Australia), NaHCO₃ (Chem-Supply, Australia), or Na₂SO₄ (Chem-Supply, Australia) in ultrapure water (≥ 18.2 M Ω cm, Millipore, Synergy 185, France), used in the preparation of all aqueous solutions.

2.2. Instrumentation

On-line spectrophotometric detection was conducted with a Pharmacia Novaspec II UV-Vis spectrophotometer (Pharmacia Biotech, Sweden) fitted with a flow-through cell made of quartz (10 mm optical path length, Starna, UK). The spectrophotometer was interfaced with a PowerChrom 280 (Model ER280) data recording system linked to a PC and run by the Chart software package (eDAQ, Australia).

The PIMs thickness was measured using an optical microscope (Model LH50A, Olympus, Japan) with a calibrated lens (Carton Optical Ind., Japan).

For method validation the samples were also analysed after off-line pre-reduction with a solution containing a mixture of 1% (w/v) KI and 0.5% (w/v) ascorbic acid by inductively coupled plasma optical emission spectrometry (ICP-OES, Model Optima 4300 DV, Perkin-Elmer) incorporating a home-made hydride generation unit.

2.3. Flow Analysis (FA) manifold

The FA manifold developed in the present study for arsenate preconcentration, separation and detection involving hydride generation is depicted in Figure 1.

Figure 1. Schematic of the FA manifold. P1-P3: peristaltic pumps; R1: gas-diffusion acceptor stream (0.2 mM KMnO_4 , 0.05 M NaOH); R2: NaBH_4 stream (0.5% (w/v) NaBH_4 , 0.05 mol L^{-1} NaOH); R3: reductant stream (4 M HCl , 1% (w/v) KI , 0.5% (w/v) ascorbic acid); R4: PIM acceptor stream (0.1 M NaCl); R5: PIM donor stream; RC: reaction coil; IV: injection valve; GDC: gas-diffusion cell; PIM: polymer inclusion membrane.

The system consisted of 3 four-channel peristaltic pumps, i.e., Pump 1 and Pump 2 (Model VS4, Watson Marlow Alitea, Sweden) and Pump 3 (Gilson Minipuls-3, France). All the pumps were fitted with Tygon tubing of suitable internal diameter (TACS, USA). Polytetrafluoroethylene (PTFE) tubing of 0.5 mm i.d. was used throughout the manifold, except for the gas-diffusion acceptor stream outlet tubing, which was of 3 m length and 0.3 mm i.d. to provide sufficient back-pressure. The latter was required to prevent the diffusion of H_2 , generated by the decomposition of NaBH_4 , across the hydrophobic microporous membranes and the filter paper of the gas-diffusion cell (GDC, Fig. 2a) into Stream R1 where it would have interfered with the analytical measurements. The following hydrophobic microporous

membranes were used in the present study: Durapore® and SureVent® membranes (Merck Millipore, USA), PTFE membranes (Reece, Australia), and polypropylene membranes (Chemplex, Zimbabwe). The flow rates of all streams were measured gravimetrically by weighing the mass of water of known temperature pumped through the corresponding tubing over a 5 min period. On-line preconcentration of arsenate was performed using a home-made extraction cell similar to the one described previously by us (L. L. Zhang, Cattrall, & Kolev, 2011), which consisted of two Perspex blocks (150 mm length, 30 mm width and 15 mm height, each) clamped together by stainless steel screws. The two channels of the extraction cell were serpentine shaped and were 157, 1 and 0.25 mm in length, width and depth, respectively. Arsine was separated in a homemade GDC (Figure 2) made of Perspex and identical to the one used previously by us (Y. Zhang, Miró, & Kolev, 2015) where arsine diffused from the gas-diffusion donor stream (Streams R2+R3+R4, Figure 1) across an assembly of a filter paper disc (No. 54, Whatman, Britain) sandwiched between two hydrophobic microporous membranes (Figure 2a) into the gas-diffusion acceptor stream (Stream R1). The filter paper was used as a physical support for the hydrophobic membranes, which otherwise could have stretched as a result of the pressure difference between the two channels of the GDC (Figure 2a) thus changing the channels' volume and impacting negatively on repeatability. The shape of the two channels of identical width and length (Figure 2b), i.e., 1.8 mm and 100 mm, respectively, ensured efficient mixing of the gas-diffusion donor and acceptor streams which improved the generation, transmembrane transfer and oxidation of arsine in the gas-diffusion acceptor stream (Stream R1) (Y. Zhang, Miró, & Kolev, 2015). The depths of the acceptor and the donor channels were 0.5 and 6 mm, respectively, and the corresponding volumes were 90 μ L and 1080 μ L, respectively. This volume difference coupled with appropriately selected flow rates of Streams R1 – R4 allowed a degree of preconcentration of arsenic as arsine in the gas-diffusion acceptor stream (Stream R1).

Figure 2. Schematic of the GDC used in the on-line separation of arsine. (a) Cross-section (donor and acceptor channels depths - 6 and 0.5 mm, respectively) and (b) top view of one of the halves of the GDC.

2.4. FA procedure

The standard/sample solution (Stream R5, Figure 1) was propelled for a predetermined period of time through the donor channel of the extraction cell where a PIM separated the sample (donor) stream (Stream R5) from the acceptor stream (Stream R4). The acceptor stream was stopped for a predetermined period of stop-flow time during the sample passage through the donor channel of the extraction cell to allow preconcentration of arsenate in the static acceptor solution located in the acceptor channel of the cell. At the end of the stop-flow time, the acceptor stream (R4) was re-started and arsenate was reduced to arsenite by merging the acceptor stream of the extraction cell (R4) with a reagent stream (R3) containing HCl, KI and ascorbic acid. Subsequently, arsine was generated by merging the combined R4+R3 stream with a sodium borohydride stream (R2). The generated arsine in the combined stream R4+R3+R2 diffused across the hydrophobic membrane of the GDC into the acceptor solution of the gas-diffusion cell (R1) where it was oxidised by KMnO_4 resulting in a decrease in the KMnO_4 absorbance, monitored continuously at 528 nm in the spectrophotometric measuring cell of the manifold. In all measurements, the analytical signal recorded was the maximum decrease in KMnO_4 absorbance relative to the baseline level.

2.5. Optimization of the FA method

The optimization of the reaction coil (RC) length (Figure 1) and the flow rate of Stream R1 and the selection of the most appropriate hydrophobic gas-diffusion membrane were carried out in a FA system similar to the one shown in Figure 1 where the extraction cell was replaced with an injection valve with a 500 μL sample loop. The standards injected in these experiments contained 1000 $\mu\text{g L}^{-1}$ As(V).

The suitability of different PIM compositions was tested in the FA manifold shown in Figure 1 using a stop-flow procedure in which 5 mL of a 1000 $\mu\text{g L}^{-1}$ As(V) standard solution were propelled at a flow rate of 0.2 mL min^{-1} through the donor channel of the extraction cell. The influence of the stop-flow time and the flow rate of Stream R5 was studied by propelling a standard solution containing 500 $\mu\text{g L}^{-1}$ As(V) through the donor channel of the extraction cell.

2.6. PIM preparation

PVC-based PIMs containing 70% (w/w) PVC and 30% (w/w) Aliquat 336 were prepared by dissolving 180 mg of Aliquat 336 in 18 mL of THF, followed by slow addition of 420 mg of PVC into the casting solution, which was constantly stirred to avoid aggregation of the polymer. Finally, the resulting mixture was poured into a 16.5 cm in diameter glass ring sitting on a flat glass plate. The ring was covered with filter paper and a watch glass to slow down the evaporation of THF in the next 15 h after which the resulting PIM was carefully peeled from the glass plate (Fontàs, Vera, Batalla, Kolev, & Anticó, 2013; Nagul, Fontàs, McKelvie, Cattrall, & Kolev, 2013).

PIMs containing 1-tetradecanol as a modifier were also prepared by the casting method outlined above. However, in this case 60 mg of this compound and 120 mg of Aliquat 336 were

dissolved in the casting solution together with 420 mg of PVC and the corresponding PIMs contained 70% (w/w) PVC, 20% (w/w) Aliquat 336 and 10% (w/w) 1-tetradecanol.

The PVDF-HFP-based membranes were prepared following the procedure described by O'Bryan et al. (O'Bryan, Cattrall, Truong, Kyratzis, & Kolev, 2016). In this method, 700 mg of PVDF-HFP and 300 mg of Aliquat 336 were dissolved in 8 mL of THF at 50 °C and the mixture was mechanically stirred until the complete dissolution of all PIM components. The casting solution was then spread onto a glass plate using a casting knife with 0.5 mm depth setting (O'Bryan, Cattrall, Truong, Kyratzis, & Kolev, 2016). The glass plate was covered with an aluminium tray to allow the slow evaporation of THF in the next 48 h after which the membrane was peeled from the glass plate.

2.7. Interference studies

The effect of common anions in appropriately selected concentration ranges (i.e., 0.15 mg L⁻¹ – 140 mg L⁻¹ in the case of H₂PO₄⁻ and 1.0 mg L⁻¹ - 40 mg L⁻¹ in the case of NO₃⁻, Cl⁻, HCO₃⁻, and SO₄²⁻) on the analytical signal for a 0.05 mg L⁻¹ (0.67 µmol L⁻¹) As(V) standard was studied.

2.8. Sample analysis

Spiked with arsenate at the µg L⁻¹ level tap and mineral water samples were analysed by both the newly developed FA method and ICP-OES. The tap water was obtained from Melbourne's public water supply, and the commercial mineral waters analysed were: Voss Still Water (Norway), Woolworths Mountain Spring Water (Australia) and Icelandic Spring Water (Iceland). All samples were analysed by the standard addition method, involving at least 3 standard additions, and the measurements were performed in triplicate (unless otherwise stated).

3. Results and discussion

3.1. Optimization of the FA system parameters

The optimization range and the initial and optimal values for each of the design and operational parameters of the newly developed FA system investigated in this study are summarized in Table 1 in the order in which the optimization was done. The initial value of a parameter was the value used in the experiments prior to the optimization of this parameter. The compositions of the Streams R1 (0.2 mmol L⁻¹ KMnO₄ and 0.05 mol L⁻¹ NaOH), R2 (0.5% (w/v) NaBH₄ and 0.05 mol L⁻¹ NaOH) and R3 (4 M HCl + 1% (w/v) KI + 0.5% (w/v) ascorbic acid) were selected on the basis of the results obtained in an earlier study involving the determination of arsenic by a gas-diffusion/hydride generation approach (Y. Zhang, Miró, & Kolev, 2015). To simplify the operation of the FA system, Streams R2, R3 and R4 were kept at the same flow rate of 0.12 mL min⁻¹.

Table 1. Optimization of the FA system for the determination of As(V).

Parameter	Range studied	Initial value	Optimal value
Reaction coil length (m)	0 – 3.00	2.50	0.25
Stream R1 flow rate (mL min ⁻¹)	0.06 – 0.46	0.24	0.06
Gas-diffusion membrane	Polypropylene		
	Durapore®		SureVent®
	SureVent®		
	PTFE		
PIM composition (% (w/w))	70 PVC, 30 A336		
	70 PVC, 20 A336, 10 1-TD		70 PVDF-HFP, 30 A336

70 PVDF-HFP, 30 A336

[NaCl] in R4 (mol L ⁻¹)	0.05 – 0.2	0.1	0.1
Stream R5 flow rate (mL min ⁻¹)	0.2 – 3.0	0.2	2.5
Stop-flow time of the acceptor stream of the extraction cell (min)	2 – 30	25	15

A336 – Aliquat 336, 1-TD – 1-tetradecanol

3.1.1. Effect of the reaction coil length, flow rate of Stream R1 and type of the gas-diffusion membrane

As mentioned earlier, the influence of these parameters was studied in an FA system, similar to the one shown in Figure 1, where the extraction cell was replaced with an injection valve with a 500 µL sample loop.

The length of the reaction coil (RC, Figure 1), where Streams R3 and R4 were merged, was varied between 0 m (i.e., no reaction coil) and 3 m. The RC length affected both the efficiency of mixing between the two streams mentioned above and the dispersion of arsenic in the donor stream of the gas-diffusion cell. As expected, a longer RC enhanced arsenic dispersion which offset any increase in the analytical signal due to better mixing between Streams R3 and R4. The highest analytical signal was obtained when the length of the RC was 0.25 m and this length of the RC was used in the subsequent experiments. The percentage of As(V) converted into As(III) under these experimental conditions was calculated as equal to 70%, by comparing the analytical signals for standards containing 1000 µg L⁻¹ of either As(III) or As(V).

As expected, higher analytical signals were recorded when lower flow rates of Stream R1 were used due to the fact that arsine generated in the RC was transferred into a smaller volume of the KMnO₄ acceptor solution of Stream R1. Experiments involving stopping Stream R1

during arsine generation were also conducted but they resulted in unstable baseline due to the transfer of greater and irreproducible amounts of H_2 into the static KMnO_4 solution located in the acceptor channel of the gas-diffusion cell (Figure 1). In addition, no enhancement in the analytical signal was observed. Hence, 0.06 mL min^{-1} was selected as the optimal flow rate of Stream R1 since this was the lowest flow rate that could be reproducibly maintained by Peristaltic pump P1.

Four different hydrophobic microporous membranes (i.e., Durapore®, SureVent®, PTFE, and polypropylene membranes) were compared with respect to their permeability to arsine, which was estimated on the basis of the corresponding analytical signal values. In each case the two channels of the gas-diffusion cell were separated by two membrane layers and a filter paper disc sandwiched between them. When the Durapore® membrane was tested a rapid formation of a brown stain on both membrane surfaces was observed due to manganese dioxide formation, and for this reason this membrane was discarded. The average analytical signals based on 10 replicate measurements of a $1000 \mu\text{g L}^{-1}$ As(V) standard for the remaining three membranes were 0.081 ± 0.004 for the polypropylene membrane, 0.101 ± 0.004 for the PTFE membrane, and 0.102 ± 0.004 for the SureVent® membrane. Although no significant difference between the last two membranes was obtained, the baseline was not very stable when using the PTFE membrane, possibly due to its malleability. SureVent® membrane was selected for further use because it was slightly thicker and more robust than the PTFE membrane and no issues with baseline stability were observed.

3.1.2. Effect of the PIM and the compositions of Stream R4

Fontàs et al. (Fontàs, Vera, Batalla, Kolev, & Anticó, 2013), reported on the successful use of a PIM composed of the base-polymer PVC and the carrier Aliquat 336 for the preconcentration of arsenate in groundwater samples. The optimal composition of this PIM, i.e.,

70% (w/w) PVC and 30% (w/w) Aliquat 336, was determined in a previous study by the same research team (Güell, Anticó, Kolev, Benavente, Salvadó, & Fontàs, 2011). In this and other studies (Güell, Anticó, Kolev, Benavente, Salvadó, & Fontàs, 2011; Güell, Fontàs, Anticó, Salvadó, Crespo, & Velizarov, 2011) 0.1 M NaCl was found to be the most suitable receiving solution for arsenate. The separation of arsenate using an Aliquat 336-based PIM involves the extraction of the HAsO_4^{2-} anion from the sample solution into the PIM, followed by the diffusion of the corresponding adduct of this anion with the quaternary alkylammonium cation of Aliquat 336 (A^+) across the membrane and the back-extraction of HAsO_4^{2-} into the acceptor solution containing NaCl as the stripping reagent (Güell, Anticó, Kolev, Benavente, Salvadó, & Fontàs, 2011). The equilibrium, described by Eq. (1), is shifted to the right (extraction into the PIM) at the sample solution/PIM interface and to the left (back-extraction into the acceptor solution) at the PIM/acceptor solution interface.

The PIM and the receiving solution, mentioned above, were initially used in the newly developed FA system for the on-line preconcentration of As(V). However, the analytical signals obtained in 3 consecutive measurements of a $1000 \mu\text{g L}^{-1}$ As(V) standard (0.09, 0.08, 0.03) were relatively low. The poor repeatability was most likely due to the leaching of the PIM liquid phase consisting of Aliquat 336 into the adjacent aqueous phases. Therefore, other PIM compositions were explored.

One of them was the composition reported by Cho et al. (Cho, Xu, Cattrall, & Kolev, 2011) for the extraction of thiocyanate from weakly alkaline aqueous solutions which consisted of 20% (w/w) Aliquat 336, 10% (w/w) 1-tetradecanol and 70% (w/w) PVC. This study demonstrated that the addition of a modifier (e.g., 1-tetradecanol) of a very low water solubility reduced significantly the leaching of the PIM liquid phase. However, the analytical signal achieved with this PIM composition (i.e., 0.041, 0.017, 0.023), though higher than the one for

the PIM composed of only 70% (w/w) PVC and 30% (w/w) Aliquat, also showed poor repeatability.

O'Bryan et al. (O'Bryan, Cattrall, Truong, Kyratzis, & Kolev, 2016) demonstrated that PVDF-HFP-based PIMs containing 30% (w/w) Aliquat and 70% (w/w) PVDF-HFP exhibited a significantly higher extraction and back-extraction rates for thiocyanate and higher stability compared to PVC-based PIMs containing the same concentration of liquid phase. This PIM provided much higher analytical signal (i.e., 0.173, 0.173, 0.172) and excellent repeatability and therefore was used in the subsequent experiments.

The concentration of NaCl in Stream R4 was varied between 0.05 and 0.20 mol L⁻¹. As expected, the analytical signal increased with increasing the NaCl concentration up to 0.1 mol L⁻¹ after which no further signal enhancement was observed. Therefore, 0.1 mol L⁻¹ was selected as the optimal NaCl concentration in Stream R4.

3.1.3. Effect of the flow rate of Stream R5 and the stop-flow time for Stream R4

It can be expected that the analytical signal will depend heavily on both the flow rate of Stream R5 and the stop-flow time (i.e., duration of the sample flow through the extraction cell) because these two parameters determine the sample volume and its contact time with the PIM. The individual effects of these two parameters on the analytical signal are not independent of each other and for this reason their combined effect was studied and the results are presented in Figure 3. It was observed that, independently of the flow rate of Stream R5, the analytical signal increased rapidly with increasing the stop-flow time up to 15 min and then it started gradually to level off. Also, it was observed that independently of the stop-flow time, the analytical signal increased with increasing the flow rate of Stream R5 up to 2.5 mL min⁻¹ after which it started decreasing. Therefore, 2.5 mL min⁻¹ was selected as the optimal flow rate. The analytical signal did not increase significantly for stop-flow times greater than 15 min (e.g., an

increase in the stop-flow time from 15 to 25 min resulted in only 10% increase in the analytical signal) and this value was selected as the optimal stop-flow time.

Figure 3. Influence of the stop-flow flow time and the flow rate of Stream R5 on the analytical signal for a 500 $\mu\text{g L}^{-1}$ As(V) standard.

3.2. Interference studies

The presence of common anions in natural water (e.g., H_2PO_4^- , Cl^- , NO_3^- , HCO_3^- , and SO_4^{2-}) which can compete with the extraction of arsenate (Fontàs, Vera, Batalla, Kolev, & Anticó, 2013), makes it necessary to investigate their potential interference. No interference effects associated with these anions were expected in the arsine generation, trans-membrane transport and detection steps. Figure 4 shows the normalized analytical signal as a function of the logarithm of the concentration ratio between each one of the anions mentioned above and arsenate. The normalized analytical signal was calculated as a fraction of the analytical signal in the absence of interfering ions. Interference effects were observed only when the concentration of the interfering ions exceeded by 2 orders of magnitude the arsenate

concentration (i.e., 50 $\mu\text{g L}^{-1}$). In the presence of significant interference effects, the standard addition method should be used.

Figure 4. Effect of the concentration of H_2PO_4^- ($\triangle$), NO_3^- ($\bullet$), SO_4^{2-} ($\diamond$), HCO_3^- ($\circ$), Cl^- ($\blacksquare$) on the normalised analytical signal for a $0.67 \mu\text{mol L}^{-1}$ ($50 \mu\text{g L}^{-1}$) As(V) standard.

3.3. Analytical figures of merit

Under optimal conditions (Table 1) the newly developed FA method is characterised by a linear range of $5.0\text{--}65 \mu\text{g L}^{-1}$ As(V) described by the following calibration equation based on 5 different concentrations:

$$A = (8.94 \times 10^{-4} \pm 1.77 \times 10^{-5}) \times C \quad (R^2=0.998) \quad (2)$$

where A is the absorbance and C is the As(V) concentration in $\mu\text{g L}^{-1}$.

The method repeatability, expressed as the relative standard deviation (RSD) of 5 replicate measurements, was calculated as equal to 1.8% for $25 \mu\text{g L}^{-1}$ and 2.8% for $50 \mu\text{g L}^{-1}$ As(V), respectively. The limit of detection (LOD) of $3.0 \mu\text{g L}^{-1}$ was calculated as the analyte

concentration corresponding to an analytical signal equal to the blank signal plus three standard deviations of the blank (Miller & Miller, 2010). The sample solution was propelled for 15 min through the PIM extraction cell while the acceptor solution was stagnant, resulting in a sampling rate of 2.8 h⁻¹.

The newly developed FA method provides better sensitivity for the determination of As(V) than other spectrophotometric FA methods (e.g., 51 µg L⁻¹ (Boonjob, Miró, & Kolev, 2013) and 21 µg L⁻¹ (Y. Zhang, Miró, & Kolev, 2015)) and sensitivity comparable to that provided by FA methods utilizing bulky and expensive atomic optical detectors (e.g., atomic fluorescence detector - 0.61 µg L⁻¹ (Caballo-Lopez & Luque de Castro, 2002) and atomic absorption detector – 0.5 µg L⁻¹ (Y. Zhang & Adeloju, 2008)).

3.4. Analysis of drinking water samples

As mentioned earlier, in most cases arsenic in drinking water consist almost entirely of As(V) (Döker & Yılmaz, 2018; Komorowicz & Barałkiewicz, 2016) and therefore the newly developed method was validated by determining the As(V) concentration in 4 drinking water samples using the standard addition method (Table 2). The standard addition method was used instead of the calibration curve method because of the high concentrations of common anions relative to the As(V) concentration. All standard additions curves were characterised by excellent linearity ($R^2 \geq 0.997$) and the repeatability of the slopes of replicate samples (n=4) expressed as RSD was 5.6%. The As concentration in the spiked samples was also determined by HG-ICP-OES using the calibration curve method. There was no statistically significant difference at the 95% confidence level between the results obtained by both methods (Table 2).

Table 2. As(V) concentration in spiked drinking water samples determined by the newly developed FA method and HG-ICP-OES.

Sample	Spiked As(V) concentration ($\mu\text{g L}^{-1}$)	Measured As(V) concentration $\pm$ SD ($\mu\text{g L}^{-1}$)	
		FA (n=3)	HG-ICP-OES (n=3)
Tap water	6.0	6.5 ± 0.5	6.6 ± 0.4
	10.0	9.1 ± 0.7	10.5 ± 0.8
	15.0	14.0*	15.8 ± 0.9
Voss mineral water	6.00	5.5 ± 0.5	6.4 ± 0.5
	10.0	10.0*	10.5 ± 0.4
Spring mineral water	9.00	8.3 ± 0.8	9.4 ± 0.6
	20.0	21.0*	21.6 ± 0.6
Icelandic mineral water	15.0	13.9 ± 0.9	13.7 ± 0.5
	25.0	22.0*	25.0 ± 2.0

* experiments performed in duplicate

HG-ICP-OES, hydride generation inductively coupled plasma optical emission spectrometry

4. Conclusions

The hydride generation FA system for the determination of arsenate in drinking waters at low $\mu\text{g L}^{-1}$ levels, developed as part of the current study, utilizes for the first time PIM-based on-line extractive separation of arsenate from the sample matrix which is subsequently reduced to arsine, detected spectrophotometrically after its on-line gas-diffusion separation. Under optimal conditions the FA system is characterized by an LOD of $3.0 \mu\text{g L}^{-1}$ and a repeatability, expressed as RSD, of 1.8% (n=5, $25 \mu\text{g L}^{-1}$) and 2.8% (n=5, $50 \mu\text{g L}^{-1}$). Lower limits of

detection could be potentially achieved by using longer stop-flow times for the extraction step, i.e., larger sample volumes. Common anions, such as phosphate, nitrate, sulphate, carbonate, and chloride, were found to interfere in the PIM-based separation process only at concentrations 100 times higher than that of arsenate. The newly developed FA system allowed the accurate determination of arsenate in drinking water spiked with As(V) at the low $\mu\text{g L}^{-1}$ level using the multi-point standard addition method. Since arsenic in most drinking waters is almost entirely composed of arsenate, it can be expected that the FA system, mentioned above, would be applicable for total arsenic determination of drinking water.

Acknowledgements

The financial support of the Spanish government through research project CTM2016-78798-C2-2-P (AEI/FEDER/UE) is acknowledged. R. Vera acknowledges a grant from Spanish Ministerio de Economía y Competitividad ref. BES-2014-068314. The authors would like to thank the Australian Research Council for financial support of this research (Grant LP100100800).

461 **References**

- 462 Almeida, M. I. G. S., Cattrall, R. W., & Kolev, S. D. (2012). Recent trends in extraction and
 463 transport of metal ions using polymer inclusion membranes (PIMs). *Journal of*
 464 *Membrane Science*, 415-416, 9-23.
- 465 Almeida, M. I. G. S., Cattrall, R. W., & Kolev, S. D. (2017). Polymer inclusion membranes
 466 (PIMs) in chemical analysis - A review. *Analytica Chimica Acta*, 987, 1-14.
- 467 Alves, V. N., Neri, T. S., Borges, S. S. O., Carvalho, D. C., & Coelho, N. M. M. (2017).
 468 Determination of inorganic arsenic in natural waters after selective extraction using
 469 *Moringa oleífera* seeds. *Ecological Engineering*, 106, 431-435.
- 470 Bissen, M., & Frimmel, F. H. (2003). Arsenic - A review. Part I: Occurrence, toxicity,
 471 speciation, mobility. *Acta Hydrochimica et Hydrobiologica*, 31(1), 9-18.
- 472 Boonjob, W., Miró, M., & Kolev, S. D. (2013). On-line speciation analysis of inorganic arsenic
 473 in complex environmental aqueous samples by pervaporation sequential injection
 474 analysis. *Talanta*, 117, 8-13.
- 475 Caballo-Lopez, A., & Luque de Castro, M. D. (2002). Hydride generation-pervaporation-
 476 atomic fluorescence detection prior to speciation analysis of arsenic in dirty samples.
 477 *Journal of Analytical Atomic Spectrometry*, 17(10), 1363-1367.
- 478 Cerda, V., & Estela, J. M. (2008). On-line sample pretreatment: dissolution and digestion. In S.
 479 D. Kolev & I. D. McKelvie (Eds.), *Advances in Flow Injection Analysis and Related*
 480 *Techniques* 1 ed., (pp. 129-158). Amsterdam: Elsevier.
- 481 Chen, G., Lai, B., Mao, X., Chen, T., & Chen, M. (2017). Continuous Arsine Detection Using
 482 a Peltier-Effect Cryogenic Trap to Selectively Trap Methylated Arsines. *Analytical*
 483 *Chemistry*, 89(17), 8678-8682.
- 484 Cho, Y., Xu, C., Cattrall, R. W., & Kolev, S. D. (2011). A polymer inclusion membrane for
 485 extracting thiocyanate from weakly alkaline solutions. *Journal of Membrane Science*,
 486 367(1), 85-90.
- 487 Dhar, R. K., Zheng, Y., Rubenstone, J., & Van Geen, A. (2004). A rapid colorimetric method
 488 for measuring arsenic concentrations in groundwater. *Analytica Chimica Acta*, 526(2),
 489 203-209.
- 490 Döker, S., & Yilmaz, M. (2018). Speciation of Arsenic in Spring, Well, and Tap Water by High-
 491 performance Liquid Chromatography–Inductively Coupled Plasma–Mass Spectrometry.
 492 *Analytical Letters*, 51(1-2), 254-264.
- 493 Farrell, J. R., Iles, P. J., & Yuan, Y. J. (1996). Determination of arsenic by hydride generation
 494 gas diffusion flow injection analysis with electrochemical detection. *Analytica Chimica*
 495 *Acta*, 334(1-2), 193-197.
- 496 Fogg, A. G., & Bsebsu, N. K. (1981). Differential-pulse voltammetric determination of
 497 phosphate as molybdovanadophosphate at a glassy carbon electrode and assessment of
 498 eluents for the flow injection voltammetric determination of phosphate, silicate,
 499 arsenate and germanate. *Analyst*, 106(1269), 1288-1295.
- 500 Fontàs, C., Vera, R., Batalla, A., Kolev, S. D., & Anticó, E. (2013). A novel low-cost detection
 501 method for screening of arsenic in groundwater. *Environmental Science and Pollution*
 502 *Research*, 21(20), 11682-11688.
- 503 Frenzel, W., Titzenthaler, F., & Elbel, S. (1994). Selective determination of arsenite by flow
 504 injection spectrophotometry. *Talanta*, 41(11), 1965-1971.
- 505 Güell, R., Anticó, E., Kolev, S. D., Benavente, J., Salvadó, V., & Fontàs, C. (2011).
 506 Development and characterization of polymer inclusion membranes for the separation
 507 and speciation of inorganic As species. *Journal of Membrane Science*, 383(1-2), 88-95.

- Güell, R., Fontàs, C., Anticó, E., Salvadó, V., Crespo, J. G., & Velizarov, S. (2011). Transport and separation of arsenate and arsenite from aqueous media by supported liquid and anion-exchange membranes. *Separation and Purification Technology*, 80(3), 428-434.
- Johnson, D. L., & Pilson, M. E. Q. (1972). Spectrophotometric determination of arsenite, arsenate, and phosphate in natural waters. *Analytica Chimica Acta*, 58(2), 289-299.
- Kamyabi, M. A., & Aghaei, A. (2016). Electromembrane extraction and spectrophotometric determination of As(V) in water samples. *Food Chemistry*, 212, 65-71.
- Komorowicz, I., & Barańkiewicz, D. (2016). Determination of total arsenic and arsenic species in drinking water, surface water, wastewater, and snow from Wielkopolska, Kujawy-Pomerania, and Lower Silesia provinces, Poland. *Environmental Monitoring and Assessment*, 188(9), 504.
- Lomonte, C., Currell, M., Morrison, R. J. S., McKelvie, I. D., & Kolev, S. D. (2007). Sensitive and ultra-fast determination of arsenic(III) by gas-diffusion flow injection analysis with chemiluminescence detection. *Analytica Chimica Acta*, 583(1), 72-77.
- Miller, J. C., & Miller, J. N. (2010). *Statistics and chemometrics for Analytical Chemistry*. Harlow, Essex England: Pearson Education Limited.
- Nagul, E. A., Fontàs, C., McKelvie, I. D., Cattrall, R. W., & Kolev, S. D. (2013). The use of a polymer inclusion membrane for separation and preconcentration of orthophosphate in flow analysis. *Analytica Chimica Acta*, 803, 82-90.
- Narusawa, Y. (1988). Flow-injection spectrophotometric determination of silicate, phosphate and arsenate with on-line column separation. *Analytica Chimica Acta*, 204, 53-62.
- Nghiem, L. D., Mornane, P., Potter, I. D., Perera, J. M., Cattrall, R. W., & Kolev, S. D. (2006). Extraction and transport of metal ions and small organic compounds using polymer inclusion membranes (PIMs). *Journal of Membrane Science*, 281(1), 7-41.
- O'Bryan, Y., Cattrall, R. W., Truong, Y. B., Kyratzis, I. L., & Kolev, S. D. (2016). The use of poly(vinylidene fluoride-co-hexafluoropropylene) for the preparation of polymer inclusion membranes. Application to the extraction of thiocyanate. *Journal of Membrane Science*, 510, 481-488.
- Rupasinghe, T., Cardwell, T. J., Cattrall, R. W., & Kolev, S. D. (2009). Determination of arsenic in industrial samples by pervaporation flow injection with amperometric detection. *Analytica Chimica Acta*, 652(1-2), 266-271.
- Rupasinghe, T., Cardwell, T. J., Cattrall, R. W., Luque de Castro, M. D., & Kolev, S. D. (2001). Pervaporation-flow injection determination of arsenic based on hydride generation and the molybdenum blue reaction. *Analytica Chimica Acta*, 445(2), 229-238.
- Rupasinghe, T., Cardwell, T. J., Cattrall, R. W., Potter, I. D., & Kolev, S. D. (2004). Determination of arsenic by pervaporation-flow injection hydride generation and permanganate spectrophotometric detection. *Analytica Chimica Acta*, 510(2), 225-230.
- Susko, M. L., Bloom, M. S., Neamtiu, I. A., Appleton, A. A., Surdu, S., Pop, C., Fitzgerald, E. F., Anastasiu, D., & Gurzau, E. S. (2017). Low-level arsenic exposure via drinking water consumption and female fecundity - A preliminary investigation. *Environmental Research*, 154(October 2016), 120-125.
- Toda, K., & Ohba, T. (2005). Highly Sensitive Flow Analysis of Trace Level Arsenic in Water Based on Vaporization-collection In-line Preconcentration. *Chemistry Letters*, 34(2), 176-177.
- Valcarcel, M., & Luque de Castro, M. D. (1987). *Flow-Injection Analysis: Principles and Applications*. Chichester: Ellis Horwood.
- Vera, R., Fontas, C., & Antico, E. (2017). Titanium dioxide solid phase for inorganic species adsorption and determination: the case of arsenic. *Environmental Science and Pollution Research*, 24(12), 10939-10948.

- Villaescusa, I., & Bollinger, J.-C. (2008). Arsenic in drinking water: sources, occurrence and health effects (a review). *Reviews in Environmental Science and Bio/Technology*, 7(4), 307-323.
- WHO. (2011). *Guidelines for drinking-water quality, 4th edition* (4th ed.): World Health Organization (WHO)
- Yaftian, M. R., Almeida, M. I. G. S., Cattrall, R. W., & Kolev, S. D. (2018). Flow injection spectrophotometric determination of V(V) involving on-line separation using a poly(vinylidene fluoride-co-hexafluoropropylene)-based polymer inclusion membrane. *Talanta*, 181(November 2017), 385-391.
- Zhang, L. L., Cattrall, R. W., Ashokkumar, M., & Kolev, S. D. (2012). On-line extractive separation in flow injection analysis based on polymer inclusion membranes : A study on membrane stability and approaches for improving membrane permeability. *Talanta*, 97, 382-387.
- Zhang, L. L., Cattrall, R. W., & Kolev, S. D. (2011). The use of a polymer inclusion membrane in flow injection analysis for the on-line separation and determination of zinc. *Talanta*, 84(5), 1278-1283.
- Zhang, Y., & Adeloju, S. B. (2008). Flow injection-hydride generation atomic absorption spectrometric determination of selenium, arsenic and bismuth. *Talanta*, 76(4), 724-730.
- Zhang, Y., Miró, M., & Kolev, S. D. (2015). Hybrid flow system for automatic dynamic fractionation and speciation of inorganic arsenic in environmental solids. *Environmental Science and Technology*, 49(5), 2733-2740.