From: NEURONAL MECHANISMS OF HEARING
Edited by Josef Syka and Lindsay Aitkin
(Plenum Publishing Corporation, 1981)

ELECTRICAL STIMULATION OF THE HUMAN COCHLEA - PSYCHOPHYSICAL AND SPEECH STUDIES

Y. C. Tong and G. M. Clark

Department of Otolaryngology
University of Melbourne
Parkville, 3052, Victoria, Australia

This report describes psychophysical and speech studies conducted on two of our post-lingually deaf patients implanted with multiple-electrode cochlear prostheses. The objectives were to study the nature of the hearing sensations produced by the individual electrodes, and to investigate the feasibility of the transmission of speech information to the higher centres by means of cadences of stimulation using one electrode at a time. Two totally deaf patients (MC1 and MC2) participated in these studies.

In the implant operation, an array of ten electrodes were inserted through an opening in the round window membrane for a distance of 15mm around the scala tympani in MC1, and 19.5 mm in MC2. The electrodes were numbered from 0 to 9 in the apical to basal direction. These electrodes, spaced 1.5 mm apart, were driven by a receiver-stimulator, which was positioned in a bed created in the mastoid bone. Residual auditory nerve fibres were activated by biphasic current pulses with each phase fixed at approximately 180 μ s. Fifteen current levels ranging from 67 μ A to 1 mA could be assigned in 67 μ A steps, and a maximum pulse rate of 1000 pps was possible (Clark et al., 1979; Tong et al., 1979).

PSYCHOPHYSICAL STUDIES

A series of psychophysical studies (Tong et al., 1979) was conducted on the two patients after the implant operation. Both time-invariant and time-varying signals were used. Time-invariant signals are those whose signal parameter values do not vary over the duration of the signal. Time-varying signals, on the other hand, are characterized by a linear variation from an initial to

a final parameter value over a specified duration. Furthermore, signals with time-varying electrode position are called single-electrode trajectories.

The results for time-invariant signals are summarized as follows:

- (1) Threshold currents for 300 ms pulse trains at 150 pps ranged from 130 μ A to 600 μ A for MC1, and 200 μ A to 470 μ A for MC2;
- (2) The loudness growth as a function of stimulus (current) level for 300 ms pulse trains was much steeper than that for acoustic stimulation;
- (3) For pulse rates below 200 pps, the perceived pitch for 300 ms pulse trains increased with pulse rate. Above 200 pps, however, the increase in pitch with pulse rate was less pronounced;
- (4) For a fixed pulse rate of stimulation, the perceived pitch varied from low to high and the sharpness from dull to sharp for 300 ms pulse trains at single-electrodes ordered in the apical to basal direction;
- (5) Vowel labels could be assigned to the sensations produced by single-electrodes at a fixed pulse rate (200 ms pulse trains);
- (6) Relative difference limens of 7 to 12% were observed at pulse rates below 250 pps for 300 ms pulse trains;
- (7) At a fixed pulse rate of stimulation, 200 ms pulse trains on different electrodes were readily discriminated. Furthermore, the signals at two adjacent electrodes were rarely confused within a restrictive range of pulse rates - tested from 105 to 165 pps and from 180 to 240 pps.

The results from time-varying signals are summarized as follows:

- (1) Relative difference limens of 9 to 13% were observed for pulse rate variation over 300 ms for pulse rates below 200 pps;
- (2) The discriminability of signals with time-varying pulse rate deteriorated with decreases in the duration of the pulse rate variation. Pulse rate variations with a variable initial pulse rate (150, 180, 210 or 240 pps) and a fixed final pulse rate (150 pps) were tested for three durations: 25, 50 and 100 ms. The discrimination performance was the poorest at 25 ms;
- (3) 100 ms single-electrode trajectories differing in the direction (apicalward, basalward) of the electrode variation were readily discriminated;
- (4) The discriminability of single-electrode trajectories was independent of the duration of the trajectories. Trajectories with a variable initial electrode position (electrodes 1, 2, 3 or 4) and fixed final electrode position (electrode 1) were tested for three durations: 25, 50 and 100 ms. The same performance was observed for each duration;
- (5) Phenomenological reports from MC1 indicated that the hearing sensations produced by the electrode trajectories were similar to those produced by acoustic speech signals characterized by second

formant frequency transitions to which phonetic labels may be assigned. The hearing sensations produced by time-varying pulse rate, on the other hand, were similar to those produced by speech signals with a variation in the fundamental (voice) frequency.

SPEECH STUDIES

In the light of the psychophysical results, a speech processing strategy, which converts the frequency of the spectral emphasis in the mid-frequency range of speech signal to single-electrode position, and the fundamental (voicing) frequency to pulse rates below 250 pps, was proposed. With this strategy, the auditory system will discriminate the electrode shifts required to convey rapidly changing segmental speech information. The suprasegmental information contained in the fundamental frequency of a speech signal, on the other hand, is encoded in the pulse rate which is more suited to slowly time-varying signals. Although the discrimination performances for electrode position and pulse rate are different, their interaction also needs to be considered in speech processing strategies. In this regard, the psychophysical results (see item 7 for time-invariant signals) indicated that, for speech signals involving a single voice with a limited range of fundamental frequency (and therefore a limited range of pulse rates below 250 pps), the information for electrode position is not likely to be confused with that for pulse rate. Furthermore, the pitch matching experiments reported in Bilger (1977) and Eddington et al., (1978) for implant patients with residual hearing in the unimplanted ear indicated that, with appropriate psychophysical procedures, it may be possible to separately identify these two types of information for pulse rates below 250 pps. It is also possible that, in the speech mode, the rapidly time-varying "electrode cues" for segmental information are readily discernible from the slowly time-varying "pulse rate cues" for suprasegmental information because of the difference in the time scale of processing. These suggestions are supported by the phenomenological reports of patient MCI described in the last section. As far as the encoding of the intensity of the speech signal is concerned, the current level of the electrical stimulus in conjunction with an appropriate compression scheme may be used.

The feasibility of this proposed speech processing strategy in the transmission of speech information to the higher centres was studied by means of a laboratory-based speech-processor. In the real-time speech processor, speech parameters were estimated and converted to electrical stimulation parameters every 10 ms, and only one electrode was activated within a 10 ms time frame. The speech parameters were the second formant frequency (F2) and its amplitude (A2), the fundamental frequency (F0), and the voicing amplitude (A0). For a given F2 estimate, the single-electrode to

be stimulated in the present time frame was selected from a pre-determined F2-to-electrode conversion table based on psychophysical results. Similarly, the current level was determined from an A2-to-current table. Furthermore, an energy threshold detector for A0 was used to determine whether voicing was present or not. If voicing was present, the pulse rate (250 pps) on the selected electrode was made proportional to F0. On the other hand, if voicing was absent a constant low pulse rate was used as it produced a sensation described as "rough" which could be associated with that of "noise" previously experienced by the patients when they had hearing.

Speech comprehension tests were conducted under three conditions: electrical stimulation alone (EO), vision alone (VO), and combined electrical stimulation and vision (EV). The first test was a confusion study (Miller and Nicely, 1955) using twelve consonants (/b/, /p/, /m/, /f/, /v/, /s/, /z/, /n/, /t/, /d/, /k/, /g/,) in a vowel-consonant-vowel (VCV) context, with the vowel fixed as a /a/ as in father. The two patients received ten randomized presentations of the 12 VCV syllables in each condition. Scoring was based on the number of consonants correctly identified. In the second test, everyday sentence lists prepared at the Central Institute for the Deaf were presented. Scoring was based on the number of key words correctly identified. The test materials were presented at 80 dB SPL by a female speaker.

The mean scores for the two patients in the first test were 38% (EO), 37% (VO) and 65% (EV). The mean scores obtained in the second test were 11% (EO), 24% (VO) and 83% (EV). These results show that the scores are highest for the EV condition, indicating the contribution of electrical stimulation of speech communication. It should also be noted that the sentence scores in the EV condition were at a level where they correlate with a satisfactory performance in understanding connected speech (Giolas, 1966). In addition to these results, the consonant confusion data obtained in the first test were also analyzed in terms of the information transmission for the articulatory features described in Miller and Nicely (1955). This analysis showed that a significant amount of information about voicing, nasality, duration, affrication and place of articulation was transmitted in the EO condition (Tong and Clark, 1980). The results of this analysis confirmed that useful speech information can be transmitted to the higher centres of the auditory system by means of the proposed strategy.

CONCLUSION

The results obtained in these studies indicate that it is possible to convey useful speech information to the higher centres of the auditory system by cadences of electrical stimulation at

different single-electrodes in the scala tympani. Furthermore, the hearing sensations produced by the electrical stimuli could be linked with those previously experienced by the patients. These results are informative because they confirmed the usefulness of a speech processing strategy employing both the "place" information relating to electrode position and the "period" information relating to pulse rate, and they formed the basis for future research. As it stands, however, the proposed speech strategy when used in conjunction with lipreading does improve significantly the patients' ability in speech communication.

REFERENCES

- Bilger, R. C., 1977, Evaluation of subjects presently fitted with auditory prosthesis, Ann. Otol., Supp. 38.
- Clark, G. M., Pyman, B. C., and Bailey, Q. R., 1979, The surgery for multiple-electrode cochlear implantations, J. Laryngol. Otol., 93: 215-223.
- Eddington, D. K., Dobelle, W. H., Brackmann, D. E., Mladejovsky, M. G., and Parkin, J. L., 1978, Auditory prostheses research with multiple channel intracochlear stimulation in man, Ann. Otol., Suppl. 51.
- Giolas, T. G., 1966, Comparative intelligibility scores of sentence test and continuous discourse, J. Aud. Res., 6:31-38.
- Miller, G. A., and Nicely, P. E., 1955, An analysis of perceptual confusions among some English consonants, J. Acoust. Soc. Am., 27: 338-352.
- Tong, Y. C., Black, R. C., Clark, G. M., Forster, I. C., Millar, J. B., O'Loughlin, B. J., and Patrick, J. F., 1979, A preliminary report on a multiple-channel cochlear implant operation, J. Laryngol. Otol., 93: 679-695.
- Tong, Y. C. and Clark, G. M., 1980, Speech comprehension with multiple channel electrical stimulation of human auditory nerve fibres. J. Physiol. and Pharmacol. Soc. Aust., (in press).

Minerva Access is the Institutional Repository of The University of Melbourne

Author/s:

Clark, Graeme M.

Title:

Electrical stimulation of the human cochlea: psychophysical and speech studies

Date:

1981

Citation:

Tong, Y. C., & Clark, G. M. (1981). Electrical stimulation of the human cochlea: psychophysical and speech studies. *Neuronal Mechanism of Hearing*, 411-415.

Persistent Link:

<http://hdl.handle.net/11343/27163>