Article Type: Point Counterpoint

POINT/COUNTERPOINT

Suggestions for topics suitable for these Point/Counterpoint debates should be addressed to Habib Zaidi, Geneva University Hospital, Geneva, Switzerland: habib.zaidi@hcuge.ch; Jing Cai, The Hong Kong Polytechnic University, Hong Kong: jing.cai@polyu.edu.hk; and/or Gerald White, Colorado Associates in Medical Physics: gerald.white@mindspring.com. Persons participating in Point/Counterpoint discussions are selected for their knowledge and communicative skill. Their positions for or against a proposition may or may not reflect their personal opinions or the positions of their employers.

In the future, emission-guided radiation therapy will play a critical role in clinical radiation oncology

Paul Keall, Ph.D.

*University of Sydney, ACRF Image X Institute, Sydney Medical School, Australia
(Tel. +61-28627-1159; Email: paul.keall@sydney.edu.au)*

Tomas Kron, Ph.D.

*Sir Peter MacCallum Cancer Institute, University of Melbourne, Melbourne, Victoria, Australia
(Tel. +61-38559-6925; Email: Tomas.Kron@petermac.org)*

Habib Zaidi, Ph.D., Moderator

OVERVIEW

Emission-guided radiation therapy entails the use of radiation emitted by molecular imaging probes (SPECT for PET) to guide delivery directly, thus avoiding the need to wait for long acquisition times to collect sufficient count statistics. The potential of biological targeting in

This is the author manuscript accepted for publication and has undergone full peer review but has not been through the copyediting, typesetting, pagination and proofreading process, which may lead to differences between this version and the [Version of Record](#). Please cite this article as [doi: 10.1002/MP.13408](https://doi.org/10.1002/MP.13408)

radiation treatment planning, taking advantage of advanced molecular imaging techniques in the form of hybrid SPECT/CT, PET/CT or PET/MRI systems, is well established. The last three decades have witnessed abundant contributions indicative of the advantages of multiparametric molecular imaging in this framework. Ultimately, molecular imaging-guided radiation therapy planning holds the promise of improved delineation of target volumes even in the presence of significant motion. Yet, despite significant advances to date, challenges remain limiting exploitation of the potential of this technology in the clinic to academic facilities with advanced technical support. Fortunately, these limitations are not preventing novel exciting developments to emerge with the aim to further boost the incorporation of multimodality imaging in the process of radiation therapy planning for various indications. In this regard, the conceptual basis of using on-board PET to guide the radiation treatment process has been around since 2010. A small company based in the US (RefleXion Medical, CA) pioneered the development of a device enabling commercial implementation of one approach to biologically-guided radiation therapy. The product was displayed for the first time at the last ASTRO 2018 annual meeting. While some think that this is relevant technology that should soon be adopted for use in the clinic, others think that the concept is still in its infancy and not mature enough for clinical adoption. This is the topic addressed in this month's Point/Counterpoint debate.

Arguing for the proposition is Paul Keall, PhD. Dr. Keall is a National Health & Medical Research Council Senior Principal Research Fellow and Director of the ACRF Image X Institute at the University of Sydney. His team of scientists and clinical partners develop and apply novel cancer imaging and targeted radiotherapy methods. One of Dr. Keall's research interests is motion management in radiotherapy. He chaired the AAPM TG264 - *Safe Clinical Implementation of MLC tracking* and is currently chairing the update to TG76 *The management of respiratory motion in radiation oncology* (TG324). Dr. Keall has published over 300 papers and has 18 awarded patents. He is a Fellow of AAPM and a member of the *Medical Physics* Editorial Board.

Arguing against the Proposition is Tomas Kron, PhD. Dr. Kron was born and educated in Germany. After completing his PhD, he migrated to Australia in 1989 where he commenced his career in radiotherapy physics. From 2001 to 2005, he lived in Canada where he worked at the

60 London Regional Cancer Centre and commissioned one of the first tomotherapy units. In 2005,
61 Tomas became principal research physicist at Peter MacCallum Cancer Centre in Melbourne,
62 Australia where he now is Director of Physical Sciences. Tomas holds academic honorary
63 appointments on a professorial level at Wollongong, RMIT and Monash Universities. He has an
64 interest in dosimetry of ionizing radiation, image guided radiotherapy, and clinical trials as a tool
65 for introducing new technology. This is demonstrated by more than 80 invited conference
66 presentations and 260 papers in refereed journals. In 2014 he was awarded an Order of Australia
67 Medal (OAM) for services to medicine, research and education.

68
69 **FOR THE PROPOSITION: Paul Keall, Ph.D.**

70 **Opening Statement**

71 It is an exciting time to be working in radiation therapy. Cancer patient survival is increasing
72 with stereotactic ablative body radiotherapy. One-to-two year survival benefits are demonstrated
73 for patients treated with limited metastatic disease. The use of radiotherapy for new conditions,
74 particularly cardiac disease, shows extremely promising results.¹ The interaction of radiotherapy
75 and immunotherapy presents major clinical opportunities. The adoption of MRI-Linacs opens
76 new capabilities for treatment conformality. All these innovations are underpinned by major
77 advances in imaging and targeted radiotherapy technology.

78 Emission-guided radiotherapy is another emerging technology that will also play a critical
79 role in clinical radiation oncology. In this context emission-guided radiotherapy is defined as in-
80 treatment room PET (or SPECT) systems. PET is arguably the most sensitive molecular imaging
81 modality. In addition to the widely used FDG radiotracer that measures tumor metabolic activity,
82 there are agents that image tumor hypoxia, proliferation, antigen levels and many other
83 biological processes. It is a natural evolution to marry the anatomic and physiologic targeting
84 capability of PET with rapidly advancing radiotherapy technology to yield emission-guided
85 radiotherapy. It is currently the norm rather than the exception to treat tumors as functionally
86 homogeneous – we prescribe the same dose to the entire planning target volume. With all the
87 imaging information at hand, it seems medieval for uniform prescriptions to heterogeneous
88 tumors to be the standard of practice. Emission-guided radiotherapy will give the treatment team
89 both the information and the confidence to target radiation most where it is needed most, and

least where it is needed least, achieving the combined goals of improved tumor control, reduced metastatic potential and normal tissue dose reduction.

Emission-guided radiotherapy has many forms – I will highlight three here.

(1) *PET-guided external beam radiotherapy*. A variety of systems have been used to determine the tumor position in real-time during radiotherapy. However, using PET imaging for this task enables both anatomic and physiologic targeting opportunities. The feasibility of PET-guided external beam radiotherapy has been demonstrated by several groups, e.g.²⁻⁴. One approach to PET-guided external beam radiotherapy is being commercially pioneered by Reflexion Medical who raised \$100M in 2018 alone to develop a product and have assembled a consortium of early adopters. PET-guided external beam radiotherapy becomes increasingly useful as the number of cancer targets increases, e.g. for treating metastatic disease. Multiple targets also pose a motion management challenge as different targets can be moving independently of each other – PET imaging will naturally be able to show the motion of individual lesions. Another intriguing potential of PET-guided external beam radiotherapy is to tailor the total dose delivered based on the quantitative biological changes during treatment as opposed to having the same dose and fraction size for all patients.

(2) *PET-guided patient dose measurements during particle therapy*. Particle therapy is promising, but to date randomized trial evidence has not demonstrated its benefit.⁵ One obvious hypothesis is that the motion and range uncertainties during treatment mask the benefit observed from planning studies. To unlock the clinical potential of particle therapy, PET is a feasible method for the non-invasive monitoring of particle therapy treatments.^{6,7} PET monitoring, coupled with beam adaptation, will reduce motion and range uncertainties. PET-guided particle therapy is being developed and investigated in many centers to enable the routine use of emission-guided radiotherapy during particle therapy.

(3) *Emission-guided radionuclide therapy*. Using a combined SPECT and PET approach, patients for liver cancer radioembolization can be imaged pre-treatment with e.g. ^{99m}Tc to predict the therapeutic dose and measured post-treatment, e.g. from the small proportion of positrons emitted from ⁹⁰Y resin microspheres.⁸ In this context, pre-treatment imaging can be used for planning, and post-treatment imaging can be used to evaluate if adjuvant treatment options are needed.

The three emission-guided radiotherapy examples given here are a subset of the ongoing development in this broader space linking emission-guidance with radiation therapy. The future technological developments, clinical applications and concomitant improvements in emission-guided radiotherapy patient outcomes are exciting. Watch this space!

AGAINST THE PROPOSITION: Tomas Kron, Ph.D.

Opening Statement

There is no doubt that emission imaging (or nuclear medicine/molecular imaging) has made a huge difference to the practice of radiation oncology.^{9,10} PET scanning and to some degree SPECT have had a tremendous impact on staging and patient selection for curative and palliative radiotherapy. There is no doubt that emission imaging impacts the way we delineate and characterize targets (and even critical structures) profoundly.^{11,12} ‘Anatomy follows function’ and therefore tumor response to therapy can be predicted earlier with emission imaging than with anatomic imaging. But as image guidance aims to help us directing radiation to a physical target every time we treat the patient¹³, do we really want this ‘glimpse into the future of anatomy’, in particular towards the end of treatment after having inflicted several logs of cell kill?

In fact, there are many fundamental problems when trying to take emission imaging into the treatment room for image guidance:

1. We talk of PET but we mean PET/CT.¹⁰ Without anatomical correlation and attenuation correction, emission-based imaging technologies have limited value in target definition let alone image guidance. As such, it is not sufficient to integrate emission detectors in the treatment unit but one also has to combine it with CT or MRI – the technical complexity and cost make this not a very convincing option.
2. Related to the previous point is the time required for imaging. Unless very high activities are administered, the efficiency of Linac-based emission detectors will be smaller than dedicated diagnostic systems and as such, acquisition time will be long – too long for motion management or repeat acquisitions during delivery.
3. FDG-PET is the most likely candidate for image guidance. However, the very success of radiotherapy in achieving several logs of cell kill already half way through the treatment course may wipe out the very signal to be used for image guidance just when we most need it.

The inflammation which is a result of radiation injury complicates matters as it is non-specific to the target.¹⁴

3. Many imaging protocols require the patient to wait a well-defined time after a tracer is injected to ensure the desired activity distribution is established. Can we expect patients to come hours before treatment and can we be sure that there are no delays in bringing the patient into the treatment room?
4. Whatever the approach, the spatial resolution of emission imaging is significantly worse than anatomical imaging.¹⁵ The (sub) millimeter accuracy desirable for modern radiotherapy approaches, such as SBRT, will always rely on anatomical imaging.
5. Finally, operators would require a completely new set of skills interpreting and quickly making decisions based on PET or SPECT images, let alone perform the injections and supervise the required waiting period.

In summary, emission-guided target identification is the future but emission-guided image guidance is less likely to be part of it. Let's not let our desire for the most fancy imaging technology use a laser where a torch is needed.

Rebuttal: Paul Keall, Ph.D.

We know that tumors and normal tissues are functionally heterogeneous. It seems backward that in the 21st century we are still treating tumors and healthy tissue as uniform structures. Emission-guided radiation therapy gives us the opportunity to move towards preferential targeting and sparing of the tissue that matters, accounting for changes with time and personalizing the dose and fractionation schedule based on individual responses.

In the opening statement against the proposition, six of the challenges that emission-guided radiation therapy faces on the journey to play a critical role in clinical radiation oncology were identified. To briefly address these six challenges:

1. *Need for attenuation correction.* As volumetric pre-treatment anatomic guidance is the current standard, pre-treatment anatomic guidance should naturally continue for emission-guided radiation therapy treatments. The pre-treatment image can be used for attenuation correction.
2. *Imaging time.* We currently accept the additional time for pre-treatment imaging as worthwhile because patient outcomes are a higher priority than imaging times. With the

181 patient interests first and foremost, time is not the priority, outcomes are. With
182 immunotherapy costing \$100,000 per patient per year in some cases, the time and cost of
183 radiotherapy is negligible when viewed in the broad lens of cancer care costs.

184 3. *Signal/noise reduction as tumor cells decrease and inflammation increases during a course of*
185 *therapy.* For single and few-fraction SABR treatments we may not expect significant changes
186 during the treatment. For prolonged fractionated treatments, the temporal changes in the
187 tumor microenvironment represent an opportunity for understanding and optimizing the
188 normal and cancer cellular kinetics on an individual basis.

189 4. *Increased time for tracer uptake.* We expect a large proportion of cancer patients to undergo
190 brutal cycles of chemotherapy to improve their survival chances. Increasing the time spent in
191 a radiotherapy department with a non-toxic radiotracer in the veins – as opposed to large
192 doses of cytotoxic drugs – is a small inconvenience when weighted against the life expectancy
193 of the patient.

194 5. *Spatial resolution.* The physiologic information from emission-guided radiation therapy
195 augments the anatomic information from treatment planning, pre-treatment and real-time
196 imaging. The additional functional information may be lower resolution than anatomic
197 guidance methods, however the additional information enables a new paradigm – selectively
198 boosting the parts of the tumor most likely to be treatment resistant and/or cause metastases.

199 6. *Educational needs.* Any new technology requires reskilling and reeducation of staff. With
200 other technology, such as knowledge-based planning, having dramatic implications for the
201 workforce, we should embrace the opportunity to learn new skills and be the experts in the
202 growing field of emission-guided radiation therapy.

203
204 To close, let us be bold. Dr von Eschenbach, when he was Director of the US National Cancer
205 Institute, stated that we are at a strategic inflection in oncology, a point of unprecedented growth
206 in knowledge, technology, and resources. The integration of these three key sectors provides an
207 opportunity for exponential progress towards the goal of eliminating suffering and death due to
208 cancer.¹⁶ Emission-guided radiation therapy combines these three sectors opening opportunities
209 for anatomic and functional targeting across cancer disease sites and disease stages. For all of the
210 reasons mentioned, emission-guided radiation therapy will play a critical role in clinical radiation
211 oncology.

Rebuttal: Tomas Kron, Ph.D.

In his opening statement Paul Keall provides an excellent summary of the excitement that surrounds modern radiotherapy and there is no doubt that functional imaging contributes significantly to this. However, of the three examples for emission guided radiotherapy given (External Beam, Particle and Radionuclide Therapy) only the latter is a convincing application of what can be termed ‘emission-guided radiation therapy’.

It is claimed that emission guided RT will give ‘the confidence to target radiation most where it is needed most, and least where it is needed least’. However, one of the limitations of PET is that one can only explore one metabolic pathway at a time. For treatment planning it may be possible to assess hypoxia, metabolic activity, proliferation and possibly EGFR status (or other factors) in sequence¹⁷ but only one tracer can be used at treatment time. Can we be certain as to what indicates where dose is needed throughout the whole treatment course?

Given the limited spatial resolution and the link of sensitivity to radionuclide dose it is also unlikely that motion management will yield anything but an internal target volume (ITV). Clearly a step back from the exciting developments in motion adaptive radiotherapy that have just become available.¹⁸

On the other hand it is granted that radionuclide therapy benefits significantly from emission imaging if personalized dose calculations are to be performed. However, this is either based on imaging with a test dose days prior to treatment or on assessment of the very activity distribution used for treatment (and of course not modifiable at this time).¹⁹ None of this would qualify as ‘emission guidance’ in the traditional sense.

It has been nearly 20 years since Clifton Ling’s seminal paper on biological targeting.²⁰ While cited and discussed by many, progress has been slow and clinical trials are just getting under way. Most of them are phase I/II studies with practice changing phase III trials in development.²¹ We are just gaining confidence in PET/SPECT based treatment planning. It looks like a lot of work is still needed before emission-guided radiation therapy can even be considered as a research tool for clinical radiation oncology.

Conflicts of Interest

Dr. Keall and Dr. Kron have no relevant conflicts of interest.

References

1. Cuculich PS, Schill MR, Kashani R, et al. Noninvasive cardiac radiation for ablation of ventricular tachycardia. *N Engl J Med*. 2017;377(24):2325-2336.
2. Fan Q, Nanduri A, Mazin S, Zhu L. Emission guided radiation therapy for lung and prostate cancers: A feasibility study on a digital patient. *Med Phys*. 2012;39(11):7140-7152.
3. Mazin SR, Nanduri AS. Emission-guided radiation therapy: Biologic targeting and adaptive treatment. *J Am Coll Radiol*. 2010;7(12):989-990.
4. Yang J, Yamamoto T, Mazin SR, Graves EE, Keall PJ. The potential of positron emission tomography for intratreatment dynamic lung tumor tracking: A phantom study . *Med Phys*. 2014;41(2):021718.
5. Liao Z, Lee JJ, Komaki R, et al. Bayesian adaptive randomization trial of passive scattering proton therapy and intensity-modulated photon radiotherapy for locally advanced non-small-cell lung cancer. *J Clin Oncol*. 2018;36(18):1813-1822.
6. Parodi K, Enghardt W, Haberer T. In-beam PET measurements of β^+ radioactivity induced by proton beams. *Phys Med Biol*. 2001;47(1):21.
7. Enghardt W, Crespo P, Fiedler F, et al. Charged hadron tumour therapy monitoring by means of PET. *Nucl Instrum Methods Phys Res A*. 2004;525(1-2):284-288.
8. Willowson KP, Hayes AR, Chan DL, et al. Clinical and imaging-based prognostic factors in radioembolisation of liver metastases from colorectal cancer: a retrospective exploratory analysis. *EJNMMI Res*. 2017;7(1):46.
9. Bomanji JB, Costa DC, Ell PJ. Clinical role of positron emission tomography in oncology. *Lancet Oncol*. 2001;2(3):157-164.
10. Pan T, Mawlawi O. PET/CT in radiation oncology. *Med Phys*. 2008;35(11):4955-4966.
11. Verma V, Choi JJ, Sawant A, et al. Use of PET and other functional imaging to guide target delineation in radiation oncology. *Semin Radiat Oncol*. 2018;28(3):171-177.

12. Siva S, Thomas R, Callahan J, et al. High-resolution pulmonary ventilation and perfusion PET/CT allows for functionally adapted intensity modulated radiotherapy in lung cancer. *Radiother Oncol.* 2015;115(2):157-162.
13. Dawson LA, Sharpe MB. Image-guided radiotherapy: rationale, benefits, and limitations. *Lancet Oncol.* 2006;7(10):848-858.
14. Jahangiri P, Pournazari K, Torigian DA, et al. A prospective study of the feasibility of FDG-PET/CT imaging to quantify radiation-induced lung inflammation in locally advanced non-small cell lung cancer patients receiving proton or photon radiotherapy. *Eur J Nucl Med Mol Imaging.* 2019, in press doi 10.1007/s00259-018-4154-5
15. Moses WW. Fundamental limits of spatial resolution in PET. *Nucl Instrum Methods Phys Res A.* 2011;648 Supplement 1:S236-S240.
16. von Eschenbach AC. NCI sets goal of eliminating suffering and death due to cancer by 2015. *J Natl Med Assoc.* 2003;95(7):637.
17. Heukelom J, Hamming O, Bartelink H, et al. Adaptive and innovative Radiation Treatment FOR improving Cancer treatment outcome (ARTFORCE); a randomized controlled phase II trial for individualized treatment of head and neck cancer. *BMC Cancer.* 2013;13:84.
18. Keall PJ, Nguyen DT, O'Brien R, et al. The first clinical implementation of real-time image-guided adaptive radiotherapy using a standard linear accelerator. *Radiother Oncol.* 2018;127(1):6-11.
19. Stokke C, Gabina PM, Solny P, et al. Dosimetry-based treatment planning for molecular radiotherapy: a summary of the 2017 report from the Internal Dosimetry Task Force. *EJNMMI Phys.* 2017;4(1):27.
20. Ling CC, Humm J, Larson S, et al. Towards multidimensional radiotherapy (MD-CRT): biological imaging and biological conformality. *Int J Radiat Oncol Biol Phys.* 2000;47(3):551-560.
21. Shi X, Meng X, Sun X, Xing L, Yu J. PET/CT imaging-guided dose painting in radiation therapy. *Cancer Lett.* 2014;355(2):169-175.

Minerva Access is the Institutional Repository of The University of Melbourne

Author/s:

Keall, P;Kron, T;Zaidi, H

Title:

In the future, emission-guided radiation therapy will play a critical role in clinical radiation oncology

Date:

2019-04-01

Citation:

Keall, P., Kron, T. & Zaidi, H. (2019). In the future, emission-guided radiation therapy will play a critical role in clinical radiation oncology. MEDICAL PHYSICS, 46 (4), pp.1519-1522. <https://doi.org/10.1002/mp.13408>.

Persistent Link:

<http://hdl.handle.net/11343/285505>