Clinical integration of automated processing for brain quantitative susceptibility mapping: multi-site reproducibility and single-site robustness.

Pascal Spincemaille^{1*}, Zhe Liu^{1,2}, Shun Zhang^{1,3}, Ilhami Kovanlikaya¹, Matteo Ippoliti⁴, Marcus Makowski⁴, Richard Watts⁵, Ludovic de Rochefort⁶, Vijay Venkatraman⁷, Patricia Desmond⁷, Mathieu D. Santin⁸, Stephane Lehericy^{8,9}, Brian H. Kopell^{10,11,12,13}, Patrice Périn¹⁴, Yi Wang^{1,2}

¹ Department of Radiology, Weill Medical College of Cornell University, New York, NY, USA

² Meinig School of Biomedical Engineering, Cornell University, Ithaca, NY, USA

³ Department of Radiology, Tongji Hospital of Tongji Medical College of Huazhong University of Science and Technology, Wuhan, Hubei, China

⁴ Department of Radiology, Charité Universitätsmedizin Berlin, Berlin, Germany

⁵ Department of Psychology, Yale University, New Haven, CT, USA

⁶ Center for Magnetic Resonance in Biology and Medicine, Aix Marseille Université, Marseille, France

⁷ Department of Medicine and Radiology, University of Melbourne, Royal Melbourne Hospital, Melbourne, Australia

⁸ Inserm U 1127, CNRS UMR 7225, Centre for NeuroImaging Research, ICM (Brain & Spine Institute), Sorbonne University, Paris, France

⁹ Neuroradiology, Hôpital Pitié-Salpêtrière, Paris, France

¹⁰ Division of Movement Disorders, Department of Neurology, Icahn School of Medicine at Mount Sinai, New York, NY, USA.

¹¹ Department of Neurosurgery, Icahn School of Medicine at Mount Sinai, New York, NY, USA.

¹² Department of Psychiatry, Icahn School of Medicine at Mount Sinai, New York, NY, USA.

¹³ Department of Neuroscience, Icahn School of Medicine at Mount Sinai, New York, NY, USA.

¹⁴ Toulouse NeuroImaging Center, Université de Toulouse, Inserm, Toulouse, France

Running title: Clinical integration of automated brain quantitative susceptibility mapping

Keywords: Quantitative Susceptibility Mapping; Software; Clinic;

***Corresponding author:** Pascal Spincemaille, Ph.D., Department of Radiology, 515 East 71st St Suite S101, New York, NY, 10021, e-mail: pas2018@med.cornell.edu, tel: +1 646 962 2630

Acknowledgments and Disclosure: This work was supported in part by the NIH R01NS090464, R01NS095562, R01CA181566, R21EB024366, and S10OD021782. Ludovic de Rochefort, Yi Wang, and

This is the author manuscript accepted for publication and has undergone full peer review but has not been through the copyediting, typesetting, pagination and proofreading process, which may lead to differences between this version and the [Version of Record](#). Please cite this article as [doi: 10.1111/jon.12658](https://doi.org/10.1111/jon.12658).

This article is protected by copyright. All rights reserved.

Pascal Spincemaille are inventors on QSM related patents issued to Cornell University. Yi Wang and Pascal Spincemaille hold equity in Medimagetric LLC. The remaining authors declare that they have no disclosures.

Abstract

Background and Purpose: Quantitative susceptibility mapping (QSM) of the brain has become highly reproducible and has applications in an expanding array of diseases. To translate QSM from bench to bedside, it is important to automate its reconstruction immediately after data acquisition. In this work, a server system that automatically reconstructs QSM and exchange images with the scanner using the DICOM standard is demonstrated using a multi-site, multi-vendor reproducibility study and a large, single-site, multi-scanner image quality review study in a clinical environment.

Methods: A single healthy subject was scanned with a 3D multi-echo gradient echo sequence at 9 sites around the world using scanners from three manufacturers. A high resolution (HiRes, $0.5 \times 0.5 \times 1 \text{ mm}^3$ reconstructed) and standard resolution (StdRes, $0.5 \times 0.5 \times 3 \text{ mm}^3$) protocol was performed. ROI analysis of various white matter and gray matter regions was performed to investigate reproducibility across sites. At one institution, a retrospective multi-scanner image quality review was carried out of all clinical QSM images acquired consecutively in one month.

Results: Reconstruction times using a GPU were $29 \pm 22 \text{ s}$ (StdRes), and $55 \pm 39 \text{ s}$ (HiRes). ROI standard deviation across sites was below 24ppb (StdRes) and 17ppb (HiRes). Correlations between ROI averages across sites were on average 0.92 (StdRes) and 0.96 (HiRes). Image quality review of 873 consecutive patients revealed diagnostic or excellent image quality in 96% of patients.

Conclusion: Online QSM reconstruction for a variety of sites and scanner platforms with low cross-site ROI standard deviation is demonstrated. Image quality review revealed diagnostic or excellent image quality in 96% of 873 patients.

Introduction

Quantitative susceptibility mapping (QSM) enables the quantitative measurement of tissue magnetic susceptibility, which is a measure of how tissue responds to an external magnetic field.^{1,2} Quantitative susceptibility mapping (QSM) has been used increasingly in clinical brain studies,^{3,4} particularly for visualizing and quantifying iron in brain diseases, including multiple sclerosis (MS),⁵⁻¹⁰ distinguishing calcifications and hemorrhage,¹¹⁻¹⁵ and preoperative planning for deep brain stimulation.¹⁶ In the brain, QSM technology has advanced to a stage of high reproducibility as demonstrated by various groups.¹⁷⁻²¹ Accordingly, it is time to consider dissemination of brain QSM to routine clinical use, especially for measuring highly paramagnetic iron and highly-concentrated diamagnetic calcium.³

QSM data acquisition has largely converged to the 3D multi-echo gradient echo sequence (mGRE), and QSM reconstructions are generally flavors of the Bayesian optimization, which is a principled approach to the ill-posed magnetic field to susceptibility source inverse problem.²² As all Bayesian

reconstruction methods are fundamentally similar, we propose for routine clinical use a basic formulation of the data fidelity and structure regularization.^{1,2,23,24} The QSM reconstruction Matlab source code, along with the detailed acquisition protocols on scanners from three major manufacturers, can be downloaded freely from <http://pre.weill.cornell.edu/mri/pages/qsm.html>. For clinical applications, a robust automated QSM reconstruction suitable for inclusion in routine clinical MRI protocols is essential. A scanner vendor independent server system is presented that is connected to scanners for automated QSM reconstruction within a few minutes of data acquisition completion. Using this setup, a single-subject multi-site, multi-vendor reproducibility study and a large, single-site, multi-scanner image quality review study in a clinical environment is performed.

Methods

QSM reconstruction server implementation

The nonlinear MEDI method^{25,26} was used to reconstruct a susceptibility map, expressed in units of parts per million (ppm) or parts per billion (ppb). The algorithm performed the following steps. Input data was complex gradient echo signal $s(\mathbf{r}, TE_j)$, with $\mathbf{r}$ the voxel position and TE_j the j th echo. For each $\mathbf{r}$, a nonlinear least squares fit of the field $b(\mathbf{r})$ (rad/s)

$$b(\mathbf{r}) = \operatorname{argmin}_{b(\mathbf{r})} \sum_j |s(\mathbf{r}, TE_j) - a(\mathbf{r}, TE_j) e^{-ib(\mathbf{r})TE_j}|^2, \quad [1]$$

was performed using the Gauss-Newton algorithm,²⁷ as well as its standard error $n(\mathbf{r})$, followed by image quality guided unwrapping.²⁸ A magnitude image was constructed as

$I(\mathbf{r}) = \left(\sum_j |s(\mathbf{r}, TE_j)|^2 \right)^{1/2}$. From $I(\mathbf{r})$, two masks were constructed. The Brain Extraction Tool (BET) algorithm in the FSL toolkit^{29,30} was used to extract a brain mask $M(\mathbf{r})$. A binary edge mask $M_G(\mathbf{r})$ was obtained by computing the 3D gradient for each of the N voxels in $I(\mathbf{r})$, and retaining the largest $0.9N$ edges. From the multi-echo data, an R_2^* map was computed using the ARLO method,³¹ from which the ventricular cerebral spinal fluid (CSF) regions were segmented, obtaining a binary mask M_{CSF} .²⁶ The Projection onto Dipole Fields³² (PDF) method was used to remove the background field: $b_L(\mathbf{r}) = b(\mathbf{r}) - d * \chi_B(\mathbf{r})$ with $\chi_B(\mathbf{r}) = \operatorname{argmin}_{\chi_B(\mathbf{r})} \|M(b(\mathbf{r}) - d * \chi_B(\mathbf{r}))\|_2^2$, where χ_B was the background susceptibility on the mask $\bar{M} = 1 - M$, $*$ denoted convolution and $d(\mathbf{r}) = \frac{1}{4\pi} \frac{3\cos^2\theta - 1}{r^3} |_{r \neq 0}$ was the dipole kernel with θ the angle between $\mathbf{r}$ and $\mathbf{B}_0$. The minimization was carried out using the conjugate gradient method.²⁷ The relative difference field $\delta_b(\mathbf{r}) = b_L(\mathbf{r})/B_0$ relative to the main magnetic field B_0 was computed (in rad/s). Next, the nonlinear Morphology Enable Dipole Inversion (MEDI)²⁵ with CSF zero referencing²⁶ was used to obtain the susceptibility map (in ppm). It solved

$$\chi^*(\mathbf{r}) = \operatorname{argmin}_{\chi(\mathbf{r})} \left\| W(\mathbf{r}) (e^{-iS*d*\chi(\mathbf{r})} - e^{-i\delta_b(\mathbf{r})}) \right\|_2^2 + \lambda_1 \|M_G(\mathbf{r}) \nabla \chi(\mathbf{r})\|_1 + \lambda_2 \|M_{CSF}(\mathbf{r}) (\chi(\mathbf{r}) - \overline{\chi_{CSF}})\|_2^2, \quad [2]$$

where S is the spherical mean value operator (radius 5mm), $W(\mathbf{r}) \propto M(\mathbf{r})/n(\mathbf{r})$ reflecting the reliability of $\delta_b(\mathbf{r})$, ∇ the 3D gradient operator and $\overline{\chi_{CSF}}$ the average of $\chi(\mathbf{r})$ over the mask M_{CSF} . The spherical mean value operator is introduced to suppress remaining background field in $\delta_b(\mathbf{r})$ but introduces an erosion (5mm) of the mask M . Eq. 2 was solved iteratively using the Gauss-Newton method, using iterative reweighting.²⁵ λ_1 was set to 0.001 and λ_2 to 0.1. The average susceptibility $\overline{\chi_{CSF}^*}$ of the ventricular CSF in the solution χ^* was computed and subtracted from this

solution to obtain the final CSF referenced susceptibility map. Pixel values on the QSM map correspond to susceptibility values measured in ppb.

The nonlinear MEDI method was implemented in C++ for both CPU and GPU. The major computational costs of the nonlinear MEDI algorithm lies in the computation of FFTs (needed to compute the dipole convolution forward problem central in QSM) and in linear algebra operations (needed for nonlinear phase to field fitting, background field removal and dipole inversion). For optimized running times, these components used the Intel Math Kernel Library (MKL)³³ on the CPU, while the cuFFT³⁴ and cuBLAS³⁵ libraries were used for the GPU. All other C++ code was identical between the CPU and GPU implementations, used multithreading where appropriate and was run on the CPU. The DCMTK toolkit library³⁶ was used for DICOM input and output. The QSM reconstruction server consisted of a DICOM server able to receive images from multiple scanners simultaneously. The DICOM protocol was chosen since this is implemented by all scanner manufacturers and requires nothing more than the setting up of DICOM destination (node) on a scanner. The technologist is then able to send the mGRE images immediately after the scan to the QSM reconstruction server, where they are stored in a temporary directory. For some scanner vendors, this was done automatically and thus required no additional action by the technologist. The QSM reconstruction server continuously polled for the presence of these temporary directories and determined whether the images were suitable mGRE images, were complete, weren't previously received and/or processed and moved the data in a first-in-first-out queue. When images were present in the queue, the C++ program described above was invoked, one at a time. The resulting QSM images were then sent back to the originating scanner using the DICOM protocol. The QSM reconstruction server has been tested on Linux (Ubuntu 16.04, Redhat 7.0) and MacOS.

Acquisition

The QSM reconstruction server was installed at 9 sites (see Table 1) using scanners from three different vendors. Using this setup, two studies were carried out.

First, between June 2017 and January 2018, a single healthy subject (male, 52 years old) visited these 9 sites, signed the local IRB-approved consent form, and was then scanned using both standard resolution (StdRes) and high resolution (HiRes) mGRE protocols at 3T. The StdRes parameters were chosen based on the clinical protocol for MS patients at Weill Cornell Medical College (WCMC). The parameters were chosen to achieve a high resolution depiction in the imaging plane (in plane resolution) as well as a sampling of the magnetic field with high signal to noise ratio (echo times up to ~45ms). The resulting long TR (~50ms) necessitated increasing the slice thickness to achieve a clinically acceptable scan time below 5 minutes. The HiRes protocol was based on a deep brain stimulation imaging protocol for preoperative planning. All imaging parameters are shown in Table 2. These parameters were considered target parameters and deviations were present depending on site and scanner vendor.

Second, at WCMC, all consecutive patients for which a QSM was acquired during the month of June 2018 were selected for a retrospective review of image quality under an IRB approved protocol. Patients were scanned on 11 different scanners from two vendors (Prisma 3T, Skyra 3T, Biograph 3T PET/MR, Aera 1.5T, Siemens Healthcare, Erlangen, Germany and MR-750 3T, Architect 1.5T, and HDx 1.5T, GE Healthcare, Waukesha, WI).

Analysis

QSM reconstruction times were measured on a single server at WCMC for all data. This server used an Intel Core i9-9940X (14 Cores, 3.30 GHz) and 128G of RAM running Ubuntu 18.04. GPU reconstructions were performed using an NVIDIA GTX 1080 Ti using 3584 CUDA cores running at 1.582 GHz and 11 GB of GDDR5X memory running at 11.4 Gbps. QSM reconstruction was performed on this server for both the HR and LR healthy volunteer data from each of the 9 sites using both the CPU and GPU implementations and running times were recorded and summarized.

For each acquisition, the magnitude images $I(\mathbf{r})$ were registered to a reference magnitude image $I_{ref}(\mathbf{r})$ taken from the HiRes protocol at site #2 using FLIRT (FSL).³⁰ A HiRes protocol was chosen as reference because it provided the highest nearly isotropic resolution, thus improving image registration performance.^{37,38} On the susceptibility map obtained with the HiRes protocol at site #2, an experienced neuroradiologist (SZ, 7 years of experience) segmented the following regions of interest (ROI): Caudate Nucleus (CN), Putamen (P), Globus Pallidus (GP), Thalamus (T), Substantia Nigra (SN), Red Nucleus (RN), Posterior limb of Internal Capsule (PIC), and Splenium of Corpus Callosum (SCC). These ROIs were then registered back to each of the susceptibility map by applying the inverse transformation. An additional erosion (radius 1 voxel) of each transformed ROI in each original susceptibility map was performed to suppress potential partial volume effects. For each susceptibility map, the ROI means were recorded and their average and standard deviation across sites were recorded to assess the reproducibility of the susceptibility mapping method. Normality of the measurements was assessed using a D'Agostino & Pearson normality test. Bland-Altman, correlation analysis and two-way ANOVA analysis was performed on the ROI measurements between sites. The significance level for all tests was set at 0.05.

For the patient study at WCMC, all QSM recons were done by a reconstruction server as described above immediately after the acquisition of the mGRE data in each patient. This server used an Intel Core i7-6850K (6 Cores, 3.60GHz), 64G of RAM, and a NVIDIA GTX 1080 Ti GPU. Image quality was scored by an experienced neuroradiologist (SZ, 7 years of experience) using a 3 point scale: 3=excellent (no obvious artifacts with excellent depiction of deep gray matter regions and cortical gray-white matter contrast), 2=diagnostic (moderate artifacts but deep gray matter regions and cortical gray-white matter contrast still visualized), and 1=poor (strong artifacts and poor cortical gray-white matter contrast). QSM artifacts include those caused by motion, implants, or hemorrhage and typically manifest as streaking artifacts. Example images for each these image scores are shown in Figure 4. Disease category, gender and age were recorded for each patient.

Results

Susceptibility maps were obtained in all cases. Reconstruction times for the StdRes protocol (N=9) were 56 ± 39 s for the CPU recon and 29 ± 22 s for the GPU recon ($p=0.003$). Reconstruction times for the HiRes protocol (N=9) were 105 ± 60 s for the CPU recon and 55 ± 39 s for the GPU recon ($p=0.0002$).

Figure 1 shows an axial slice of the susceptibility maps through the midbrain of the subject imaged using the StdRes protocol. The corresponding HiRes images are shown in Figure 2. For both protocols, an excellent visual correspondence across the brain between the 9 sites is observed.

The measurements (in ppb) for the CN, P, GP, T, SN, RN, DN, PIC and SCC ROIs for each scan and protocol passed the normality test ($p > 0.11$). The ROI measurement reproducibility across scanners is shown in Figure 3a for the StdRes protocol and Figure 3b for the HiRes protocol. The standard deviation of the ROI measurements across the 9 scanners ranged between 3.6 and 24.3 ppb for the StdRes protocol and between 5 and 17.6 ppb for the HiRes protocol. These numbers are consistent with the inter-scanner and same scanner reproducibility reported in the literature.¹⁸ Averaged across all possible pairs of sites, bias (in absolute value) was 4.5 ± 2.8 ppb (StdRes protocol) and 3.4 ± 2.4 ppb (HiRes protocol), limits of agreement were $\pm 18.7 \pm 5.6$ ppb (StdRes protocol) and $\pm 14.3 \pm 3.8$ ppb (HiRes protocol) around the bias, regression slope was 0.98 ± 0.20 for the StdRes protocol and 0.96 ± 0.09 for the HiRes protocol, and the corresponding average R^2 was 0.92 ± 0.05 for the StdRes protocol and 0.96 ± 0.02 for the HiRes protocol. ANOVA analysis revealed that site accounted for 1.56% of the total variance ($P < 0.0001$) for the StdRes protocol and 0.57% ($P < 0.0001$) for the HiRes protocol.

Results for the image quality review of QSM acquired in patients are shown in Table 3. 873 patients were selected suffering from a variety of conditions. Image quality was consistently high, with 775, 67, and 31 patients having excellent image, moderate and poor image quality, respectively. Example QSM images are shown in Figures 5 through 8. In these Figures, susceptibility values are shown within an eroded (smaller) brain mask, i.e., voxels at the cortex have been removed. Figure 5 shows a QSM of an arteriovenous malformation (arrow) in the left frontal lobe, showing the nidus of abnormal blood vessels. Figure 6 shows a QSM of an atypical meningioma located on left frontoparietal convexity. Hyperintense foci of intratumoral hemorrhage are bright on QSM. Figure 7 shows a QSM of a multiple sclerosis (MS) patient. The arrow points to a lesion that is hyperintense on QSM but non-enhancing on the contrast enhanced T1 weighted image. Recent efforts indicate that enhancement or lack thereof on QSM of lesions seen on T2 can predict T1 enhancement.³⁹⁻⁴¹ Figure 8 shows an intracranial hemorrhage. QSM is hyperintense in the hemorrhage due to its strong paramagnetism.

Discussion

The results presented in this work show the feasibility of a clinically practical implementation of brain QSM. The same subject was scanned on scanners from 3 manufacturers at 9 different sites. ROI analysis revealed excellent reproducibility across sites, similar to intra-scanner and inter-scanner reproducibility reported previously.¹⁷⁻²¹ Image quality in 873 consecutive patients at one site in a month revealed diagnostic or excellent quality in 96% of patients.

The data exchange with the QSM reconstruction server uses the DICOM protocol, which requires no additional software on the scanner. The QSM reconstruction itself is fully automated, including the referencing of the susceptibility values with respect the ventricular cerebral spinal fluid, chemically close to pure water. This automated zero-reference during reconstruction improves image quality, provides absolute susceptibility values and allows longitudinal and cross-center studies, which are important for monitoring disease progression and therapy of many neurological diseases including multiple sclerosis,⁴¹⁻⁴³ cerebral cavernous malformation,⁴⁴ and Parkinson's disease,⁴⁵ and for

mapping cerebral metabolic rate of oxygen consumption,⁴⁶ and biodistribution of magnetic theranostics.⁴⁷

The presented implementation is easily integrated into a clinical workflow. Indeed, at WCMC, automated QSM reconstructions are performed routinely for over 10 scanners from two vendors and two field strengths. Gradient echo acquisitions are typically near the start of the exam such that QSM is available on the scanner by the end of the exam. Then, the technologist can send all data to PACS without further delay. However, connecting a server to a scanner can be problematic due to IT restrictions, maintenance of connectivity, and human error. To avoid these problems, we have now ported the codes for automated QSM reconstruction directly on some scanners (#1 in Table 1). In fact, the computing power on the latest scanners is higher than that of our server and allows QSM reconstruction of a StdRes acquisition within 3 minutes while scanning of other image contrasts continues without interruption.

Several studies have shown reproducibility of QSM in various settings and subject cohorts: in 14 healthy subjects scanned on a single 3T scanner four times over 7 days at a single site,¹⁷ in 10 MS patients on a single 3T scanner and 10 healthy subjects on a 1.5T and two 3T scanners from two vendors, all at a single site,¹⁸ in 8 healthy subjects on a 3T scanner at a single site,¹⁹ in 22 healthy subjects on both 1.5T and 3T from the same vendor at a single site,²⁰ in 9 healthy subjects both within site and between sites using 3T scanners.²¹ Reproducibility of QSM was shown in a Gadolinium phantom imaged at 7 different sites, using 12 clinical and 3 preclinical scanners, 3 field strengths (1.5T, 3T, and 7T) and 4 vendors.⁴⁸ The current study shows reproducibility in a single healthy volunteer imaged at 8 different sites using 9 3T scanners from 3 different vendors.

The Bayesian approach to QSM allows an optimal construction from noisy data and ill-posed system matrix (dipole kernel) under the prior knowledge of structural information expressed in an L1 norm.⁴⁹ It also allows constraining the ventricular CSF to have a uniform value suitable for referencing the final susceptibility values.²⁶ The field (phase) noise is commonly approximated as Gaussian, which breaks down at low signal to noise ratio; the use of a nonlinear data fidelity term ensures always the correct noise weighting for data fidelity using complex MRI data.²⁵ The proposed QSM implementation uses an image quality guided unwrapping method for phase processing.⁵⁰ The Laplacian based unwrapping method⁵¹ has been used for QSM because of its ease of implementation and speed, but can suffer from substantial errors.⁵² There are several ways to speed up QSM reconstruction. Using a GPU, we find an overall 3 to 5-fold reduction in reconstruction time. Additionally, using the alternating directions of multipliers method has been shown to provide a further speed up.⁵³

There are a number of limitations in this study. The spherical mean value operation (S in Eq. 2) is used to suppress remaining background field in the estimated local field, but comes at the cost of an eroded brain mask: a border of certain width (5mm in our implementation) is removed from the original brain mask computed using BET. This brain erosion may make it difficult to study lesions near the brain border (Figure 5), particularly near the temporal bones and nasal cavity. The erosion needs to be readjusted to poor SNR in this region, or QSM without brain erosion needs to be further developed to be robust and automated.⁵⁴ Finally, it is known that the magnetic susceptibility of white matter is anisotropic, i.e., it depends on the orientation with respect to the magnetic field. In

principle, anisotropic susceptibility can be resolved by acquiring data in multiple head orientations,^{55,56} but this is not clinically feasible. Using a single orientation and the isotropic signal model in Eq. 2 is known to be a source of shadow artifacts that is yet to be suppressed satisfactorily in QSM.²³ In this study, a single high resolution acquisition at one scanner was used as the reference for image registration. This may have introduced a bias, and results may change should using another reference image. We have attempted to minimize this bias by choosing a high resolution acquisition, which is nearly isotropic, as reference, following the recommendations by the authors of the FSL FLIRT image registration tool used in this work.^{30,37,38} In conclusion, this work demonstrates a clinically practical online QSM reconstruction method for a variety of scanner vendors with low cross-site ROI standard deviation. Excellent QSM image quality was observed in a consecutive patient cohort affected by a variety of diseases.

Table 1. Scanners and sites with the quantitative susceptibility mapping reconstruction server installed used in this work.

1	Weill Cornell Medical College, New York, NY	General Electric (DV25)
2	Weill Cornell Medical College, New York, NY	Siemens (VE11B)
3	Aix-Marseille University, Marseille, France	Siemens (VB17)
4	Charité – Universitätsmedizin Berlin, Germany	Siemens (VE11)
5	Centre Hospitalier Universitaire, Toulouse, France	Philips Medical Systems (5.1.7.2)
6	University of Vermont Medical Center, Burlington, VT	Philips Medical Systems (5.3.1.0)
7	Icahn School of Medicine at Mount Sinai, New York, NY	General Electric (DV25)
8	Royal Melbourne Hosp, Melbourne, VIC, Australia	Siemens (VE11B)
9	Hôpital Pitié-Salpêtrière, Paris, France	Siemens (VD13)

Table 2. Imaging parameters for the standard resolution (StdRes) and high resolution (HiRes) gradient echo acquisition for quantitative susceptibility mapping.

	StdRes	HiRes
Acquired voxel size	0.95 x 0.75 x 4.0 mm ³	0.8 x 0.8 x 1.0 mm ³
Reconstructed voxel size	0.75 x 0.75 x 3.0 mm ³	0.8 x 0.8 x 1.0 mm ³
Acquisition matrix	320 x 320 x 56	320 x 320 x 172

FOV (Readout/Phase)	240/195 mm	256/208 mm
TR	49 ms	49 ms
Number of TEs	10	10
TE1/ Δ TE	6.69/4.06 ms	6.73/4.06 ms
Coil	32 channel head coil	32 channel head coil
Flip angle	15°	15°
Reconstruction	Real/Imaginary or Magnitude/Phase	Real/Imaginary or Magnitude/Phase
Partial k-space (k_y/k_z)	100/100 %	75/75 %
Acceleration factor	2	2
Flow compensation	Readout	Readout
Readout bandwidth	260 Hz/pixel	260 Hz/pixel
Scan time	4m32s	12m19s

Abbreviations: FOV = field of view, TR = repetition time, TE = echo time, TE1 = first echo time, Δ TE = echo spacing,

Table 3. Patient demographics and image quality scores. Image quality scores used a 3-point scale: 3=excellent, 2=diagnostic, and 1=poor.

Disease category	Number	Sex (Male/Female)	Age (years, M $\pm$ SD)	Image Quality (M $\pm$ SD)
Cerebral vascular diseases (including arteriovenous malformation, e.g. Figure 5)	104	47/57	62.29 $\pm$ 19.38	2.85 $\pm$ 0.44
Brain tumor (including meningioma, e.g. Figure 6)	262	122/140	53.92 $\pm$ 20.27	2.85 $\pm$ 0.47
Demyelinating diseases (including multiple sclerosis, e.g. Figure 7)	109	31/78	43.5 $\pm$ 14.56	2.93 $\pm$ 0.3
Neurodegenerative diseases	45	23/22	62.09 $\pm$ 24.22	2.93 $\pm$ 0.33
Psychiatric diseases	12	5/7	54.25 $\pm$ 21.9	2.33 $\pm$ 0.89

Intracranial hemorrhage (e.g. Figure 8)	36	17/19	52.82±25.47	2.53±0.77
Intracranial infection	14	4/10	43.79±21.27	2.93±0.27
Seizure	29	19/10	33.35±24.41	2.97±0.19
Chiari malformation	11	3/8	30.91±22.68	2.91±0.3
Intracranial benign cyst	17	7/10	25.6±20.4	2.88±0.33
Nonspecific neurologic complaints (e.g., dizziness/headache)	234	91/143	48.55±22.92	2.85±0.41
Total	873	369/504	50.87±22.05	2.85±0.41

Abbreviations: M=mean, SD=standard deviation

Figure 1 Axial susceptibility map for a single subject using the standard resolution protocol. Numbers indicate site (see Table 1). All images shown are co-registered, showing similar basal ganglia.

Figure 2 Coronal susceptibility map for a single subject using the high resolution protocol. Numbers indicate site (see Table 1). All images shown are co-registered, clearly depicting the subthalamic nuclei (STN) and the globus pallidus (GP), important targets for deep brain stimulation. Other brain regions well visualized are the putamen (PU), external and internal globus pallidus (GPe and GPi) and the substantia nigra (SN).

Figure 3 Reproducibility of susceptibility χ (expressed in units of parts per billion or ppb) for a single subject at multiple sites for A) the standard resolution protocol and B) the high resolution protocol. The analysed regions of interest are: Caudate Nucleus (CN), Putamen (P), Globus Pallidus (GP), Thalamus (T), Substantia Nigra (SN), Red Nucleus (RN), Dentate Nucleus (DN), Posterior limb of Internal Capsule (PIC), and Splenium of Corpus Callosum (SCC). Both left (L) and right (R) values are shown where appropriate.

Autho

Figure 4 Example images for each of the scores using the image scoring analysis. A) poor image quality (score 1) in a male 64 year old patient with Multiple Sclerosis (MS), caused by large motion during the acquisition. B) diagnostic image quality (score 2) in a 55 year old female patient with MS, caused by moderate motion during the acquisition. C) excellent image quality (score 3) in a 54 year old female patient with MS.

Figure 5 A) Brain quantitative susceptibility map of a 46 year old female with an arteriovenous malformation (arrow) in the left frontal lobe, showing the nidus of abnormal blood vessels B) corresponding transverse relaxation (T2) weighted image.

Figure 6 A) Brain quantitative susceptibility map (QSM) of a 61 year old male with an atypical meningioma located on left frontoparietal convexity. QSM demonstrates hyperintense foci of intratumoral hemorrhage B) corresponding contrast enhanced longitudinal relaxation (T1) weighted image.

Figure 7 A) Brain quantitative susceptibility map (QSM) of a 46 year old male with multiple sclerosis (MS). The MS lesion indicated by the arrow in the right corona radiata is hyperintense on QSM and non-enhancing on B) the contrast enhanced longitudinal relaxation (T1) weighted image.

Figure 8 A) Brain quantitative susceptibility map of a 49 year old male with intracranial hemorrhage (arrow). The large hyperintensity indicates the strong paramagnetism of the hemorrhage B) the corresponding contrast enhanced longitudinal relaxation (T1) weighted image.

References

1. Kressler B, De Rochefort L, Spincemaille P, et al. Estimation of sparse magnetic susceptibility distributions from MRI using non-linear regularization. Proceedings of the 16th International Society of Magnetic Resonance in Medicine Annual Meeting. Toronto, Canada; 2008:1514.
2. de Rochefort L, Liu T, Kressler B, et al. Quantitative susceptibility map reconstruction from MR phase data using bayesian regularization: Validation and application to brain imaging. *Magn Reson Med* 2010;63:194-206.
3. Wang Y, Spincemaille P, Liu Z, et al. Clinical quantitative susceptibility mapping (QSM): Biometal imaging and its emerging roles in patient care. *J Magn Reson Imaging* 2017;46:951-71.
4. Eskreis-Winkler S, Zhang Y, Zhang J, et al. The clinical utility of QSM: Disease diagnosis, medical management, and surgical planning. *NMR Biomed* 2017;30.
5. Dayan M, Hurtado Rua SM, Monohan E, et al. MRI analysis of white matter myelin water content in multiple sclerosis: A novel approach applied to finding correlates of cortical thinning. *Front Neurosci* 2017;11:284.
6. Chen W, Zhang Y, Mu K, et al. Quantifying the susceptibility variation of normal-appearing white matter in multiple sclerosis by quantitative susceptibility mapping. *AJR Am J Roentgenol* 2017:1-6.
7. Zhang Y, Gauthier SA, Gupta A, et al. Magnetic susceptibility from quantitative susceptibility mapping can differentiate new enhancing from nonenhancing multiple sclerosis lesions without gadolinium injection. *AJNR Am J Neuroradiol* 2016.

8. Zhang Y, Gauthier SA, Gupta A, et al. Longitudinal change in magnetic susceptibility of new enhanced multiple sclerosis (MS) lesions measured on serial quantitative susceptibility mapping (QSM). *Journal of Magnetic Resonance Imaging* 2016;44:426-32.
9. Zhang Y, Gauthier SA, Gupta A, et al. Quantitative susceptibility mapping and R2* measured changes during white matter lesion development in multiple sclerosis: Myelin breakdown, myelin debris degradation and removal, and iron accumulation. *AJNR Am J Neuroradiol* 2016;37:1629-35.
10. Chen W, Gauthier SA, Gupta A, et al. Quantitative susceptibility mapping of multiple sclerosis lesions at various ages. *Radiology* 2014;271:183-92.
11. Ciraci S, Gumus K, Doganay S, et al. Diagnosis of intracranial calcification and hemorrhage in pediatric patients: Comparison of quantitative susceptibility mapping and phase images of susceptibility-weighted imaging. *Diagn Interv Imaging* 2017;98:707-14.
12. Zeineddine HA, Girard R, Cao Y, et al. Quantitative susceptibility mapping as a monitoring biomarker in cerebral cavernous malformations with recent hemorrhage. *J Magn Reson Imaging* 2017.
13. Chen W, Zhu W, Kovanlikaya I, et al. Intracranial calcifications and hemorrhages: Characterization with quantitative susceptibility mapping. *Radiology* 2014;270:496-505.
14. Sun H, Kate M, Gioia LC, et al. Quantitative susceptibility mapping using a superposed dipole inversion method: Application to intracranial hemorrhage. *Magn Reson Med* 2016;76:781-91.
15. Liu T, Surapaneni K, Lou M, et al. Cerebral microbleeds: Burden assessment by using quantitative susceptibility mapping. *Radiology* 2012;262:269-78.
16. Liu T, Eskreis-Winkler S, Schweitzer AD, et al. Improved subthalamic nucleus depiction with quantitative susceptibility mapping. *Radiology* 2013;269:216-23.
17. Santin MD, Didier M, Valabregue R, et al. Reproducibility of R2 * and quantitative susceptibility mapping (QSM) reconstruction methods in the basal ganglia of healthy subjects. *NMR Biomed* 2017;30.
18. Deh K, Nguyen TD, Eskreis-Winkler S, et al. Reproducibility of quantitative susceptibility mapping in the brain at two field strengths from two vendors. *J Magn Reson Imaging* 2015;42:1592-600.
19. Feng X, Deistung A, Reichenbach JR. Quantitative susceptibility mapping (QSM) and R2(*) in the human brain at 3T: Evaluation of intra-scanner repeatability. *Z Med Phys* 2018;28:36-48.
20. Hinoda T, Fushimi Y, Okada T, et al. Quantitative susceptibility mapping at 3 T and 1.5 T: Evaluation of consistency and reproducibility. *Invest Radiol* 2015;50:522-30.
21. Lin PY, Chao TC, Wu ML. Quantitative susceptibility mapping of human brain at 3T: A multisite reproducibility study. *AJNR Am J Neuroradiol* 2015;36:467-74.
22. Langkammer C, Schweser F, Shmueli K, et al. Quantitative susceptibility mapping: Report from the 2016 reconstruction challenge. *Magn Reson Med* 2018;79:1661-73.
23. Kee Y, Liu Z, Zhou L, et al. Quantitative susceptibility mapping (QSM) algorithms: Mathematical rationale and computational implementations. *IEEE Trans Biomed Eng* 2017;64:2531-45.
24. Wang Y, Liu T. Quantitative susceptibility mapping (QSM): Decoding MRI data for a tissue magnetic biomarker. *Magn Reson Med* 2015;73:82-101.

25. Liu T, Wisnieff C, Lou M, et al. Nonlinear formulation of the magnetic field to source relationship for robust quantitative susceptibility mapping. *Magn Reson Med* 2013;69:467-76.
26. Liu Z, Spincemaille P, Yao Y, et al. MEDI+0: Morphology enabled dipole inversion with automatic uniform cerebrospinal fluid zero reference for quantitative susceptibility mapping. *Magn Reson Med* 2018;79:2795-803.
27. Nocedal J, Wright SJ. Numerical optimization. 2nd ed. New York: Springer; 2006:xxii, 664 p.
28. Cusack R, Papadakis N. New robust 3-D phase unwrapping algorithms: Application to magnetic field mapping and undistorting echoplanar images. *Neuroimage* 2002;16:754-64.
29. Smith SM. Fast robust automated brain extraction. *Hum Brain Mapp* 2002;17:143-55.
30. Jenkinson M, Beckmann CF, Behrens TE, et al. FSL. *Neuroimage* 2012;62:782-90.
31. Pei M, Nguyen TD, Thimmappa ND, et al. Algorithm for fast monoexponential fitting based on auto-regression on linear operations (arlo) of data. *Magn Reson Med* 2015;73:843-50.
32. Liu T, Khalidov I, de Rochefort L, et al. A novel background field removal method for MRI using projection onto dipole fields (PDF). *NMR Biomed* 2011;24:1129-36.
33. Intel math kernel library. <https://software.intel.com/en-us/mkl>.
34. cuFFT library. <https://developer.nvidia.com/cufft>.
35. cuBLAS library. <https://developer.nvidia.com/cublas>.
36. DCMTK - dicom toolkit. <http://dicom.offis.de/dcmtoolkit.php.en>.
37. Jenkinson M, Smith S. A global optimisation method for robust affine registration of brain images. *Med Image Anal* 2001;5:143-56.
38. Jenkinson M, Bannister P, Brady M, et al. Improved optimization for the robust and accurate linear registration and motion correction of brain images. *Neuroimage* 2002;17:825-41.
39. Gupta A, Al-Dasuqi K, Xia F, et al. The use of noncontrast quantitative MRI to detect gadolinium-enhancing multiple sclerosis brain lesions: A systematic review and meta-analysis. *AJNR Am J Neuroradiol* 2017;38:1317-22.
40. Zhang S, Nguyen TD, Zhao Y, et al. Diagnostic accuracy of semiautomatic lesion detection plus quantitative susceptibility mapping in the identification of new and enhancing multiple sclerosis lesions. *Neuroimage Clin* 2018;18:143-8.
41. Zhang Y, Gauthier SA, Gupta A, et al. Magnetic susceptibility from quantitative susceptibility mapping can differentiate new enhancing from nonenhancing multiple sclerosis lesions without gadolinium injection. *AJNR Am J Neuroradiol* 2016;37:1794-9.
42. Chen W, Gauthier SA, Gupta A, et al. Quantitative susceptibility mapping of multiple sclerosis lesions at various ages. *Radiology* 2014;271:183-92.
43. Kaunzner UW, Kang Y, Zhang S, et al. Quantitative susceptibility mapping identifies inflammation in a subset of chronic multiple sclerosis lesions. *Brain* 2019;142:133-45.
44. Tan H, Liu T, Wu Y, et al. Evaluation of iron content in human cerebral cavernous malformation using quantitative susceptibility mapping. *Invest Radiol* 2014;49:498-504.
45. Murakami Y, Kakeda S, Watanabe K, et al. Usefulness of quantitative susceptibility mapping for the diagnosis of parkinson disease. *AJNR Am J Neuroradiol* 2015;36:1102-8.
46. Zhang J, Liu T, Gupta A, et al. Quantitative mapping of cerebral metabolic rate of oxygen (CMRO₂) using quantitative susceptibility mapping (QSM). *Magn Reson Med* 2015;74:945-52.

47. Liu T, Spincemaille P, de Rochefort L, et al. Unambiguous identification of superparamagnetic iron oxide particles through quantitative susceptibility mapping of the nonlinear response to magnetic fields. *Magn Reson Imaging* 2010;28:1383-9.
48. Deh K, Kawaji K, Bulk M, et al. Multicenter reproducibility of quantitative susceptibility mapping in a gadolinium phantom using MEDI+0 automatic zero referencing. *Magn Reson Med* 2019;81:1229-36.
49. Liu J, Liu T, de Rochefort L, et al. Morphology enabled dipole inversion for quantitative susceptibility mapping using structural consistency between the magnitude image and the susceptibility map. *Neuroimage* 2012;59:2560-8.
50. Xu W, Cumming I. A region-growing algorithm for insar phase unwrapping. *IEEE Transactions on Geoscience and Remote Sensing* 1999;37:124-34.
51. Schofield MA, Zhu Y. Fast phase unwrapping algorithm for interferometric applications. *Opt Lett* 2003;28:1194-6.
52. Robinson SD, Bredies K, Khabipova D, et al. An illustrated comparison of processing methods for MR phase imaging and QSM: Combining array coil signals and phase unwrapping. *NMR Biomed* 2017;30:e3601.
53. Milovic C, Bilgic B, Zhao B, et al. Weak-harmonic regularization for quantitative susceptibility mapping. *Magn Reson Med* 2019;81:1399-411.
54. Liu Z, Kee Y, Zhou D, et al. Preconditioned total field inversion (TFI) method for quantitative susceptibility mapping. *Magn Reson Med* 2017;78:303-15.
55. Wisnieff C, Liu T, Spincemaille P, et al. Magnetic susceptibility anisotropy: Cylindrical symmetry from macroscopically ordered anisotropic molecules and accuracy of MRI measurements using few orientations. *Neuroimage* 2013;70:363-76.
56. Liu C. Susceptibility tensor imaging. *Magn Reson Med* 2010;63:1471-7.

Minerva Access is the Institutional Repository of The University of Melbourne

Author/s:

Spincemaille, P;Liu, Z;Zhang, S;Kovanlikaya, I;Ippoliti, M;Makowski, M;Watts, R;de Rochefort, L;Venkatraman, V;Desmond, P;Santin, MD;Lehericy, S;Kopell, BH;Peran, P;Wang, Y

Title:

Clinical Integration of Automated Processing for Brain Quantitative Susceptibility Mapping: Multi-Site Reproducibility and Single-Site Robustness

Date:

2019-08-04

Citation:

Spincemaille, P., Liu, Z., Zhang, S., Kovanlikaya, I., Ippoliti, M., Makowski, M., Watts, R., de Rochefort, L., Venkatraman, V., Desmond, P., Santin, M. D., Lehericy, S., Kopell, B. H., Peran, P. & Wang, Y. (2019). Clinical Integration of Automated Processing for Brain Quantitative Susceptibility Mapping: Multi-Site Reproducibility and Single-Site Robustness. JOURNAL OF NEUROIMAGING, 29 (6), pp.689-698. <https://doi.org/10.1111/jon.12658>.

Persistent Link:

<http://hdl.handle.net/11343/286250>