Electrode Discrimination by Early-Using the Cochlear Limited Multiple-Electrode Cochlear Implant

P. A. Busby and G. M. Clark

Objective: The aims of this study were to determine whether electrode discrimination by early-deafened subjects using the Cochlear Limited prosthesis varied at different locations on the electrode array, was influenced by the effects of auditory deprivation and experience with electric stimulation, and was related to speech perception.

Design: Difference limens for electrode discrimination were measured in 16 early-deafened subjects at three positions on the array: electrodes 18 (apical), 14 (mid), and 8 (basal). Electrodes were stimulated using random variations in current level to minimize the influence of loudness cues. Assessed were correlations between the difference limens, subject variables related to auditory deprivation (age at onset of deafness, duration of deafness, and age at implantation) and auditory experience (duration of implant use and the total time period of auditory experience), and speech perception scores from two closed-set and two open-set tests.

Results: The average difference limens across the three positions were less than two electrodes for 75% of subjects, with average limens between 2 and 6.5 electrodes for the remaining 25% of subjects. Significant differences across the three positions were found for 69% of subjects. The average limens and those at the basal position positively correlated with variables related to auditory deprivation, with larger limens for subjects implanted at a later age and with a longer duration of deafness. The average limens and those at the apical position negatively correlated with closed-set speech perception scores, with lower scores for subjects with larger limens, but not with open-set scores. Speech scores also negatively correlated with variables related to auditory deprivation.

Conclusions: These findings showed that early-deafened subjects were generally successful in electrode discrimination although performance varied across the array for over half the subjects. Discrimination performance was influenced by the effects of auditory deprivation, and both electrode discrimination and variables related to auditory deprivation influenced closed-set speech perception.

(*Ear & Hearing* 2000;21:291-304)

Department of Otolaryngology, University of Melbourne, Victoria, Australia.

One of the more important parameters for coding acoustic information in multiple-electrode cochlear implants is stimulation of the different electrodes. Most speech processing strategies assign or map the outputs of a series of bandpass filters to the different electrodes in accordance with the tonotopic organization of the cochlea (Eddington, 1980; McDermott, McKay, & Vandali, 1992; Skinner et al., 1994; Tong, Blamey, Dowell, & Clark, 1983; Tong, Clark, Blamey, Busby, & Dowell, 1982; Wilson, Finley, Lawson, Wolford, Eddington, & Rabinowitz, 1991). Although the overall order of pitch percepts for stimulating electrodes along the array is usually consistent with the tonotopic order in postlinguistically deaf adult subjects, it also has been the case that not all electrodes are perceptually distinct from each other, as shown in a number of studies using several different procedures such as pitch estimation, pitch ranking, electrode identification, and electrode discrimination (Busby, Whitford, Blamey, Richardson, & Clark, 1994; Collins, Zwolan, & Wakefield, 1997; Cohen, Busby, Whitford, & Clark, 1996; Dorman, Smith, McCandless, Dunnivant, Parkin, & Dankowski, 1990; Eddington, Dobelle, Brackman, Mladejovsky, & Parkin, 1978; Henry, McKay, McDermott, & Clark, 1997; Nelson, Van Tasell, Schroder, Soli, & Levine, 1995; Shannon, 1983; Tong & Clark, 1985; Zwolan, Collins, & Wakefield, 1997). It has also been reported in some studies with postlinguistically deaf adults using the Cochlear Limited prosthesis of 22 electrodes that speech perception scores were higher in subjects with a greater number of perceptually distinct electrodes (Collins et al., 1997; Nelson et al., 1995). In addition, Henry et al. (1997) showed in a group of eight postlinguistically deaf adults using the Cochlear Limited prosthesis that speech perception scores correlated with electrode discrimination at the apical and mid positions on the array, but not the basal position. However, although Zwolan et al. (1997) showed no correlation between speech perception scores and electrode discrimination in postlinguistically deaf adults using the Cochlear Limited prosthesis, scores improved for seven of the nine subjects tested when nondiscriminable electrodes were excluded from the speech coding strategy.

Thus, it appears that the number of perceptually distinct electrodes available for speech coding could be an important factor in speech perception.

Considerably less is known about the ability of early-deafened subjects to perceive differences between electrodes. For these subjects, the onset of the profound-total hearing loss was before birth or at a very early age and they typically have had very limited auditory experience. They are, therefore, an important subject group for investigation as their performance is likely to be influenced by the effects of auditory deprivation at or near birth and electric stimulation as the primary means of auditory experience. It is possible that early onset of profound deafness and lack of auditory stimulation from a very early age could result in incomplete maturation of the auditory system and/or degeneration of underlying auditory processes, as demonstrated in experimental animal studies (Hardie, Marsti-McClintock, Aitkin, & Shepherd, 1998; Harrison, Nagasawa, Smith, Stanton, & Mount, 1991; Moore, 1990). It also appears that the effects of auditory deprivation are greater during development in the very young animal than in the older animal (Hashisaki & Rubel, 1989; Moore, 1990; Tierney, Russell, & Moore, 1997). However, auditory deprivation from birth may not lead to a complete loss of the tonotopic projections in the auditory pathway as studies with the congenitally deaf white cat have shown some tonotopic organization at the auditory cortex (Hartmann, Shepherd, Heid, & Klinke, 1997), with evidence of intact afferent projections at least to the level of the inferior colliculus (Heid, Jahn-Siebert, Klinke, Hartmann, & Langner, 1997). There may also be changes in the auditory pathway as a consequence of electric stimulation. For instance, Snyder, Rebscher, Cao, Leake, and Kelly (1990) showed an expansion in the representation of the site of stimulation at the inferior colliculus after chronic electric stimulation in the cochlea. Thus, it may be the case that the ability of early-deafened subjects to perceive differences between electrodes could differ from that of postlinguistically deaf adults. Also, differences between early-deafened subjects may be related to the effects of auditory deprivation and/or experience with audition via electric stimulation.

Early investigations with small numbers of early-deafened subjects have suggested that they have been less successful than postlinguistically deaf adults in the perception of differences between electrodes, and that performance was related to the effects of auditory deprivation in some cases. Eddington et al. (1978) reported that pitch estimates were not tonotopically ordered in the one early-deafened adult tested using a prosthesis of six electrodes. In preliminary studies with four early-deaf-

ened adults and children using the Cochlear Limited prosthesis that investigated electrode identification (Busby, Tong, & Clark, 1992; Tong, Busby, & Clark, 1988), subjects were not able to successfully identify two electrodes that were separated by eight electrodes or identify three electrodes separated by four electrodes. This level of performance was considerably poorer than that of six postlinguistically deaf adults using the Cochlear Limited prosthesis who were generally able to identify seven electrodes separated by one or two electrodes (Tong & Clark, 1985).

We have also investigated electrode discrimination in small groups of early-deafened subjects implanted with the Cochlear Limited prosthesis using both single electrodes and electrode trajectories as stimuli (Busby & Clark, 1996; Busby, Tong, & Clark, 1993). For electrode trajectories, subjects discriminated between electrodes stimulated in an apical-to-basal direction and the same electrodes stimulated in a basal-to-apical direction, where it was hypothesized that discrimination judgments would be related to the ability to discriminate the direction of pitch change. In the first study of trajectory discrimination, percent correct scores were obtained for different ranges of electrodes in the trajectories, from two to 10 electrodes (Busby et al., 1993). Only one of the four early-deafened subjects tested was able successfully discriminate all the trajectories and gave results similar to those from a group of four postlinguistically deaf adults. This early-deafened subject was deafened at a later age and had been deaf for a shorter period of time than the other early-deafened subjects. For the other three early-deafened subjects, scores were considerably lower for trajectories using a small range of electrodes but tended to increase with increases in the range of electrodes. In addition, speech perception scores for a single-electrode strategy were significantly lower than those for a multiple-electrode strategy for the four postlinguistically deaf adults and the one early-deafened subject with high electrode trajectory discrimination scores, suggesting that they were using speech information coded on electrodes for speech perception. The three early-deafened subjects with low trajectory discrimination scores had similar scores for the two strategies, suggesting that they were not effectively using the speech information coded on electrodes. The multiple-electrode strategy coded the first and second formant frequencies on two separate electrodes, in addition to the fundamental frequency as pulse rate and the amplitudes of the two formants as electric stimulation levels on the two electrodes. The single-electrode strategy coded the fundamental frequency and the amplitudes of the two formants on a single electrode.

In our second study (Busby & Clark, 1996), the discrimination of single electrodes and electrode trajectories was measured at the apical, mid, and basal positions of the Cochlear Limited array in six early-deafened subjects to compare between time-invariant and time-varying electrode perception at different locations on the array. Also investigated was whether performance was similar for two data collection paradigms: an adaptive procedure to obtain difference limens, and a method of constant stimuli to provide psychometric functions. Difference limens were less than two to three electrodes for most subjects, although limens of four to eight electrodes were recorded for one subject who was implanted at a later age and was deaf for a longer period of time than the other subjects. Discrimination performance was similar for single electrodes and electrode trajectories. There were also differences across the three positions on the array for some subjects. The psychometric functions generally showed that discrimination scores improved with increases in the distance between electrodes, and that the difference limens from the adaptive procedure were similar to the corresponding point on the psychometric functions. These findings were consistent with those from postlinguistically deaf adults using the Cochlear Limited prosthesis (Zwolan et al., 1997), indicating that adaptive procedures are useful for estimating electrode discrimination in cochlear implant subjects.

The objectives of this study with a larger group of early-deafened subjects were to determine whether electrode discrimination varied at the apical, mid, and basal positions of the Cochlear Limited array of 22 electrodes, whether performance was related to auditory deprivation and experience with electric stimulation, and whether electrode discrimination was related to speech perception performance. In our previous studies, it was not possible to examine these issues because of the small number of subjects tested. Group data from 16 subjects are presented in this study, consisting of data from six subjects in our previous study (Busby & Clark, 1996) and new data from an additional 10 subjects. It was hypothesized that electrode discrimination would vary across the array in early-deafened subjects, consistent with findings from postlinguistically deaf adults (Collins et al., 1997; Henry et al., 1997; Nelson et al., 1995; Zwolan et al., 1997). It was also hypothesized that electrode discrimination performance would be poorer as a result of long-term auditory deprivation and/or improve with postoperative experience. The importance of these variables in speech perception has also been reported in several studies (cf. Proceedings of the NIH Consensus Statement, Reference Note 1). For instance, Fryauf-Bertschy, Tyler,

Kelsay, Gantz, and Woodworth (1997) showed that open-set word recognition scores were higher in children implanted before 5 yr of age, thus with a shorter duration of auditory deprivation, and that children required several years of experience with the implant before significant scores were obtained, thus postoperative experience resulted in improvements in performance. In addition, it was hypothesized that speech perception performance would be related to electrode discrimination, as demonstrated in some studies with postlinguistically deaf adults (Collins et al., 1997; Henry et al., 1997; Nelson et al., 1995).

METHODS

Subjects

The histories of the 16 subjects are summarized in Table 1, ordered according to their performance in electrode discrimination. The average age of the subjects at the time of diagnosis of the profound-total bilateral sensorineural hearing loss was 17.3 mo (SD = 11.4 mo), with a range of 6 to 47 mo. Three subjects were deafened as a result of meningitis; at 39, 47, and 24 mo of age for P9, P10, and P12, respectively. The other subjects were congenitally deaf from a variety of etiologies, where almost half of these were unknown ($N = 6$). The subjects were implanted with the Cochlear Limited prosthesis at an average age of 9.3 yr (SD = 4.9 yr), with a range of 3.5 to 20 yr. The Cochlear Limited array consists of 22 banded electrodes at a uniform spacing of 0.75 mm (Clark et al., 1987). In some cases, some electrodes were not available for stimulation as they were external to the scala tympani, stimulation produced unpleasant sensations, or there was an open or short circuit. The number of electrodes used by the subjects ranged from 15, for P6 and P8, to 20. The average age of the subjects at the time of testing was 13.2 yr (SD = 4.2 yr), with a range of 8.7 to 21.7 yr. The average duration of implant use at the time of testing was 3.9 yr (SD = 1.8 yr), with a range of 1.0 to 7.1 yr. Note that electrode discrimination data for six subjects (P1, P4, P8, P12, P13, and P16) were previously reported in our preliminary study (Busby & Clark, 1996).

Stimuli

Stimuli were biphasic current pulses with a pulse duration of 200 μ sec/phase. The Cochlear Limited prosthesis can deliver electric current between 15 and 1500 μ A, which is converted to a scale of 239 levels with approximately a 2.5% increase in current for each level (Skinner et al., 1991), although there are some minor variations across individual prostheses.

TABLE 1. Summary of subject histories

Subject	Age at Confirmation of Profound-Total Hearing Loss (mo)	Cause of Deafness	Age at Implantation (yr)	Age at Testing (yr)	Duration of Implant Use (yr)
P1	6	Rubella	6.6	11.1	4.5
P2	15	Congenital, unknown	11.7	13.8	2.2
P3	16	Cytomegalovirus	4.6	11.7	7.1
P4	24	Klippel-Feil and Mondini syndromes	8.6	11.3	2.7
P5	11	Congenital, unknown	6.6	11.0	4.5
P6	10	Cytomegalovirus	7.8	12.1	4.3
P7	9	Congenital, unknown	8.2	11.9	3.7
P8	6	Usher's syndrome	14.8	20.3	5.5
P9	39	Meningitis	4.8	9.8	5.1
P10	47	Meningitis	5.2	10.4	5.2
P11	18	Usher's syndrome	20.0	21.0	1.0
P12	24	Meningitis	3.5	8.7	5.3
P13	15	Rubella	5.2	9.6	4.4
P14	12	Congenital, unknown	10.4	11.4	1.0
P15	12	Congenital, unknown	13.6	14.7	1.1
P16	12	Congenital, unknown	17.6	21.7	4.1

theses. In this report, electric stimulation levels are described using these current levels. The duration of all stimuli was 300 msec and the pulse rate was 250 pulses/s.

Bipolar stimulation was used for all subjects. For most subjects, except P4 and P6, the two electrodes in the bipolar pair were separated by one electrode, corresponding to "BP+1" in the Cochlear Limited clinical procedures. For P4 and P6, the two electrodes were separated by two electrodes, corresponding to "BP+2" in the clinical procedures. It was necessary to use BP+2 stimulation for these two subjects to ensure that the comfortable listening levels were within the operating range of the implant. Electrodes are labeled according to the basal member of the bipolar pair and are numbered 22 to 1 in an apical-basal direction, consistent with the Cochlear Limited clinical numbering method.

Difference limens for electrode discrimination were obtained for three reference electrodes, electrodes 18, 14, and 8, which are labeled in this report as the apical, mid, and basal positions, respectively. In the SPECTRA speech processing strategy of the Cochlear Limited prosthesis, the acoustic frequency bandwidths for electrodes 18, 14, and 8 are 550 to 750 Hz, 1350 to 1550 Hz, and 3079 to 3571 Hz, respectively, in BP+1 stimulation when using all 20 available electrodes. In BP+2 stimulation when using all 19 electrodes, the frequency bandwidths for these electrodes are lower: 350 to 550 Hz, 1150 to 1350 Hz, and 2680 to 3079 Hz, respectively. In the adaptive procedure described below, the comparison electrodes were basal to the reference electrode. Thus, the corresponding acoustic frequency bandwidths of the filters used in the SPECTRA strategy for the comparison electrodes were higher than those for the reference electrodes.

Threshold and Comfortable Listening Levels

Before data collection for each subject, threshold (T) and comfortable listening (C) levels were obtained for all electrodes available for stimulation. These levels were used to determine the stimulation levels on the reference and comparison electrodes. The T and C levels were measured using single presentations of the stimuli and the step-size was two current levels. The T levels were measured using the Hughson-Westlake ascending technique (Carhart & Jerger, 1959). The C levels were measured using the ascending-descending technique used in our previous studies with postlinguistically deaf and early-deafened subjects (Busby et al., 1994, Busby & Clark, 1996). Current was increased until stimulation became too loud or uncomfortable, and it was then decreased to a level corresponding to a comfortable listening level. The C levels were further adjusted to ensure that the electrodes were approximately of equal loudness. First, a group of four consecutive electrodes were stimulated, with a 500 msec interval between successive electrodes, and the subject indicated whether any electrodes required adjustment for equal loudness. Second, all electrodes available for stimulation were sequentially presented in an apical-basal direction and the subject indicated whether any electrodes needed further adjustment to the C level.

Procedure

An adaptive three-interval forced choice procedure was used to obtain the electrode difference limens. Each trial consisted of three intervals separated by 500 msec. Two of the intervals were the reference electrode and one was the comparison electrode, and the interval with the comparison was

TABLE 2. Maximum ranges for the randomized current level variations and dynamic ranges at the three positions on the array. The numbers in parentheses at each position indicate the electrodes used to calculate the mean dynamic ranges

Subject	Apical (18-14)		Mid (14-10)		Basal (8-4)	
	Maximum Variation	Dynamic Range	Maximum Variation	Dynamic Range	Maximum Variation	Dynamic Range
P1	0-20	78	0-20	78	0-20	73
P2	0-15	35	0-15	36	0-15	37
P3	0-20	38	0-20	59	0-20	54
P4	0-15	33	0-10	16	0-10	17
P5	0-15	23	0-10	20	0-10	15
P6	0-20	50	0-20	65	0-20	57
P7	0-20	32	0-20	31	0-15	27
P8	0-20	34	0-20	31	0-20	29
P9	0-15	32	0-15	33	0-15	40
P10	0-20	59	0-20	46	0-15	33
P11	0-20	55	0-20	62	0-20	44
P12	0-20	35	0-20	36	0-20	37
P13	0-10	20	0-10	17	0-10	18
P14	0-10	13	0-10	13	1-10	15
P15	0-15	22	0-10	18	0-10	18
P16	0-15	22	0-10	19	0-10	15

randomly selected in each trial. The comparison electrode was basal to the reference electrode. Each interval was cued by a number (1 to 3) graphically displayed on a computer screen. The subject selected the number corresponding to the comparison in each trial using a verbal response or computer mouse. Visual feedback was given after each trial. A two-down, one-up adaptive procedure that estimated the 70.7% correct point on a psychometric function (Levitt, 1971) was used to obtain the difference limen. The comparison was chosen to be clearly different from the reference at the beginning of the adaptive procedure. The step-size of change was one electrode for all reversals. Each run was terminated after 10 reversals and the average of the last six was taken as the result. Limens were separately obtained for the apical, mid, and basal reference electrodes for each subject.

To ensure that the effects of any loudness differences between the reference and comparison stimuli were minimized and that discrimination judgments were primarily based on perceptual cues related to stimulation of two different sites in the cochlea, each of the three intervals in each trial of the adaptive procedure was presented using a randomly varied current level (Busby & Clark, 1996). Different ranges of random variation in current levels below the C level were used: 0 (no variation), and between 0 to 5, 0 to 10, 0 to 15, and 0 to 20 current levels. However, for some subjects it was not possible to collect data at the larger ranges (0 to 15 and 0 to 20 current levels) because of a small dynamic range between T and C levels. Shown in Table 2 are the maximum ranges used in data collection for each of the three reference electrodes and the dynamic

ranges for the same electrodes. The dynamic range values have been averaged over five electrodes basal to the reference, as five electrodes were typically used in the adaptive procedure. The average maximum amount of current level variation across subjects was 0 to 50% of the dynamic range (SD = 13%), with a range of 0 to 26% to 0 to 69%.

The electrode difference limens collected using no (0) current level variation were not included in data analysis or the reporting of results because loudness cues may have been available to the subjects. This was because complete loudness balancing between electrodes may not have been achieved due to the relatively large step-size of actual current between the current level steps of the Cochlear Limited prosthesis. Also, it is possible that children were not able to provide accurate and consistent judgments of equal loudness, and that there would be across-electrode and between-subject variability in these judgments. In 25% of cases, across subjects and electrodes, the difference limens at 0 current level variation were more than one electrode smaller than the average limen where current level variation was used. Similar findings were reported in our previous study with children (Busby & Clark, 1996). Thus, only the difference limens for 0 to 5 to 0 to 20 current level variation were used.

Subject Variables and Speech Perception Scores

The following data were obtained from the clinic histories: age at onset of profound deafness, duration of deafness before implantation, age at implantation, and duration of implant use. Also calculated

TABLE 3. Correlations between subject variables and the corresponding levels of significance

	Age at Onset of Deafness	Duration of Deafness	Age at Testing	Duration of Implant Use	Age at Implantation
Duration of deafness	$R^2 = 38.1\%$ $p = 0.011 (-)$				
Age at testing	$R^2 = 14.4\%$ $p = 0.147$	$R^2 = 82.7\%$ $p < 0.001 (+)$			
Duration of implant use	$R^2 = 12.5\%$ $p = 0.179$	$R^2 = 35.2\%$ $p = 0.015 (-)$	$R^2 = 7.0\%$ $p = 0.322$		
Age at implantation	$R^2 = 20.0\%$ $p = 0.082$	$R^2 = 96.0\%$ $p < 0.001 (+)$	$R^2 = 87.9\%$ $p < 0.001 (+)$	$R^2 = 34.1\%$ $p = 0.018 (-)$	
Total time of auditory experience	$R^2 = 56.7\%$ $p = 0.001 (+)$	$R^2 = 53.1\%$ $p = 0.001 (-)$	$R^2 = 14.3\%$ $p = 0.149$	$R^2 = 77.7\%$ $p < 0.001 (+)$	$R^2 = 40.5\%$ $p = 0.008 (-)$

The (+) and (-) indicate significant positive and negative correlations, respectively.

was the total time period of auditory experience (age at onset of deafness plus duration of implant use) to determine whether total auditory experience (acoustic and electric) was a variable influencing electrode discrimination. Age at the time of testing was also included as a subject variable to determine whether subject differences in electrode discrimination were influenced by any cognitive or task-dependent factors related to chronological age.

For measures of speech perception performance, however, clinic data were not available for all subjects for all the tests used in clinic assessments. The speech perception data reported here were collected on average within 0.38 yr (SD = 0.42 yr) of the electrode discrimination data. Selected were data from two closed-set monosyllabic word tests presented in the audition alone condition: the Picture Vocabulary test (Plant, 1984), which consists of 12 words presented using a 12-alternative forced-choice paradigm ($N = 14$), and the NU-CHIPS test (Elliott & Katz, 1980), which consists of 50 words presented using a 4-alternative forced-choice paradigm ($N = 9$). Data from two open-set tests were also selected: word scores for BKB sentences (Bench & Bamford, 1979) presented in the audition alone ($N = 15$) and auditory-visual ($N = 15$) conditions, and phoneme scores for PBK words (Haskins, Reference Note 2) presented in the audition alone ($N = 12$) and auditory-visual ($N = 13$) conditions. In addition, the lipreading enhancement scores were calculated from the word scores for BKB sentences ($N = 12$) and phoneme scores for PBK words ($N = 10$) using:

Lipreading Enhancement = $(AV - V)/(100 - V)$;
where V and AV were scores for the vision alone and auditory-visual conditions, respectively, providing the percent score of the possible improvement over the vision alone score (Tyler, Opie, Fryauf-Bertschy, & Gantz, 1992).

Correlations between electrode discrimination performance, subject variables, and speech perception scores were determined using the Pearson prod-

uct moment coefficient of correlation. These are expressed using the coefficient of determination (R^2) to indicate the proportion of accounted for variance. For each subject at each of the three reference electrodes, the average difference limen was calculated for the different amounts of current level variation. In addition for each subject, the grand mean over the three positions and the different amounts of current level variation was calculated. These values were used in the correlation analyses.

There were a number of important correlations between subject variables, shown in Table 3, which influenced their use in data analysis. For the congenitally deaf subjects, who constituted 81.25% of subjects, values for duration of deafness and age at implantation were the same, thus a high positive correlation was found ($R^2 = 96.0\%$, $p < 0.001$). Also for the congenitally deaf subjects, values were the same for duration of implant use and total auditory experience, thus the positive correlation was high ($R^2 = 77.7\%$, $p < 0.001$). There were also other correlations between variables for our group of subjects. Both age at implantation and duration of deafness negatively correlated with duration of implant use and total auditory experience, indicating that subjects with a longer period of auditory deprivation had less auditory experience. Age at testing positively correlated with age at implantation and duration of deafness, indicating that subjects with a longer period of auditory deprivation were tested at an older age. In addition for the congenitally deaf subjects, the age at onset of deafness was 0 yr, thus correlations with this variable were unlikely to be particularly meaningful.

RESULTS

Figure 1 shows the average electrode difference limens for the different amounts of current level variation at each of the three reference positions for the 16 subjects, with the error bars indicating the

TABLE 4. Difference limens for electrode discrimination at each reference position, the mean difference limen, and speech perception scores for closed-set (Picture Vocabulary) and open-set (BKB sentences) tests presented in the audition alone condition

Subject	Electrode Difference Limens				Speech Perception (% correct)	
	Apical	Mid	Basal	Mean	Picture Vocabulary	BKB Sentences
P1	1.04	0.50	0.50	0.68	91.7	51.0
P2	1.06	0.67	0.78	0.83	88.0	4.0
P3	0.92	1.04	0.92	0.96	79.0	2.0
P4	0.67	0.56	2.54	1.18	87.5	0.0
P5	0.56	1.00	2.42	1.21	75.0	—
P6	1.71	1.29	1.09	1.36	87.5	12.0
P7	0.75	1.04	2.72	1.39	—	52.0
P8	1.38	1.21	1.63	1.40	62.5	2.0
P9	2.49	0.67	1.06	1.40	100.0	46.0
P10	1.28	0.67	2.39	1.44	—	84.0
P11	0.67	1.17	4.50	1.75	54.2	0.0
P12	1.38	1.88	2.04	1.76	100.0	53.0
P13	0.59	4.25	1.50	2.11	79.2	34.0
P14	1.66	4.67	2.91	3.08	100.0	72.0
P15	5.78	1.00	1.67	3.24	46.0	2.0
P16	6.50	5.09	3.92	5.36	33.0	0.0

standard error. The ordering of subjects is from smallest to largest average limen across all three positions. The limens for each position and the grand mean across the three positions are also given in Table 4. The smallest average limen over the three positions was 0.68 electrodes for P1 and the largest average limen was 5.36 electrodes for P16. These average limens were less than two electrodes for 12 subjects (P1 through P12) and greater than two electrodes for the remaining four subjects (P13 through P16). To determine whether limens differed across the three positions, the data for each subject were separately analyzed using a 1-way analysis of variance. Post hoc analyses of the means using the Tukey test were used to determine which positions

contributed to the significant effect at a level of 0.05. For 11 subjects, marked by asterisks in Figure 1, a significant difference across the three positions was found. For five subjects (P1, P6, P9, P15, and P16), the apical limen was larger than the mid and/or basal limen. For one subject (P13), the mid limen was larger than the apical and basal limens, and the basal limen was also larger than the apical limen. For the remaining five subjects where a significant difference was found (P4, P5, P7, P10, and P11), the basal limen was larger than the mid and apical limen, but there was no difference between the mid and apical limens. Note that in two cases, these differences between the three positions were less than one electrode and could be interpreted as having no importance: P1 and P6. On the other hand, considerably larger differences of greater than three electrodes, which could be taken as having some importance, were recorded in several instances: P11 for the difference between the basal and both the mid and apical limens, P13 and P14 for the difference between the mid and apical limens, and P15 for the difference between the apical and both the mid and basal limens. Note that limens for the three positions did not significantly correlate with each other ($R^2 < 16.2\%$, $p > 0.124$).

Figure 2 shows the average electrode difference limens across the three positions as functions of the six subject variables investigated. Also shown in each graph are the correlation coefficients and corresponding levels of significance. A significant positive correlation, shown by the linear regression line, was found between age at implantation and average difference limen ($R^2 = 25.6\%$, $p = 0.046$), indicating that subjects who were implanted at a later age had larger average limens. The correlations between electrode discrimination and the two variables of duration of deafness and age at testing approached significance ($R^2 = 22.4\%$, $p = 0.064$ and $R^2 = 22.2\%$,

Figure 1. Electrode difference limens from the 16 subjects for the apical, mid, and basal reference electrodes. The error bars show the standard errors over the different amounts of current level variation used. The asterisks indicate a significant difference between the three positions.

Figure 2. Average electrode difference limens as functions of the six subject variables. Given in each graph are the correlation coefficient and level of significance. A linear regression line has been fitted in the one case of a significant correlation.

$p = 0.065$, respectively), which was consistent with the high correlations between these two subject variables and age at implantation (Table 3). Note that the limens for the three meningitic subjects (P9, P10, and P12) were less than 2.5 electrodes, and these limens were similar to those from a number of congenitally deaf subjects (Table 4). This suggested that any degenerative processes as a consequence of meningitis did not markedly influence this measure of electrode discrimination for these three subjects.

The electrode difference limens for each of the three reference positions were separately correlated with age at implantation to determine whether there were any differences with respect to location in the cochlea (Fig. 3). A significant positive correlation, as shown by the linear regression line, was recorded for the basal limen ($R^2 = 35.7\%$, $p = 0.015$), indicating that subjects implanted at an older age had larger limens at the basal position. The basal limens also positively correlated with duration of deafness ($R^2 = 29.4\%$, $p = 0.030$) and age at testing ($R^2 = 27.7\%$, $p = 0.036$), consistent with the high correlations between these two subject variables and age at implantation (Table 3). Note that the limens for the apical and mid positions did not correlate with any subject variables ($R^2 < 18.9\%$, $p > 0.092$). Thus these findings indicated that between-subject variability in electrode discrimination at only the basal portion of the cochlea was related to the duration of auditory deprivation.

Figure 4 shows the average difference limens

across the three positions as functions of the eight measures of speech perception. Table 4 also shows the scores for the Picture Vocabulary and BKB sentence tests in the audition alone condition. Significant negative correlations, as shown by the linear regression lines, were found for the two closed-

Figure 3. Electrode difference limens at each of the three reference electrodes as functions of age at implantation. Given in each graph are the correlation coefficient and level of significance. A linear regression line has been fitted in the one case of a significant correlation.

Figure 4. Speech perception scores as functions of the average electrode difference limen. Given in each graph are the correlation coefficient and level of significance. Linear regression lines have been fitted in the two cases of a significant correlation.

set tests, Picture Vocabulary ($R^2 = 38.7\%$, $p = 0.018$) and NU-CHIPS ($R^2 = 52.0\%$, $p = 0.028$), indicating that subjects with larger limens had lower scores. Correlations with scores for the open-set BKB sentence and PBK word tests were not significant ($R^2 < 21.0\%$, $p > 0.134$). Scores for the two closed-set tests were positively correlated with each other ($R^2 = 88.0\%$, $p = 0.001$), and were also negatively correlated with three subject variables as shown in Table 5: duration of deafness, age at implantation, and age at testing. These three subject variables were highly correlated with each other (Table 3).

To determine the contributions of subject variables and the average difference limens to the closed-set speech scores, two-predictor multiple regression models were fitted to the data, each test separately. Age at testing was taken as the representative subject variable as the correlations between this variable and the speech scores were higher than those of the other variables (Table 5). The models were:

Picture Vocabulary score (%) = $131 - 3.26$ (age at testing) - 5.23 (average limen)

NU-CHIPS score (%) = $126 - 3.59$ (age at testing) - 5.20 (average limen)

For the Picture Vocabulary test, the model ac-

counted for 74.7% of the variance ($p = 0.001$), but only the coefficient for age at testing was significant: age at testing ($t = -3.96$, $p = 0.002$) and average limen ($t = -1.82$, $p = 0.096$). Also calculated was the independent contribution of the average limen after partialing out the contribution of age at testing. The average limen accounted for an additional 23.1% of the variance, but the difference between a one-predictor model using age at testing and the two-

TABLE 5. Correlation coefficients, and the corresponding levels of significance, between subject variables and scores for the Picture Vocabulary and NU-CHIPS tests

Subject variable	Picture Vocabulary	NU-CHIPS
Age at onset of deafness	$R^2 = 19.1\%$ $p = 0.118$	$R^2 = 17.5\%$ $p = 0.262$
Duration of deafness	$R^2 = 60.0\%$ $p = 0.001$ ()	$R^2 = 66.2\%$ $p = 0.008$ (-)
Age at implantation	$R^2 = 58.7\%$ $p = 0.001$ (-)	$R^2 = 67.1\%$ $p = 0.007$ ()
Duration of implant use	$R^2 = 3.8\%$ $p = 0.504$	$R^2 = 2.4\%$ $p = 0.694$
Total time of auditory experience	$R^2 = 11.3\%$ $p = 0.240$	$R^2 = 7.1\%$ $p = 0.490$
Age at testing	$R^2 = 67.1\%$ $p < 0.001$ ()	$R^2 = 69.4\%$ $p = 0.005$ (-)

The () indicates a significant negative correlation.

Figure 5. Speech perception scores for the Picture Vocabulary and NU-CHIPS tests as functions of the electrode difference limens at each of the three reference electrodes. Given in each graph are the correlation coefficient and level of significance. Linear regression lines have been fitted in the two cases of a significant correlation.

predictor model was not significant [$F(1,11) = 3.91$, $p = 0.074$]. For the NU-CHIPS test, the model accounted for 75.3% of the variance ($p = 0.015$), but both coefficients were not significant: age at testing ($t = -2.38$, $p = 0.055$) and average limen ($t = -1.20$, $p = 0.277$). This suggested multicollinearity between variables and the correlation between the average difference limen and age at testing, for the nine subjects where a score for NU-CHIPS was recorded, approached significance ($R^2 = 41.2\%$, $p = 0.062$). An additional 19.3% of the variance was accounted for by the average limen after partialing out the contribution of age at testing, but the difference between a one-predictor model using age at testing and the two-predictor model was not significant [$F(1,6) = 2.63$, $p = 0.156$]. Similar findings were found when using the other two variables that significantly correlated with these speech scores: age at implantation and duration of deafness (Table 5).

Scores for the Picture Vocabulary and NU-CHIPS tests were separately correlated with the electrode difference limens at each of the three reference positions to determine whether there were any differences with respect to location in the cochlea (Fig. 5). Significant negative correlations were recorded between both speech tests and the apical position ($R^2 > 41.3\%$, $p < 0.021$), indicating that subjects with larger limens had lower speech scores. However, this finding should be interpreted with caution as the correlation was influenced by low scores from

two subjects with large limens. The correlations between the speech scores for these two tests and the mid and basal positions were not significant ($R^2 < 30.4\%$, $p > 0.057$).

The difference limens for the apical position did not correlate with the three subject variables that also correlated with scores for the two closed-set tests, which are shown in Table 5. To determine the contributions of subject variables and the apical limen to the speech scores, two-predictor multiple regression models were fitted to the data, and age at testing was taken as the representative subject variable. The models were:

Picture Vocabulary score (%) = $128 - 3.23$ (age at testing) $- 4.01$ (apical limen)

NU-CHIPS score (%) = $123 - 3.42$ (age at testing) $- 4.19$ (apical limen)

For the Picture Vocabulary test, the model accounted for 77.6% of the variance ($p < 0.001$), and the coefficients for each predictor variable were significant: age at testing ($t = -4.24$, $p = 0.001$) and apical limen ($t = -2.27$, $p = 0.044$). An additional 31.6% of the variance was accounted for by the apical limen after partialing out the contribution of age at testing, and the difference between a one-predictor model using age at testing and the two-predictor model was significant [$F(1,11) = 6.01$, $p = 0.032$]. For the NU-CHIPS test, the model accounted for 78.7% of the variance ($p = 0.010$), but only the coefficient for age at testing was significant: age at testing ($t = -2.53$, $p = 0.044$) and apical limen ($t =$

-1.62, $p = 0.157$). An additional 30.9% of the variance was accounted for by the apical limen after partialing out the contribution of age at testing, but the difference between a one-predictor model using age at testing and the two-predictor model was not significant [$F(1,6) = 4.80, p = 0.071$]. Multicollinearity between variables again possibly influenced these findings as the correlation between the apical limen and age at testing for the nine subjects with a NU-CHIPS score approached significance ($R^2 = 36.8\%, p = 0.083$). These results were similar to those when using the other two variables that significantly correlated with these speech scores: age at implantation and duration of deafness (Table 5).

DISCUSSION

The main findings of this study were that most of the early-deafened subjects were able to successfully discriminate electrodes, with average difference limens across the three positions on the array of less than two electrodes for 75% of subjects, but there were significant differences between the three positions for 69% of subjects. The average limens and those at the basal position were positively related to the duration of auditory deprivation, with larger limens for subjects implanted at a later age and with a longer duration of deafness. The average limens and those at the apical position were negatively related to closed-set speech perception scores, with lower scores for subjects with larger limens, but no relationship was found between the limens and open-set speech scores. In addition, the closed-set speech perception scores were negatively related to the duration of auditory deprivation, with lower scores for subjects implanted at a later age and with a longer duration of deafness.

Although most of the early-deafened subjects successfully discriminated between electrodes, performance was not uniform across the three positions on the array for over half the subjects (Fig. 1 and Table 4). Difference limens were larger at either the apical or basal position in almost all cases where a significant difference between the three positions was found. The one exception was P13 where limens at the mid position were larger than those at the apical and basal positions. Although some of these differences across the three positions were less than one electrode, there were cases in which the differences were three electrodes or greater. These findings were consistent with those reported for postlinguistically deaf adults, where variability in the perception of differences between electrodes at different locations on the Cochlear Limited array has also been found (Busby et al., 1994; Collins et al., 1997; Cohen et al., 1996; Henry et al., 1997; Nelson et al.,

1995; Zwolan et al., 1997). This variability in electrode perception along the array for both early-deafened and postlinguistically deaf adult subjects could be related to variability in spiral ganglion survival as a function of location in the cochlea. It has been shown that electrode discrimination in the monkey (Pfungst, Glass, Spelman, & Sutton, 1985) and pitch ranking in the human (Clark et al., 1988) were poorer at locations in the cochlea with very poor spiral ganglion survival. It is also possible that these differences along the array could be influenced by the integrity of the auditory pathway and cortex (Heffner & Heffner, 1986).

The significant correlation between the average limen and age at implantation (Fig. 2) suggested that electrode discrimination may be sensitive to the effects of auditory deprivation. The results also suggested that this relationship was most apparent at the basal portion of the array rather than the apical and mid locations (Fig. 3). Note that our findings do not indicate that electrode discrimination at the apical and mid positions was unaffected by the consequences of auditory deprivation, but that between-subject variability in the apical and mid limens was not related to the (indirect) measures of auditory deprivation used in our study. It may be the case that spiral ganglion cell survival in the basal portion of the cochlea is sensitive to auditory deprivation in early-deafened subjects. It is also possible that performance at the basal position may have been influenced by the consequences of electrode insertion trauma, such as fibrous tissue growth and new bone formation (Nadol, 1997; Shepherd, Matsushima, Martin, & Clark, 1994). These factors would change the electrical resistance between the active and return electrodes, and between stimulating electrodes and stimulated nerve fibers, which could produce abnormal current flow to more distant fibers (Clark et al., 1988), and therefore influence electrode discrimination. It is interesting to note that a negative correlation between the basal limens (Table 4) and the dynamic range of hearing for basal electrodes (Table 2) approached significance ($R^2 = 24.6\%, p = 0.051$). Thus, limens tended to be larger for subjects with smaller dynamic ranges. It is possible that these two psychophysical measures are influenced by similar underlying mechanisms, such as spiral ganglion cell survival and the consequences of insertion trauma. For the apical and mid positions, however, the negative correlations between limens and dynamic ranges were not significant ($R^2 = 10.3\%, p = 0.225$ and $R^2 = 22.6\%, p = 0.063$, respectively).

The lack of a significant correlation between the difference limens and postoperative experience suggested that any stimulation-induced changes did not

influence electrode discrimination in our subjects. Some studies in the neonatally deafened animal have suggested that chronic electric stimulation may lead to an increase in spiral ganglion cell survival at the site of stimulation (Leake, Hradek, Rebscher, & Snyder, 1991; Leake, Hradek, & Snyder, 1999). However, it is not known whether similar increases in cell survival occur in human subjects.

The significant correlations between the average difference limens and those at the apical position with speech perception scores for the two closed-set tests (Fig. 5) suggested that a considerable amount of speech information is conveyed by stimulation of electrodes along the array. These findings were consistent with those from Henry et al. (1997) who found significant correlations between electrode discrimination at the apical and mid positions on the array and the phoneme scores for an open-set monosyllabic word test by postlinguistically deaf adults using the Cochlear Limited prosthesis. Although the results of our study and those of Henry et al. (1997) suggest that the ability to discriminate electrodes is an important factor in speech perception, the tests differed across the two studies. The lack of significant correlations between electrode discrimination and open-set speech perception scores in our study (Fig. 4) may have been because of the greater contributions of other variables, such as the language skills and receptive vocabulary of early-deafened subjects (Sarant, Blamey, Cowan, & Clark, 1997), when compared with postlinguistically deaf adults.

In our study, a greater contribution to between-subject variability in the closed-set speech scores was from three subject variables (Table 5) that were also highly correlated with each other: age at implantation, duration of deafness, and age at testing (Table 3). It is likely that the two variables related to auditory deprivation were the more relevant factors influencing these speech scores than age at testing because the speech tests were age-appropriate for our subjects. Note that because of the greater contribution of subject variables, the contribution of the difference limens was reduced when using regression models with two predictor variables.

The correlations between the apical limens and closed-set speech scores were also influenced by outliers, thus limiting the significance of the relationship between electrode discrimination and speech perception. There are, however, alternative interpretations of these results. It is possible that the ability to discriminate between electrodes does not markedly influence closed-set speech perception, at least for these two speech tests, unless electrode discrimination was poor, greater than three to four electrodes. This would be consistent with data from postlinguistically deaf adults, which suggested that

consonant identification was more dependent on temporal information than on spectral information (Nelson et al., 1995). Therefore, the spread of speech scores for subjects with limens of less than two electrodes (Figs. 4 and 5) may be related to the perception of temporal speech information, and that the poor performance of the subjects with the large limens could be related to the perception of spectral information.

One factor that limited our findings was that electrode discrimination was measured at only three positions on the array. It is likely that an increase in the number of reference electrodes along the array would provide a more accurate measure of discrimination performance to relate to measures of auditory deprivation and experience, and to correlate with speech perception scores. However, it may not be realistic to collect this amount of data from younger early-deafened subjects, where attention to the task can vary. For instance, it would be necessary to obtain nine difference limens if the reference electrodes were every second electrode on the array, between electrodes 20 and 4 in BP+1 stimulation, which would then be multiplied by a factor of at least two when collecting data for two or more different amounts of random variation in current level. As younger children are now receiving multiple-electrode cochlear implants, more efficient methods of data collection are required so that electrode discrimination can be measured for a greater number of reference electrodes.

The information from electrode discrimination studies could also be incorporated into speech processing strategies so that speech information is presented using perceptually distinct channels. It would also be important to ensure that electrode perception is measured at several locations on the array so that a reasonable estimate of variability is obtained. The usefulness of this approach has been shown with postlinguistically deaf adults using the Cochlear Limited prosthesis, where scores for modified speech processing strategies that only used perceptually distinct electrodes were higher than scores for the standard strategies using all available electrodes, without training (Zwolan et al., 1997). It is possible that further improvements in performance could be obtained after training and experience (Tyler, Preece, Lansing, Otto, & Gantz, 1986). The performance of early-deafened subjects using a similar approach, however, has yet to be determined and would be an important topic for future research.

ACKNOWLEDGMENTS:

This work was supported by the National Health and Medical Research Council of Australia, the Human Communication Re-

search Centre at the Department of Otolaryngology, University of Melbourne, and Lions Clubs International. We would like to thank the anonymous reviewers for their constructive comments, Mark Harrison for assistance in software development, and the Cochlear Implant Clinic of the Royal Victorian Eye Ear Hospital, Melbourne, for access to clinic data.

Address for correspondence: Peter A. Busby, Ph.D. Department of Otolaryngology, University of Melbourne, Parkville, 3052, Victoria, Australia.

Received March 23, 1999; accepted March 1, 2000

REFERENCES

- Bench, R. J., & Bamford, J. (Eds.) (1979). *Speech-Hearing Tests and the Spoken Language of Hearing-Impaired Children*. London: Academic.
- Busby, P. A., & Clark, G. M. (1996). Electrode discrimination by early-deafened cochlear implant patients. *Audiology*, 35, 8–22.
- Busby, P. A., Tong, Y. C., & Clark, G. M. (1992). Psychophysical studies using a multiple-electrode cochlear implant in patients who were deafened early in life. *Audiology*, 31, 95–111.
- Busby, P. A., Tong, Y. C., & Clark, G. M. (1993). Electrode position, repetition rate, and speech perception by early- and late-deafened cochlear implant patients. *Journal of the Acoustical Society of America*, 93, 1058–1067.
- Busby, P. A., Whitford, L. A., Blamey, P. J., Richardson, L. M., & Clark, G. M. (1994). Pitch perception for different modes of stimulation using the Cochlear multiple-electrode prosthesis. *Journal of the Acoustical Society of America*, 95, 2658–2669.
- Carhart, R., & Jerger, J. (1959). Preferred method for clinical determination of pure tone thresholds. *Journal of Speech and Hearing Disorders*, 24, 330–345.
- Clark, G. M., Blamey, P. J., Busby, P. A., Dowell, R. C., Franz, B. K., Musgrave, G. N., Nienhuys, T. G., Pyman, B. C., Roberts, S. A., Tong, Y. C., Webb, R. L., Kuzma, J. A., Money, D. K., Patrick, J. F., & Seligman, P. M. (1987). A multiple-electrode intracochlear implant for children. *Archives of Otolaryngology*, 113, 825–828.
- Clark, G. M., Shepherd, R. K., Franz, B. K.-H., Dowell, R. C., Tong, Y. C., Blamey, P. J., Webb, R. L., Pyman, B. C., McNaughtan, J., Bloom, D. M., Kakulas, B. A., & Siejka, S. (1988). The histopathology of the human temporal bone and auditory nervous system following cochlear implantation in a patient. *Acta Otolaryngologica. Supplement (Stockholm)*, 488, 1–65.
- Cohen, L. T., Busby, P. A., Whitford, L. A., & Clark, G. M. (1996). Cochlear implant place psychophysics. 1. Pitch estimation with deeply inserted electrodes. *Audiology and Neuro-Otology*, 1, 265–277.
- Collins, L. M., Zwolan, T. A., & Wakefield, G. H. (1997). Comparison of electrode discrimination, pitch ranking, and pitch scaling data in postlingually deafened adult cochlear implant subjects. *Journal of the Acoustical Society of America*, 101, 440–455.
- Dorman, M. F., Smith, L., McCandless, G. A., Dunnivant, G., Parkin, J., & Dankowski, K. (1990). Pitch scaling and speech understanding by patients who use the Ineraid cochlear implant. *Ear and Hearing*, 11, 310–315.
- Eddington, D. K. (1980). Speech discrimination in deaf subjects with cochlear implants. *Journal of the Acoustical Society of America*, 68, 885–891.
- Eddington, D. K., Dobelle, W. H., Brackman, D. E., Mladejovsky, M. G., & Parkin, J. L. (1978). Auditory prostheses research with multiple channel intracochlear stimulation in man. *Annals of Otology, Rhinology, and Laryngology, Suppl.*, 53, 5–39.
- Elliott, L. L., & Katz, D. R. (1980). *Northwestern University Children's Perception of Speech (NU-CHIPS)*. St. Louis: Auditec.
- Fryauf-Bertschy, H., Tyler, R. S., Kelsay, D. M. R., Gantz, B. J., & Woodworth, G. G. (1997). Cochlear implant use by prelingually deafened children: The influences of age at implant and length of device use. *Journal of Speech, Language, and Hearing Research*, 40, 183–199.
- Hardie, N. A., Martsi-McClintock, A., Aitkin, L. M., & Shepherd, R. K. (1998). Neonatal sensorineural hearing loss affects synaptic density in the auditory midbrain. *NeuroReport*, 9, 2019–2022.
- Harrison, R. V., Nagasawa, A., Smith, D. W., Stanton, S., & Mount, R. J. (1991). Reorganization of auditory cortex after neonatal high frequency cochlear hearing loss. *Hearing Research*, 54, 11–19.
- Hartmann, R., Shepherd, R. K., Heid, S., & Klinke, R. (1997). Response of the primary auditory cortex to electrical stimulation of the auditory nerve in the congenitally deaf white cat. *Hearing Research*, 112, 115–133.
- Hashisaki, G. T., & Rubel, E. W. (1989). Effects of unilateral cochlea removal on anteroventral cochlear nucleus neurons in developing gerbils. *Journal of Comparative Neurology*, 303, 2–15.
- Heffner, H. E., & Heffner, R. S. (1986). Hearing loss in Japanese macaques following bilateral auditory cortex lesions. *Journal of Neurophysiology*, 55, 256–271.
- Heid, S., Jahn-Siebert, T. K., Klinke, R., Hartmann, R., & Langner, G. (1997). Afferent projection patterns in the auditory brainstem in normal and congenitally deaf white cats. *Hearing Research*, 110, 191–199.
- Henry, B., McKay, C., McDermott, H., & Clark, G. (1997). Speech cues for cochlear implantees: spectral discrimination. In G. M. Clark (Ed.), *Cochlear Implants, Proceedings of the XVI World Congress of Otorhinolaryngology, Head and Neck Surgery* (pp. 89–93). Bologna, Italy: Monduzzi Editore.
- Leake, P. A., Hradek, G. T., Rebscher, S. J., & Snyder, R. L. (1991). Chronic intracochlear electrical stimulation induces selective survival of spiral ganglion cells in neonatally deafened cats. *Hearing Research*, 54, 251–271.
- Leake, P. A., Hradek, G. T., & Snyder, R. L. (1999). Chronic electrical stimulation by a cochlear implant promotes survival of spiral ganglion neurons after neonatal deafness. *Journal of Comparative Neurology*, 412, 543–562.
- Levitt, H. (1971). Transformed up-down methods in psychoacoustics. *Journal of the Acoustical Society of America*, 49, 467–476.
- McDermott, H. J., McKay, C. M., & Vandali, A. E. (1992). A new portable sound processor for the University of Melbourne/Nucleus Limited multielectrode cochlear implant. *Journal of the Acoustical Society of America*, 91, 3367–3371.
- Moore, D. R. (1990). Auditory brainstem of the ferret: bilateral cochlear lesions in infancy do not effect the number of neurones projecting from the cochlear nucleus to the inferior colliculus. *Developmental Brain Research*, 54, 125–130.
- Nadol, J. B. (1997). Patterns of neural degeneration in the human cochlea and auditory nerve: implications for cochlear implantation. *Otolaryngology - Head and Neck Surgery*, 117, 220–228.
- Nelson, D. A., Van Tasell, D. J., Schroder, A. C., Soli, S., & Levine, S. (1995). Electrode ranking of "place pitch" and speech recognition in electrical hearing. *Journal of the Acoustical Society of America*, 98, 1987–1999.
- Pfingst, B. E., Glass, I., Spelman, F. A., & Sutton, D. (1985). Psychophysical studies of cochlear implants in monkeys: clinical implications. In R. A. Schindler & M. M. Merzenich (Eds.), *Cochlear Implants* (pp. 305–321). New York: Raven.

- Plant, G. (1984). A diagnostic speech test for severely and profoundly hearing-impaired children. *Australian Journal of Audiology*, 6, 1-9.
- Sarant, J. Z., Blamey, P. J., Cowan, R. S., & Clark, G. M. (1997). The effect of language knowledge on speech perception—what are we really assessing. *American Journal of Otolaryngology*, 18(Suppl.), S135-S137.
- Shannon, R. V. (1983). Multichannel electrical stimulation of the auditory nerve in man. I. Basic psychophysics. *Hearing Research*, 11, 157-189.
- Shepherd, R. K., Matsushima, J., Martin, R. L., & Clark, G. M. (1994). Cochlear pathology following chronic electrical stimulation of the auditory nerve. II: Deafened kittens. *Hearing Research*, 81, 150-166.
- Skinner, M. W., Clark, G. M., Whitford, L. A., Seligman, P. M., Staller, S. J., Shipp, D. B., Shalloo, J. K., Everingham, C., Menapace, C. M., Arndt, P. L., Antogenelli, T., Brimacombe, J. A., Pijl, S., Daniels, P., George, C. R., McDermott, H. J., & Beiter, A. L. (1994). Evaluation of a new spectral peak coding strategy for the Nucleus 22 channel cochlear implant system. *American Journal of Otolaryngology*, 15(Suppl. 2), 15-27.
- Skinner, M. W., Holden, L. K., Holden, T. A., Dowell, R. C., Seligman, P. M., Brimacombe, J. A., & Beiter, A. L. (1991). Performance of postlinguistically deaf adults with the wearable speech processor (WSP III) and mini speech processor (MSP) of the Nucleus multi-electrode cochlear implant. *Ear and Hearing*, 12, 3-22.
- Snyder, R. L., Rebscher, S. J., Cao, K., Leake, P. A., & Kelly, K. (1990). Chronic intracochlear electrical stimulation in the neonatally deafened cat. I. Expansion of central representation. *Hearing Research*, 50, 7-34.
- Tierney, T. S., Russell, F. A., & Moore, D. R. (1997). Susceptibility of developing cochlear nucleus neurons to deafferentation-induced death abruptly ends just before the onset of hearing. *Journal of Comparative Neurology*, 378, 295-306.
- Tong, Y. C., Blamey, P. J., Dowell, R. C., & Clark, G. M. (1983). Psychophysical studies evaluating the feasibility of a speech processing strategy for a multiple-channel cochlear implant. *Journal of the Acoustical Society of America*, 74, 73-80.
- Tong, Y. C., Busby, P. A., & Clark, G. M. (1988). Perceptual studies on cochlear implant patients with early onset of profound hearing impairment prior to the normal development of auditory, speech, and language skills. *Journal of the Acoustical Society of America*, 84, 951-962.
- Tong, Y. C., & Clark, G. M. (1985). Absolute identification of electric pulse rates and electrode positions by cochlear implant patients. *Journal of the Acoustical Society of America*, 77, 1881-1888.
- Tong, Y. C., Clark, G. M., Blamey, P. J., Busby, P. A., & Dowell, R. C. (1982). Psychophysical studies for two multiple-channel cochlear implant patients. *Journal of the Acoustical Society of America*, 71, 153-160.
- Tyler, R. S., Opie, J. M., Fryauf-Bertschy, H., & Gantz, B. J. (1992). Future directions for cochlear implants. *Journal of Speech-Language Pathology and Audiology*, 16, 151-164.
- Tyler, R. S., Preece, J. P., Lansing, C. R., Otto, S. R., & Gantz, B. J. (1986). Previous experience as a confounding factor in comparing cochlear-implant processing schemes. *Journal of Speech and Hearing Research*, 29, 282-287.
- Wilson, B. S., Finley, C. C., Lawson, D. T., Wolford, R. D., Eddington, D. K., & Rabinowitz, W. M. (1991). New levels of speech recognition with cochlear implants. *Nature*, 352, 236-238.
- Zwolan, T. A., Collins, L. M., & Wakefield, G. H. (1997). Electrode discrimination and speech recognition in postlingually deafened adult cochlear implant subjects. *Journal of the Acoustical Society of America*, 102, 3673-3685.

REFERENCE NOTES

- 1 Proceedings of the NIH Consensus Development Conference on Cochlear Implants in Adults and Children (1995). National Institutes of Health: Bethesda, Maryland.
- 2 Haskins, H. (1949). A phonetically balanced test of speech discrimination for children. Unpublished thesis, Northwestern University.

Minerva Access is the Institutional Repository of The University of Melbourne

Author/s:

Busby, P. A.;Clark, Graeme M.

Title:

Electrode discrimination by early-deafened subjects using the Cochlear Limited multiple electrode cochlear implant

Date:

2000

Citation:

Busby, P. A., & Clark, G. M. (2000). Electrode discrimination by early-deafened subjects using the Cochlear Limited multiple electrode cochlear implant. *Ear and Hearing*, 21, 291-304.

Persistent Link:

<http://hdl.handle.net/11343/28757>