Title Page

Title: Three Dimensional Printing in Orthopaedic Surgery: Review of Current and Future Applications

Short Title: Three Dimensional Printing in Orthopaedic Surgery

Authors

Mulford JS, MBBS(Hons) FRACS. Launceston General Hospital

Babazadeh S, MBBS. Launceston General Hospital

Mackay N, MBBS, Launceston General Hospital

Key words: Orthopaedics, Printing, Orthopedics, Printing, Three-Dimensional.

- The corresponding author is not in recipient of a research scholarship.
- The paper was presented to the Australian Knee Society October 17th Byron Bay.
- There are 6 figures included.
- Word Count Abstract: 66
- Word count text: 1391

Corresponding author:

Jonathan Mulford jonathanmulford1971@gmail.com

This is the author manuscript accepted for publication and has undergone full peer review but has not been through the copyediting, typesetting, pagination and proofreading process, which may lead to differences between this version and the Version of Record. Please cite this article as doi: [10.1111/ans.13533](https://doi.org/10.1111/ans.13533)

170 St John Street Launceston, Tasmania. 0363347332. Fax

The corresponding author is responsible for the integrity of the study and for communicating with the other authors about editorial decisions and the nature of any revisions

Abstract

Three-dimensional (3D) printing is a rapidly evolving technology with the potential for significant contributions to surgical practice. There are many current applications for 3D printed technology with future applications being explored. This technology has applications in preoperative planning, education, custom manufacturing (implants, prosthetics and surgical guides) and exciting potential for biological applications. This article reviews the current and future applications of 3D technology in orthopaedic surgery.

Word Count 66.

Introduction

3D technology was first introduced two decades ago, when it was seen as unobtainable, expensive, and futuristic with limited clinical application [1-3]. In recent times there has been an explosion in 3D printing technology applications. Its use is becoming more widespread in surgery and the technology will likely play a pivotal role in orthopaedic practice.

What is 3D printing?

Three dimensional printing converts a computer generated 3D image into a physical model. It is also referred to as rapid prototyping or additive manufacturing [4]. The object is built one thin layer at a time. Typically the manufacture of 3D models can be based on 3D Digital Imaging and Communications in Medicine (DICOM) format data from CT or MRI images.

The printer builds the model by a series of cross sectional layers of either liquid, powder or sheet material such as plastics/polymers and metals. The final shape is created when the layers are joined [4]. This process can be used to create unique patient-specific materials and may be more cost effective than conventional manufacturing of implants[5]. 3D printing can create any complex shape and allows solid and porous sections to be combined to provide optimal strength and performance [4]. Numerous different materials are utilised including plastics, polymers, glass, ceramic, metals [4, 6, 7] and biological materials [8].

Educational Applications

Three-dimensional models can be used to educate patients, students, trainees and surgeons. The visual 3D model can play an important role in the discussion and consent process for complex surgical cases. The model assists in demonstrating the complexity and severity of injuries to the patient and family, as well as help understand the nature of the procedure. These models can then also be used for teaching students and trainees [9, 10].

Surgical Planning

Currently, 3D printing widely available in surgical planning for a variety of orthopaedic procedures. These cases vary from complex fracture patterns to revision arthroplasty surgery.

Three-dimensional printing is especially useful in complex trauma cases. The 3D printed models provide a visual and tactile aid in conceptualising complex fracture patterns [11]. The model can be sterilised and reviewed intra-operatively as required [12, 13]. Preoperative review of the 3D model can allow the surgeon to anticipate intra-operative difficulties, selection of optimal surgical approach and the need for specific equipment. Challenging pelvic fractures provide an example of these concepts (Figure 1) [13-17].

Paediatric orthopaedic surgeons [12, 18] have utilised 3D printed models to assist in the management of complex spine scoliosis, coalition in the foot, Perthes and Blounts disease. The models were used to assist in preoperative planning, communication with the patient, for reference during surgery with reported improvements in the safety of the procedure and reducing operative time.

Simple and complex osteotomies can be planned using models preoperatively. The surgeon can study the deformity and plan the surgery with a computer model. This includes the exact placement of implants and the ideal osteotomy site. 3D printing can produce jigs to allow for predrilling of holes for customised plates with built-in osteotomy guides (Figure 2) [19].

There are published examples where 3D technology has been utilised in complex cases of upper limb [20-27] and lower limb osteotomies [19, 28] [29]. These

articles support the notion that this technology simplifies complex surgery, providing confidence that goals of surgery are being achieved and reduce operative time. Trials comparing routine preoperative planning with the use of 3D printing are required [30].

Preoperative computer-assisted planning and custom 3D printed guides have also been described in pedicle screw placement [31], mosaic arthroplasty [32] and musculoskeletal tumour excision [33, 34].

In Australia the 3D printed guides that have been discussed are readily available, however cost is an issue. Typically these guides do not have a rebate from either Medicare or insurance companies, as they are not implantable devices. The cost can be up to \$2000 AUD.

Most implant companies have 3D printed guides available to assist in standard knee joint arthroplasty [35]. A guide to assist with hip resurfacing has been described [36]. This process is commonly called Patient Specific Instrumentation (PSI). Patients have either CT scan or MRI scan to produce DICOM images. 3D images are then created and a preoperative surgical plan constructed to achieve perfect implant placement. Disposable cutting blocks are then fabricated to match and conform to the patient's anatomy using 3D printing technology. The proposed benefits include improved reproducibility of component alignment, reduced surgical time, and optimised efficiency and cost-effectiveness. Despite these proposed benefits it has yet to be proven to be better than standard techniques[37, 38]. The Australian National joint registry in 2014 estimated this technique was used in 6.8% of all primary total knee procedures. The costs associated with utilizing this this

technique are largely absorbed into the cost of the implant. There is a cost to acquire the imaging required to create the PSI as this requires an MRI or CTscan.

Custom Implants

3D printing has allowed the emergence of custom implants [39]. Customized implants for joint arthroplasty are useful when the patient does not fit the standard range of implant size or their disease pattern requires specific implant characteristics in order to achieve correction and function [4]. This has largely been utilised for revision implants in difficult cases (Figure 3). In some instances, revision surgery may not have been possible with traditional implants. DICOM images are utilised to create an anatomical model on a computer. An implant can then be custom designed to fit the anatomical problem. 3D printing has allowed for the manufacture of complex external and internal structures to in an attempt to improve osteointegration and accurately fit the defect site [40].

3D printing is also used in primary implants, for example the manufacture of primary tibial base-plate devices (Stryker -Kalamazoo, MI, USA) [41].

Many of the proposed benefits[42] remain unproven in comparison with traditional techniques at this stage, however there is a paucity of publications at this stage with a lack of long-term follow up in comparison to traditional techniques [43-45]. Onerecent article suggests higher manipulation rates following knee replacement surgery with custom implants with no short term benefits[46] however more research is required.

Custom implants can be extended to other body parts. Published case reports include custom calcaneus implant with pre-planned soft tissue attachments was created with 3D printing technology for an isolated calcaneus tumour [47](Figure 4). While custom implants are extremely useful in some cases, the technique is not considered the gold standard at this stage and therefore requires careful discussion with each individual patient.

Custom External Fixators and Orthotics

There has been utilisation of 3D printed customised external fixators to assist with fracture reduction [48]. It has been proposed that this technique could incorporate stress adjustment, healing factors and nanotechnology in the future to potentially optimise the fracture healing process.

Many 3D printed braces, splints and orthotics are now available commercially; however, there is no literature regarding the outcomes and benefits of these braces.

Future Applications

The use and application of 3D technology in orthopaedics is changing rapidly. Its use can be split into two broad categories. The first encompasses industrial applications of the process, resulting in complex models, prostheses, biological scaffolds and nanotechnology [49]. The second is in every-day clinical use; whereby it can be used for education and surgical planning.

At this stage the industrial uses of 3D printing in orthopaedics is associated with significant costs. The incorporation of this technology into every-day practice will be influenced by whether 3D printing techniques can be used to reduce current manufacturing costs, simplify surgery and ultimately improve surgical outcomes. Everyday clinical use of 3D printers for education and surgical planning is becoming more feasible. The cost of 3D printers has reduced dramatically in recent times with free software available. Aside from the paucity of clinical trials, the main practical disadvantage is the length of time taken to prepare images and print the model.

Biological tissue 3-D printing has been discussed for some time [50] and animal models demonstrate its feasibility [8, 51][52-54]. This technology is able to print cells or incorporate cells into a 3D scaffold to allow implantation within the body. Examples include regeneration of liver tissue [55], customizable blood vessels [56], skin, bone and cartilage [49] and scaffolds that could include delivery systems for healing factors or nanoparticles [49]. Regulatory barriers limit its widespread clinical use [40, 57] as there are ethical dilemmas with the use of stem cells. Translation of the discoveries in the laboratory to human patients still required more high quality trials [58, 59].

Current challenges for 3D cell printing include long-term stability and vascularisation of the engineered tissue [45, 60, 61]. Currently no option of printing the cells in the operating room exists as biological printers at this time are large requiring sophisticated computers. These techniques however have the ability to

transform the treatment of musculoskeletal conditions that currently have limited options.

3D printed bone scaffolds are being utilised to optimise the possibility of union particularly where bone defects exist and atrophic non-unions have occurred. Bone repair scaffolds can be constructed/generated to fit very specific defects [62] whilst growth factors and cell types can be introduced to enhance healing [40, 63-65].

Osteochondral and chondral graft options are also being investigated with multiple osteochondral tissue equivalents being considered in animal models [66, 67] [45]. A sheep model has shown incorporation of 3D printed meniscal grafts [68] (Figure 5).

Conclusion

3D printing applications in orthopaedic surgery are exciting and have the potential to alter future orthopaedic practice. The application of 3D technology in orthopaedics is only recent and hence there is a lack of long term follow up in comparison to traditional techniques. Limitations such as the quality of the model, cost of production and production time are likely to continue to evolve and improve. 3D printing applications have the potential to improve surgical outcomes via surgical education, pre-operative planning, custom manufacturing (implants, prosthetics and surgical guides) with future biological applications.

References

1. Brown GA, Firoozbakhsh K, DeCoster TA, Reyna JR, Jr., Moneim M: **Rapid prototyping: the future of trauma surgery?** *J Bone Joint Surg Am* 2003, **85-A Suppl 4**:49-55.
2. Goffin J, Van Brussel K, Martens K, Vander Sloten J, Van Audekercke R, Smet MH: **Three-dimensional computed tomography-based, personalized drill guide for posterior cervical stabilization at C1-C2.** *Spine (Phila Pa 1976)* 2001, **26**(12):1343-1347.
3. Peters P, Langlotz F, Nolte LP: **Computer assisted screw insertion into real 3D rapid prototyping pelvis models.** *Clin Biomech (Bristol, Avon)* 2002, **17**(5):376-382.
4. Rengier F, Mehndiratta A, von Tengg-Kobligh H, Zechmann CM, Unterhinninghofen R, Kauczor HU, Giesel FL: **3D printing based on imaging data: review of medical applications.** *Int J Comput Assist Radiol Surg* 2010, **5**(4):335-341.
5. Kern R: **3D printed implants hit the market, pave the way for more personalized devices.** *The Gray Sheet* 2013, **39**(44).
6. Fuller SM, Butz DR, Vevang CB, Makhoul MV: **Application of 3-dimensional printing in hand surgery for production of a novel bone reduction clamp.** *J Hand Surg Am* 2014, **39**(9):1840-1845.
7. Salmi M, Paloheimo KS, Tuomi J, Wolff J, Makitie A: **Accuracy of medical models made by additive manufacturing (rapid manufacturing).** *J Craniomaxillofac Surg* 2013, **41**(7):603-609.
8. Fedorovich NE, Alblas J, Hennink WE, Oner FC, Dhert WJ: **Organ printing: the future of bone regeneration?** *Trends Biotechnol* 2011, **29**(12):601-606.
9. Michalski MH, Ross JS: **The shape of things to come: 3D printing in medicine.** *JAMA* 2014, **312**(21):2213-2214.
10. Guitton TG, Brouwer K, Lindenhovius AL, Dyer G, Zurakowski D, Mudgal CS, Ring DC: **Diagnostic accuracy of two-dimensional and three-dimensional imaging and modeling of radial head fractures.** *J Hand Microsurg* 2014, **6**(1):13-17.
11. Rengier F, Mehndiratta A, Tengg-Kobligh HV, Zechmann C: **3D printing based on imaging data: review of medical application.** *International Journal of Computer Assisted Radiology & Surgery* 2010, **5**:335/341.
12. Starosolski ZA, Kan JH, Rosenfeld SD, Krishnamurthy R, Annapragada A: **Application of 3-D printing (rapid prototyping) for creating physical models of pediatric orthopedic disorders.** *Pediatr Radiol* 2014, **44**(2):216-221.
13. Chung KJ, Hong do Y, Kim YT, Yang I, Park YW, Kim HN: **Preshaping plates for minimally invasive fixation of calcaneal fractures using a real-size 3D-printed model as a preoperative and intraoperative tool.** *Foot Ankle Int* 2014, **35**(11):1231-1236.
14. Hurson C, Tansey A, O'Donnchadha B, Nicholson P, Rice J, McElwain J: **Rapid prototyping in the assessment, classification and preoperative planning of acetabular fractures.** *Injury* 2007, **38**(10):1158-1162.
15. Niikura T, Sugimoto M, Lee SY, Sakai Y, Nishida K, Kuroda R, Kurosaka M: **Tactile surgical navigation system for complex acetabular fracture surgery.** *Orthopedics* 2014, **37**(4):237-242.
16. Wu XB, Wang JQ, Zhao CP, Sun X, Shi Y, Zhang ZA, Li YN, Wang MY: **Printed three-dimensional anatomic templates for virtual preoperative planning before reconstruction of old pelvic injuries: initial results.** *Chin Med J (Engl)* 2015, **128**(4):477-482.
17. Jeong HS, Park KJ, Kil KM, Chong S, Eun HJ, Lee TS, Lee JP: **Minimally invasive plate osteosynthesis using 3D printing for shaft fractures of clavicles: technical note.** *Arch Orthop Trauma Surg* 2014, **134**(11):1551-1555.

18. Guarino J, Tennyson S, McCain G, Bond L, Shea K, King H: **Rapid prototyping technology for surgeries of the pediatric spine and pelvis: benefits analysis.** *J Pediatr Orthop* 2007, **27**(8):955-960.
19. Victor J, Premanathan A: **Virtual 3D planning and patient specific surgical guides for osteotomies around the knee: a feasibility and proof-of-concept study.** *Bone Joint J* 2013, **95-B**(11 Suppl A):153-158.
20. Zhang YZ, Lu S, Chen B, Zhao JM, Liu R, Pei GX: **Application of computer-aided design osteotomy template for treatment of cubitus varus deformity in teenagers: a pilot study.** *J Shoulder Elbow Surg* 2011, **20**(1):51-56.
21. Murase T, Oka K, Moritomo H, Goto A, Yoshikawa H, Sugamoto K: **Three-dimensional corrective osteotomy of malunited fractures of the upper extremity with use of a computer simulation system.** *J Bone Joint Surg Am* 2008, **90**(11):2375-2389.
22. Oka K, Murase T, Moritomo H, Yoshikawa H: **Corrective osteotomy for malunited both bones fractures of the forearm with radial head dislocations using a custom-made surgical guide: two case reports.** *J Shoulder Elbow Surg* 2012, **21**(10):e1-8.
23. Miyake J, Murase T, Oka K, Moritomo H, Sugamoto K, Yoshikawa H: **Computer-assisted corrective osteotomy for malunited diaphyseal forearm fractures.** *J Bone Joint Surg Am* 2012, **94**(20):e150.
24. Stockmans F, Dezillie M, Vanhaecke J: **Accuracy of 3D Virtual Planning of Corrective Osteotomies of the Distal Radius.** *J Wrist Surg* 2013, **2**(4):306-314.
25. Omori S, Murase T, Kataoka T, Kawanishi Y, Oura K, Miyake J, Tanaka H, Yoshikawa H: **Three-dimensional corrective osteotomy using a patient-specific osteotomy guide and bone plate based on a computer simulation system: accuracy analysis in a cadaver study.** *Int J Med Robot* 2014, **10**(2):196-202.
26. Takeyasu Y, Oka K, Miyake J, Kataoka T, Moritomo H, Murase T: **Preoperative, computer simulation-based, three-dimensional corrective osteotomy for cubitus varus deformity with use of a custom-designed surgical device.** *J Bone Joint Surg Am* 2013, **95**(22):e173.
27. Storelli DA, Bauer AS, Lattanza LL, McCarroll HR, Jr.: **The use of computer-aided design and 3-dimensional models in the treatment of forearm malunions in children.** *Tech Hand Up Extrem Surg* 2015, **19**(1):23-26.
28. Schwartz A, Money K, Spangehl M, Hattrup S, Claridge RJ, Beauchamp C: **Office-based rapid prototyping in orthopedic surgery: a novel planning technique and review of the literature.** *Am J Orthop (Belle Mead NJ)* 2015, **44**(1):19-25.
29. Dobbe JG, Pre KJ, Kloen P, Blankevoort L, Streekstra GJ: **Computer-assisted and patient-specific 3-D planning and evaluation of a single-cut rotational osteotomy for complex long-bone deformities.** *Med Biol Eng Comput* 2011, **49**(12):1363-1370.
30. Leong NL, Buijze GA, Fu EC, Stockmans F, Jupiter JB, Distal Radius Malunion collaborative g: **Computer-assisted versus non-computer-assisted preoperative planning of corrective osteotomy for extra-articular distal radius malunions: a randomized controlled trial.** *BMC Musculoskelet Disord* 2010, **11**:282.
31. Owen BD, Christensen GE, Reinhardt JM, Ryken TC: **Rapid prototype patient-specific drill template for cervical pedicle screw placement.** *Comput Aided Surg* 2007, **12**(5):303-308.
32. Kunz M, Waldman SD, Rudan JF, Bardana DD, Stewart AJ: **Computer-assisted mosaic arthroplasty using patient-specific instrument guides.** *Knee Surg Sports Traumatol Arthrosc* 2012, **20**(5):857-861.

33. Tam MD, Laycock SD, Bell D, Chojnowski A: **3-D printout of a DICOM file to aid surgical planning in a 6 year old patient with a large scapular osteochondroma complicating congenital diaphyseal aklasia.** *J Radiol Case Rep* 2012, **6**(1):31-37.
34. Khan FA, Lipman JD, Pearle AD, Boland PJ, Healey JH: **Surgical technique: Computer-generated custom jigs improve accuracy of wide resection of bone tumors.** *Clin Orthop Relat Res* 2013, **471**(6):2007-2016.
35. Nam D, McArthur BA, Cross MB, Pearle AD, Mayman DJ, Haas SB: **Patient-specific instrumentation in total knee arthroplasty: a review.** *J Knee Surg* 2012, **25**(3):213-219.
36. Raaijmakers M, Gelaude F, De Smedt K, Clijmans T, Dille J, Mulier M: **A custom-made guide-wire positioning device for hip surface replacement arthroplasty: description and first results.** *BMC Musculoskelet Disord* 2010, **11**:161.
37. Sassoon A, Nam D, Nunley R, Barrack R: **Systematic review of patient-specific instrumentation in total knee arthroplasty: new but not improved.** *Clin Orthop Relat Res* 2015, **473**(1):151-158.
38. Victor J, Dujardin J, Vandenuecker H, Arnout N, Bellemans J: **Patient-specific guides do not improve accuracy in total knee arthroplasty: a prospective randomized controlled trial.** *Clin Orthop Relat Res* 2014, **472**(1):263-271.
39. Harrysson OL, Hosni YA, Nayfeh JF: **Custom-designed orthopedic implants evaluated using finite element analysis of patient-specific computed tomography data: femoral-component case study.** *BMC Musculoskelet Disord* 2007, **8**:91.
40. Derby B: **Printing and prototyping of tissues and scaffolds.** *Science* 2012, **338**(6109):921-926.
41. Kern R: **3-D Printed Implants Hit the Market, Pave the Way For More Personalized Devices.** *The Grey Sheet* 2013, **39**(44):1-3.
42. Doyle: **3 ways 3D printing could revolutionize healthcare.** *HealthBiz Decoded* 2014.
43. Starosolski Z, Kan J, Rosenfeld S: **Application of 3D printing (rapid prototyping) for creating physical models of pediatric orthopaedic disorders.** *Pediatric Radiology* 2013, **44**:216-221.
44. Maruthappu M, Keogh B: **How might 3D printing affect clinical practice?** *BMJ* 2014, **349**:g7709.
45. Gibbs DM, Vaezi M, Yang S, Oreffo RO: **Hope versus hype: what can additive manufacturing realistically offer trauma and orthopedic surgery?** *Regen Med* 2014, **9**(4):535-549.
46. White PB, Ranawat AS: **Patient-Specific Total Knees Demonstrate a Higher Manipulation Rate Compared to "Off-the-Shelf Implants".** *The Journal of arthroplasty* 2015.
47. Imanishi J, Choong PF: **Three-dimensional printed calcaneal prosthesis following total calcaneotomy.** *Int J Surg Case Rep* 2015, **10**:83-87.
48. Qiao F, Li D, Jin Z, Gao Y, Zhou T, He J, Cheng L: **Application of 3D printed customized external fixator in fracture reduction.** *Injury* 2015, **46**(6):1150-1155.
49. Di Bella C, Fosang A, Donati DM, Wallace GG, Choong PF: **3D Bioprinting of Cartilage for Orthopedic Surgeons: Reading between the Lines.** *Frontiers in surgery* 2015, **2**:39.
50. Boland T, Xu T, Damon B, Cui X: **Application of inkjet printing to tissue engineering.** *Biotechnol J* 2006, **1**(9):910-917.
51. Murphy SV, Atala A: **3D bioprinting of tissues and organs.** *Nat Biotechnol* 2014, **32**(8):773-785.
52. Tasoglu S, Demirci U: **Bioprinting for stem cell research.** *Trends Biotechnol* 2013, **31**(1):10-19.

53. Cui X, Breitenkamp K, Finn MG, Lotz M, D'Lima DD: **Direct human cartilage repair using three-dimensional bioprinting technology.** *Tissue Eng Part A* 2012, **18**(11-12):1304-1312.
54. Markstedt K, Mantas A, Tournier I, Martinez Avila H, Hagg D, Gatenholm P: **3D Bioprinting Human Chondrocytes with Nanocellulose-Alginate Bioink for Cartilage Tissue Engineering Applications.** *Biomacromolecules* 2015, **16**(5):1489-1496.
55. Lee JW, Choi YJ, Yong WJ, Pati F, Shim JH, Kang KS, Kang IH, Park J, Cho DW: **Development of a 3D cell printed construct considering angiogenesis for liver tissue engineering.** *Biofabrication* 2016, **8**(1):015007.
56. Pinnock CB, Meier EM, Joshi NN, Wu B, Lam MT: **Customizable engineered blood vessels using 3D printed inserts.** *Methods* 2015.
57. Montmartin M, Meyer C, Euvrard E, Pazart L, Weber E, Benassarou M: **[3D printing in health care facilities: What legislation in France?].** *Rev Stomatol Chir Maxillofac Chir Orale* 2015.
58. Knoepfler PS: **From bench to FDA to bedside: US regulatory trends for new stem cell therapies.** *Advanced drug delivery reviews* 2015, **82-83**:192-196.
59. Wei CC, Lin AB, Hung SC: **Mesenchymal stem cells in regenerative medicine for musculoskeletal diseases: bench, bedside, and industry.** *Cell transplantation* 2014, **23**(4-5):505-512.
60. Lethaus B, Poort L, Bockmann R, Smeets R, Tolba R, Kessler P: **Additive manufacturing for microvascular reconstruction of the mandible in 20 patients.** *J Craniomaxillofac Surg* 2012, **40**(1):43-46.
61. Miller JS, Stevens KR, Yang MT, Baker BM, Nguyen DH, Cohen DM, Toro E, Chen AA, Galie PA, Yu X *et al*: **Rapid casting of patterned vascular networks for perfusable engineered three-dimensional tissues.** *Nat Mater* 2012, **11**(9):768-774.
62. Ricci JL, Clark EA, Murriky A, Smay JE: **Three-dimensional printing of bone repair and replacement materials: impact on craniofacial surgery.** *J Craniofac Surg* 2012, **23**(1):304-308.
63. Sawkins MJ, Mistry P, Brown BN, Shakesheff KM, Bonassar LJ, Yang J: **Cell and protein compatible 3D bioprinting of mechanically strong constructs for bone repair.** *Biofabrication* 2015, **7**(3):035004.
64. Gao G, Schilling AF, Yonezawa T, Wang J, Dai G, Cui X: **Bioactive nanoparticles stimulate bone tissue formation in bioprinted three-dimensional scaffold and human mesenchymal stem cells.** *Biotechnol J* 2014, **9**(10):1304-1311.
65. Smith DM, Cray JJ, Jr., Weiss LE, Dai Fei EK, Shakir S, Rottgers SA, Losee JE, Campbell PG, Cooper GM: **Precise control of osteogenesis for craniofacial defect repair: the role of direct osteoprogenitor contact in BMP-2-based bioprinting.** *Ann Plast Surg* 2012, **69**(4):485-488.
66. Zhang W, Lian Q, Li D, Wang K, Hao D, Bian W, He J, Jin Z: **Cartilage repair and subchondral bone migration using 3D printing osteochondral composites: a one-year-period study in rabbit trochlea.** *Biomed Res Int* 2014, **2014**:746138.
67. Fedorovich NE, Schuurman W, Wijnberg HM, Prins HJ, van Weeren PR, Malda J, Alblas J, Dhert WJ: **Biofabrication of osteochondral tissue equivalents by printing topologically defined, cell-laden hydrogel scaffolds.** *Tissue Eng Part C Methods* 2012, **18**(1):33-44.
68. Lee CH, Rodeo SA, Fortier LA, Lu C, Erisken C, Mao JJ: **Protein-releasing polymeric scaffolds induce fibrochondrocytic differentiation of endogenous cells for knee meniscus regeneration in sheep.** *Science translational medicine* 2014, **6**(266):266ra171.

Author Manuscript

Figures

Figure 1. 3D model of fractured pelvis to assist with surgical planning and surgery.

“This figure is based on a study first published in Chin Med J, 2015, 128(4): 477-482”

Figure 2. 3D model of planned osteotomy with jig with osteotomy guide.

(©Materialise 2015 All rights reserved)

Figure 3. Patient Specific 3D printed guides for knee replacement surgery.

(©Materialise 2015 All rights reserved)

Figure 4. 3D printed calcaneus replacement (Printed with permission from Imanishi, Choong PF [47]).

Figure 5. 3D printed meniscus in a sheep model (Printed with permission from Lee et al.[68])

Fig1 Pelvis R1.jpg

fig 6 Meniscus.jpg

Fig 2 Osteotomy R1.jpg

Fig 3 Custom copy R1.jpg

Fig 5 Calc.jpg

Minerva Access is the Institutional Repository of The University of Melbourne

Author/s:

Mulford, JS;Babazadeh, S;Mackay, N

Title:

Three-dimensional printing in orthopaedic surgery: review of current and future applications

Date:

2016-09-01

Citation:

Mulford, J. S., Babazadeh, S. & Mackay, N. (2016). Three-dimensional printing in orthopaedic surgery: review of current and future applications. ANZ JOURNAL OF SURGERY, 86 (9), pp.648-653. <https://doi.org/10.1111/ans.13533>.

Persistent Link:

<http://hdl.handle.net/11343/291159>