Long title: CT brain perfusion: a static phantom study of contrast-to-noise ratio and radiation dose

Short title: CTP contrast-to-noise ratio and radiation dose

Stewart M. Midgley Ph.D (1, 2), Dr. Damien L Stella (1), Dr. Bruce C.V. Campbell (3),
Francesca Langenberg (1), and Paul F Einsiedel (1)

(1) Department of Radiology,
Melbourne Health and University of Melbourne
Parkville, VIC 3050, Australia

Tel.: +61-3-9342-7255

Fax: +61-3-9342-8369

email: stewart.midgley@monash.edu

(2) South Australian Medical Imaging,
Flinders Medical Centre, Bedford park SA 5042

Tel: +61 8 8204 4642

Fax: +61 8 8204 6193

email: sami.dimp.fmc@gmail.com

(3) Department of Medicine and Neurology,
Melbourne Health and University of Melbourne,
Parkville, VIC 3050, Australia

Tel.: +61-3-8344 6252

Fax: +61-3-9347 1863

Type of Manuscript: Original Article

Conflicts of interest

The authors declare that they have no financial or other conflict of interest.

This is the author manuscript accepted for publication and has undergone full peer review but has not been through the copyediting, typesetting, pagination and proofreading process, which may lead to differences between this version and the [Version of Record](#). Please cite this article as [doi: 10.1111/1754-9485.12561](https://doi.org/10.1111/1754-9485.12561)

This article is protected by copyright. All rights reserved

Funding information

Unfunded research with no award or funding

Author Manuscript

Received Date : 28-Jul-2016

Revised Date : 11-Oct-2016

Accepted Date : 26-Oct-2016

5 Article type : Radiology Original Article

CT brain perfusion: a static phantom study of contrast-to-noise ratio and radiation dose

Abstract

10 **Objectives:** CT perfusion (CTP) is increasingly employed in the diagnosis and management of ischemic stroke but radiation dose can be significant and optimising contrast to noise ratio (CNR) is challenging. This study aimed to quantify and optimise the balance between CNR as a surrogate for image quality and radiation dose.

20 **Methods:** A perspex head phantom with vials of dilute contrast agent was scanned using a Siemens Definition Flash 128-slice scanner. The CTP protocol exposure parameters were adjusted over 70-120 kVp and 150-285 mAs. Measurements were obtained for the average dose per slice in Hounsfield Units (HU) for iodinated contrast agent, and the image noise for background regions of perspex. The CNR was measured as a function of the volumetric CT dose index (CTDIvol) and kVp.

25 **Results:** A change from 120 to 80 kVp, achieved the same CNR with 60% reduction in dose. Alternatively, for the same dose, the change from 120 to 80 kVp improved CNR by +58%. A change from 80 to 70 kVp whilst operating at the same CNR, led to 13% reduction in dose. Alternatively, maintaining the same dose while changing from 80 to 70 kVp improved the CNR by +7%.

30 **Conclusion:** Lower beam energies achieved the same CNR with less dose, or improved CNR at the same dose. A reduction from 80 kVp to 70 kVp may be clinically useful to optimise CTP acquisitions.

keywords CT brain perfusion, contrast-to-noise ratio, radiation dose

Author Manuscript

Introduction

40 In suspected ischemic stroke, non-contrast CT brain may be normal in the first few hours after onset when potential reperfusion therapies are considered. CT perfusion (CTP) has the capacity to visualise the blood flow abnormality in ischemic stroke and therefore confirm the diagnosis as well as providing information about the severity of hypoperfusion and likely salvageability of brain tissue [1, 2]. The concept that patients with a large
45 irreversibly injured region of brain, identified using CTP, are less likely to respond favourably to reperfusion was utilised in the selection criteria for two of the recent positive clinical trials of endovascular thrombectomy [3, 4]. CTP is incorporated in the routine imaging of ischemic stroke at some centres [5] but concerns are often raised regarding radiation dose [6] and contrast-to-noise ratio (CNR) can be limited as a result [7]. The
50 acquisition involves repeated CT scanning of the brain during the passage of an intravenous bolus of iodinated contrast. The four dimensional data set is then processed to create statistical parametric maps (SPM) typically including cerebral blood flow (CBF), cerebral blood volume (CBV), mean transit time (MTT) and either time to peak (TTP) or time to maximum (Tmax) for ease of interpretation.

55 The scan protocol for CTP has evolved with advances in CT scanner technology. Very early CTP studies utilised the same scan parameters as CT brain studies operating at 120 kVp [8]. Faster detection systems have enabled dynamic studies at lower kVp, leading to the CTP scan protocol using 80 kVp to achieve reduced radiation dose to the patient and
60 improved sensitivity to iodine based contrast agent [9]. Modern CT scanners with multiple rows of detectors are capable of rapid volume scanning, and advances in the X-ray source technology have now enabled two manufacturers to reduce their minimum beam energy from 80 to 70 kVp. Whilst lower energies deliver an improvement in radiographic contrast, the beam is less penetrating, so there is a reduction in the number of transmitted photons
65 with an accompanying increase in noise. This loss of image quality can be offset by operating at higher mAs but at the cost of an increase in radiation dose to the patient.

Recent clinical studies have examined the influence on image quality and radiation dose offered by 70 kVp over higher energy spectra. Qualitative measures of image quality for
70 brain angiography [10, 11] for 70 kVp over 120 kVp, found increased noise, increased CT numbers expressed in Hounsfield Units (HU), with similar image quality and significant dose

75 saving (85% less). Comparison between 80 kVp with iterative reconstruction against 70 kVp with filtered backprojection (FBP) and different mAs for CTP achieved similar image quality at less dose and no significant differences in parametric maps [12]. For CTP at the same mAs, 70 kVp versus 80 kVp [13] uses 35% less dose, achieving similar image quality in parametric maps for CBF and CBV, with modest change in MTT for white matter.

80 This phantom based study aims to quantify the trade off between both image quality expressed as the CNR and radiation dose in the context of CTP. The absence of a suitable CTP phantom, prevents a dynamic study and subsequent analysis of SPM. Instead we focus on a single static frame to quantify radiographic contrast, image noise and patient dose metrics across a clinically relevant range of exposure settings kVp and mAs. The relationships between all parameters will be used to balance the trade-offs between CNR and radiation dose.

85

Materials and methods

90 The study was conducted using a Siemens Definition Flash 128-slice CT scanner (Siemens. Erlangen, Germany), with the CTP protocol technique factors summarised in table 1. The parameters chosen were centred around our clinical protocol utilising 80 kVp and 200 effective mAs (equal to the ratio of mAs to helical pitch), with a maximum mAs of 285 chosen as this was the system limit at 70 kVp. Whilst the Siemens flash scanner is Dual Source, during perfusion imaging it only uses a single source to operate as a conventional CT scanner, and acquires 30 temporal frames over 10 cm axial length reconstructed to 10x10 mm thick slices. Note that helical scanning involves over-beaming to collect an extra half rotation of data at the start and end of the scan [14], and this leads to the scanned volume being longer than the reconstructed volume. No dose modulation techniques were utilised and the system did not offer iterative reconstruction for CT perfusion.

100

The perspex CTDI head phantom (CIRS. Norfolk, USA) used is 16 cm diameter by 14 cm axial length with a physical density and composition approximating brain tissue. It was necessary to wrap the phantom with neoprene rubber (see figure 1) to suppress a strong cupping artefact arising from the skull attenuation correction employed by the CTP image reconstruction algorithm.

105

Vials of varying dilutions of iodine based contrast agent (table 2) were loaded into the five apertures contained within the perspex cylinder and the phantom scanned at four beam energies 70, 80, 100 and 120 kVp, each at three different exposures varying between 150-285 mAs. The five concentrations evaluated ranged between 1-10% to best mimic the *in vivo* situation, as radiographic contrast agent undergoes volumetric dilution to approximately 5% following IV injection and mixing with blood during a first pass through the heart and lungs [15].

The reconstructed data was exported to a workstation for quantitative region of interest (ROI) analysis with the open source tool box provided by ImageJ [16]. A circular ROI was drawn to enclose the contents of each vial excluding any air and the mean HU was used as a measure of radiographic contrast relative to the perspex background. A measure of image noise was obtained as the standard deviation for HU within a large circular ROI of perspex located midway between the centre and edge. Image quality was quantified using the contrast to noise ratio (CNR) as a figure of merit, evaluated as the ratio of mean HU for iodine to the standard deviation for nearby regions of perspex. Radiation dose is reported by the scanner as the volumetric CT dose index (CTDIvol) and is summarised in table 3. The effective dose in milliSieverts (mSv) was calculated using the CTExpo (V2.3.1) CT dose calculator [17] and modern tissue weighting factors [18].

Results

The mean HU for each iodine concentration established for each kVp are presented in figure 2, with error bars representing one standard deviation for the iodine ROI. The relationship between HU and iodine concentration is linear, up to a maximum available value of 3072 (as 12 bits are used to span the range -1024 to +3072), which translates to concentrations 96-180 mg/ml for the respective range 70-120 kVp.

Measured image noise is presented in figure 3 in units of HU. The curves shown were obtained by fitting to a polynomial function of the reciprocal of the square root of CTDIvol, assuming that Poisson counting statistics govern the detection system.

Measured CNR is presented as a function of CTDIvol in figure 4 for each iodine

140 concentration, beam quality (kVp) and quantity (mAs). Non-linear regression analysis employed a Marquardt-Levenberg algorithm, to fit curves [19] to the measured CNR at each kVp against the square root of CTDI, and weighted by the reciprocal of HN uncertainties of 2-10% presented in figure 3. As detailed statistical analysis was not required, we did not use the 1-3% uncertainties (ratio of one standard deviation to the
145 mean) for the fitted coefficients.

Tables 4 and 5 show changes in dose and CNR required to maintain the same CNR and dose respectively relative to variations in kVp from the clinical protocol of 80 kVp and 200 mAs. The fitted curves were used to draw the horizontal and vertical arrows joining
150 different kVp results. For the same CNR, lower kVp delivers less dose; -39% (120 to 80 kVp), -42% (100 to 80 kVp) and -13% (80 to 70 kVp). For the same dose, lower energies achieve better CNR: +58% (120 to 80 kVp), +32% (100 to 80 kVp) and +7% (80 to 70 kVp).

155 Table 6 summarises the calculation steps for estimating effective dose (E) based upon exposure factors required to achieve the same CNR at each kVp. The clinical protocol (table 1) was scaled to estimate CTDIvol using results in table 4, with effective mAs estimated using table 3, DLP calculated for 10 cm scan length repeated over 30 temporal frames and effective dose was calculated as described in the materials and methods.

160

Discussion

The detection task for CTP is to resolve small changes in HU due to the presence or absence of iodinated contrast agent in minute blood vessels within the brain, from the
165 background image noise. For CTP, each volume scan is repeated over multiple temporal frames, so there is a need to optimise the trade off between CNR and radiation dose.

Our phantom based measurements quantify how the radiographic contrast improves at lower kVp. Figure 2 shows 50% improvement in contrast by changing from 120 to 80 kVp
170 and a further 20% improvement by changing 80 to 70 kVp. Lower kVp is also accompanied by a reduction in dose with a subsequent increase in background noise due to there being fewer transmitted photons (see Figure 3).

Our findings in Figure 4 show that using our current clinical CTP scan protocol (80 kVp, 200 mAs and CTDIvol of 8.5 mGy) as a comparator to achieve the same CNR (and presumed clinical diagnostic utility) at 70 kVp a dose saving of 13% is achieved. Whilst this is a relatively small improvement, for a CTP protocol with scan length of 10 cm with 30 temporal frames we estimate the total effective dose saving in moving to a 70 kVp CTP protocol is about 0.6 mSv. Our results also show that performing CTP at 100 to 120 kVp results in unacceptably high patient doses of 10.4 to 14.6 mSv in achieving the same CNR as for 80 kVp. Similar changes in dose for the same CNR (horizontal displacements in figure 4) are reported in recent clinical and phantom based studies [10-11] for the CT angiography protocol. Clinical CTP using the same mAs for reduced kVp [13], is represented in figure 4 as the diagonal shifts between kVp curves to reduced dose and CNR. These trends are due to there being fewer photons in the transmitted spectrum leading to reduced dose and increased noise. When changing 80 kVp to 70 kVp, even fewer photons are available because the K-edge for a Tungsten anode is 69.5 keV so characteristic x-ray production is almost absent. Alternately (if clinically desired), in moving from an 80 to a 70 kVp CTP protocol whilst maintaining the same radiation exposure, (vertical displacement in figure 4) provides a 7% improvement in CNR. Statistical parametric analysis delivers quantitative information about dynamic flow parameters, which are not available from our static phantom study. We would anticipate improved CNR to deliver SPM with similar mean values and reduced uncertainty (i.e., less noise).

It should be noted that the results in this study are not unexpected on the basis of well established radiation physics models [14, 20], but they do provide experimental evidence that they may be applicable to the clinical scenario of brain CTP given the phantom we used. An assumption in the study design is that the concentration of intra arterial iodinated contrast within microvasculature (not ever evaluated as far as we are aware) mirrors that as established in larger central vessels (average of 5%). In addition, we also assume that the image noise for regions of brain tissue is similar to that for regions of perspex. We believe these are reasonable assumptions.

Our measurements considered typical in vivo arterial concentrations of contrast agent with relative dilution of 0.9-9%. In the theoretical scenario where mAs could be adjusted such that CT scans at 70 and 120 kVp achieved the same CNR we would expect the diagnostic outcome to be similar as these images are utilized to calculate change over time (eg.

Blood flow). The absolute numbers generated are not relevant, as these images are not used for diagnostic purposes in their own right (the individual images generated in clinical CTP are purposely as low dose as possible and hence are considered 'non diagnostic'). CT perfusion aims to detect small changes in blood flow in the brain parenchyma in order to accurately assess cerebral blood flow and the other perfusion parameters. Increased image noise reduces the ability to detect these changes as the tissue enhancement in white matter may be only a few HU. The criteria for visibility [14, 20] is a CNR greater than 5, which is approached at lower concentrations and low doses. Extrapolation of the results in figure 4 (c) and (d) suggest such a limit is reached at 70-80 kVp for approximately 0.5% relative dilution and 5 mGy CTDIvol. The CNR for microvasculature in vivo might be reduced due to partial volume effects. Strategies for improving the CNR include increasing the iodine delivery rate (by raising one or both of the iodinated contrast concentration or the rate of intravenous injection), and/or reducing the image noise by temporal filtering and using iterative reconstruction over filtered back projection (not an option on our scanner for perfusion CT at the time of the study).

Conclusions

Our head phantom based measurements have demonstrated that changing the CTP protocol from 80 to 70 kVp can deliver a 13% dose reduction (for the same CNR) or a 7% improvement in CNR (for the same dose). We believe these results can be used to justify an in vivo investigation of CTP at 70 kVp rather than the currently used value of 80 kVp.

Acknowledgements

The authors thank and acknowledge assistance from CT radiographers Alida Finegan and Megan Dwyer.

References

- 1 Campbell BCV, Christensen S, Levi CR, Desmond PM, Donnan GA, Davis SM, and Parsons MW. Comparison of Computed Tomography Perfusion and Magnetic Resonance Imaging Perfusion-Diffusion Mismatch in Ischemic Stroke. *Stroke* 2012; 43(10): 2648-2653. doi: 10.1161/STROKEAHA.112.660548

- 2 Bivard A., Spratt N, Levi C and Parsons M. Perfusion computer tomography: imaging and clinical validation in acute ischaemic stroke. *Brain* 2011; 134(11): 3408-3416. doi: 10.1093/brain/awr257.
- 3 Campbell BCV, Mitchell PJ, Kleinig TJ, *et al*, Endovascular Therapy for Ischemic Stroke with Perfusion-Imaging Selection *New Engl J Med* March 12, 2015; 372(11); 1009-1018. DOI: 10.1056/NEJMoa1414792
- 4 Saver JL, Goyal M, Bonafe A, *et al* Stent-Retriever Thrombectomy after Intravenous t-PA vs. t-PA Alone in Stroke *N Engl J Med* June 11, 2015; 372:(24) 2285-2295. DOI: 10.1056/NEJMoa1415061
- 5 Wintermark M, Sanelli PC, Albers GW, *et al*, Imaging Recommendations for Acute Stroke and Transient Ischemic Attack Patients: A Joint Statement by the American Society of Neuroradiology, the American College of Radiology, and the Society of NeuroInterventional Surgery. *Am J Neuroradiol* 2013; 34(11); E117-E127. DOI: 10.3174/ajnr.A3690
- 6 Wintermark M. and Lev M.H, FDA Investigates the safety of brain perfusion CT. *Am. J. Neuroradiol.* 2010; 31(1); 2-3. doi: 10.3174/ajnr.A1967
- 7, Schaefer P.W., Douza L, Kamalian S *et al*/ Limited Reliability of Computed Tomographic Perfusion Acute Infarct Volume Measurements Compared With Diffusion-Weighted Imaging in Anterior Circulation. *Stroke* 2015; 46; 2,419-424. doi: [10.1161/STROKEAHA.114.007117](https://doi.org/10.1161/STROKEAHA.114.007117)
- 8, Cohnena M., Wittsack HJ, Assadi S *et al*, Radiation Exposure of Patients in Comprehensive Computed Tomography of the Head in Acute Stroke. *Am J Neuroradiol* 2006; 27(8); 1741-1745. PMID: [16971627](https://pubmed.ncbi.nlm.nih.gov/16971627/)
- 9, Wintermark M., Maeder P, Verdun FR, *et al*, Using 80 kVp versus 120 kVp in Perfusion CT Measurement of Regional Cerebral Blood Flow. *Am J Neuroradiol* 2000; 21(10); 1881-

1884. PMID: [11110541](#)

- 10, Cho E.S., Chung Ts, Ahn SJ, *et al.* Cerebral computed tomography angiography using
280 a 70 kVp protocol: improved vascular enhancement with a reduced volume of contrast
medium and radiation dose, *Eur Radiol* 2015: 25(5); 1421-1430. doi: 10.1007/s00330-014-
3540-z
- 11, Chen GZ, Zhang LJ, Wichmann JL, *et al.* Radiation dose and image quality of 70 kVp
285 cerebral CT angiography with optimized sinogram-affirmed iterative reconstruction:
comparison with 120 kVp cerebral CT angiography, *Eur Radiol* 2015: 25(5); 1453-1463.
doi: 10.1007/s00330-014-3533-y
- 12, Corcuera-Solano I., McLellan Am, Doshi AH, Pawha PS and Tanenbaum LN, Whole-
290 Brain Adaptive 70-kVp Perfusion Imaging with Variable and Extended Sampling Improves
Quality and Consistency While Reducing Dose, *Am J Neuroradiol* 2014: 35 (11); 2045-51.
doi: 10.3174/ajnr.A4043
- 13, Li ZN, LI H, Zhang K, *et al* Improvement of image quality and radiation dose of CT
295 perfusion of the brain by means of low-tube voltage (70 KV) *European Radiology* August
2014, Volume 24, Issue 8, pp 1906-1913. DOI 10.1007/s00330-014-3247-1
- 14, Bushberg J.T., Seibert JA, Leidholdt EM, and Boone JMI, The Essential Physics of
300 Medical Imaging, third edition}, (Lippincott Williams & Wilkins, Philadelphia 2012) 92; 176-
180, 338-340
- 15, AAPM, Report 15. Performance evaluation and quality assurance in digital subtraction
angiography, (American Association of Physicists in Medicine New York 1985); 19-20.
- 305 16, Schneider, C.A., Rasband W.S. and Eliceiri, K.W., NIH to imageJ: 25 years of image
analysis, *Nature Methods*. 2012: 9(7); 671-675. doi:10.1038/nmeth.2089

17, Stamm G. and Nagel H.D., (2002) CT-expo a novel program for dose evaluation in CT [Article In German]. *Fortschritte auf dem Gebiete der Rontgenstrahlen und der Nuklearmedizin* 2002; 174(12): 1570-1576. DOI: [10.1055/s-2002-35937](https://doi.org/10.1055/s-2002-35937)

315 18, International Commission on Radiological Protection (ICRP). The 2007 Recommendations of the International Commission on Radiological Protection. *Ann ICRP*. 2007;37:1–332. ICRP publication 103.

19, Janert P.K., Gnuplot in action: understanding data with graphs, (Manningham Publications, Greenwich CT, 2010) 190-199.

20, Flower M.A., 2012 Webb's Physics of medical imaging second edition, (CRC Press, Boca Raton, 2012), 24, 34--36, 745.

Table 1 Exposure parameters for the clinical CTP protocol at the time of the study.

Scan length (mm)	118
kVp selected	70, 80, 100, 120
Gantry rotation (s)	0.285
Helical pitch, P	0.5
Effective mAs	150, 200, 285
Dose modulation	OFF
Collimation x Pitch (mm)	128 x 0.6 x 0.5= 38.4
Reconstruction filter	H20f
Reconstructed voxel side (mm)	0.39
Reconstructed slice thickness (mm)	10
Reconstruction volume axial length (mm)	100

330 **Table 2** Measured mass density ρ and estimates for iodine concentrations.

Dilution (%)	ρ (g/cm ³)	Iodine (mg/ml)
100	1.402	350
8.7	1.015	30.5
7.1	1.012	25
3.9	1.005	13.5
1.9	1.000	6.6
0.94	0.994	3.3

Table 3 CTDIvol reported by the scanner as a function of kVp and effective mAs

Effective mAs	CTDIvol (mGy)			
	70	80	100	120
150	4.13	6.35	12.3	19.5
200	5.40	8.46	16.4	26.0

285	7.84	12.1	23.4	37.1
-----	------	------	------	------

Table 4 Relative change in CTDIvol for fixed CNR and variable kVp.

I (mg/ml)	70 to 80 kVp	90 to 100 kVp	100 to 120 kVp
30.5	1.14	1.71	2.50
25.0	1.16	1.66	2.38
13.5	1.16	1.68	2.27
6.6	1.14	1.68	2.27
3.3	1.18	1.98	3.13
mean	1.15	1.74	2.50

335

Table 5 Relative CNR change, for fixed CTDIvol and variable kVp.

I (mg/ml)	70 to 80 kVp	80 to 100 kVp	100 to 120 kVp
30.5	1.06	0.76	0.63
25.0	1.06	0.78	0.65
13.5	1.07	0.77	0.66
6.6	1.07	0.77	0.66
3.3	1.08	0.79	0.56
mean	1.07	0.77	0.63

Table 6 Estimation of effective dose (E) at each kVp for 30 frames, 10 cm scan length and same CNR.

kVp	70	80	100	120
Effective mAs	270	200	180	160
CTDIvol (mGy)	7.4	8.5	14.8	21.3
DLP (mGycm)	2230	2540	4440	6240
E (mSv)	5.3	5.9	10.4	14.6

Figure 1 The CTDI head phantom wrapped with neoprene rubber, showing vials of dilute contrast agent inserted into the five aperture. Regions of interest (red) enclose the vial contents avoiding any trapped air (9 o'clock) whilst the 6 o'clock vial is in an adjacent slice.

Figure 2 Measured HU as a function of iodine concentration and kVp,

Figure 3 Image noise in regions of perspex as a function of the radiation dose per slice (CTDIvol), kVp (lines) and mAs (points), with the filled black circle denoting the CTP protocol used for this study.

Figure 4 CNR as a function of radiation dose, kVp (curves) and mAs (points) for (a) 25 mg/ml (7.1%), (b) 13.5 mg/ml (3.9%), (c) 6.6 mg/ml (1.9%) and (d) 3.3 mg/ml (0.9%). Error

355 bars denote one standard deviation for each iodine ROI. A filled black circle identifies the CTP scan protocol used in this study (80 kVp and 200 mAs for 8.5 mGy). The vertical and horizontal arrows denote CNR for the same dose and dose for the same CNR respectively.

360 end

Author Manuscript

ara_12561_f1.eps

ara_12561_f2.eps

ara_12561_f4a.eps

ara_12561_f4b.eps

ara_12561_f4c.eps

ara_12561_f4d.eps

Minerva Access is the Institutional Repository of The University of Melbourne

Author/s:

Midgley, SM;Stella, DL;Campbell, BCV;Langenberg, F;Einsiedel, PF

Title:

CT brain perfusion: A static phantom study of contrast-to-noise ratio and radiation dose

Date:

2017-06-01

Citation:

Midgley, S. M., Stella, D. L., Campbell, B. C. V., Langenberg, F. & Einsiedel, P. F. (2017).
CT brain perfusion: A static phantom study of contrast-to-noise ratio and radiation dose.
JOURNAL OF MEDICAL IMAGING AND RADIATION ONCOLOGY, 61 (3), pp.361-366.
<https://doi.org/10.1111/1754-9485.12561>.

Persistent Link:

<http://hdl.handle.net/11343/292165>